

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134321

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/16/2025	18:25
2	CAL2	1	Water	NA	NA	20.2	7.00	01/16/2025	18:26
3	CAL3	1	Water	NA	NA	20.2	10.02	01/16/2025	18:27
4	ICV	1	Water	NA	NA	20.2	7.01	01/16/2025	18:29
5	CCV1	1	Water	NA	NA	20.2	2.01	01/16/2025	18:30
6	Q1092-02	1	Solid	20.02	20	20.1	6.30	01/16/2025	18:35
7	Q1092-03	1	Solid	20.03	20	20.6	7.22	01/16/2025	18:37
8	Q1115-02	1	Solid	20.02	20	20.6	6.72	01/16/2025	18:40
9	Q1115-05	1	Solid	20.04	20	20.5	8.76	01/16/2025	18:42
10	Q1115-08	1	Solid	20.02	20	20.9	8.60	01/16/2025	18:47
11	Q1115-11	1	Solid	20.03	20	20.4	9.52	01/16/2025	18:55
12	Q1119-01	1	Solid	20.02	20	20.3	5.56	01/16/2025	19:00
13	Q1119-01DUP	1	Solid	20.03	20	20.4	5.57	01/16/2025	19:01
14	CCV2	1	Water	NA	NA	20.3	12.02	01/16/2025	19:05

WORKLIST(Hardcopy Internal Chain)

18
~~AP 201310~~
 UB 134321

WorkList Name : corrosivity Q1092 WorkList ID : 186949 Department : Wet-Chemistry Date : 01-16-2025 09:55:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1119-01	FILTER CAKE	Solid	Corrosivity	Cool 4 deg C	ARAM01	M11	01/16/2025	9045D
Q1092-02	428	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/15/2025	9045D
Q1092-03	429	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/15/2025	9045D
Q1115-02	366	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	01/16/2025	9045D
Q1115-05	VNJ-214	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	01/16/2025	9045D
Q1115-08	72-11995	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	01/16/2025	9045D
Q1115-11	VNJ-213	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	01/16/2025	9045D

Date/Time 01-16-25 18:20

Raw Sample Received by: SP WOC

Raw Sample Relinquished by: CP SM

Date/Time 01-16-25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: SP WOC