



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Verification

**Client:** Aramark Uniforms

**SDG No.:** Q1119

**Project:** Monthly 2025

**RunNo.:** LB134321

| Analyte                   |      | Units | Result | True Value | %<br>Recovery | Acceptance<br>Window (%R) | Analysis<br>Date |
|---------------------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID:<br>Corrosivity | ICV  | pH    | 7.01   | 7          | 100           | 90-110                    | 01/16/2025       |
| Sample ID:<br>Corrosivity | CCV1 | pH    | 2.01   | 2.00       | 101           | 90-110                    | 01/16/2025       |
| Sample ID:<br>Corrosivity | CCV2 | pH    | 12.02  | 12.00      | 100           | 90-110                    | 01/16/2025       |

## Initial and Continuing Calibration Verification

**Client:** Aramark Uniforms

**SDG No.:** Q1119

**Project:** Monthly 2025

**RunNo.:** LB134328

| Analyte                                    | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|--------------------------------------------|-------|--------|------------|------------|------------------------|---------------|
| Sample ID: <b>ICV1</b><br>Reactive Cyanide | mg/L  | 0.097  | 0.099      | 98         | 85-115                 | 01/17/2025    |
| Sample ID: <b>CCV1</b><br>Reactive Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 01/17/2025    |
| Sample ID: <b>CCV2</b><br>Reactive Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 01/17/2025    |
| Sample ID: <b>CCV3</b><br>Reactive Cyanide | mg/L  | 0.24   | 0.25       | 96         | 90-110                 | 01/17/2025    |