

### Duplicate Sample Summary

|                   |                  |                                         |          |
|-------------------|------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Aramark Uniforms | <b>SDG No.:</b>                         | Q1119    |
| <b>Project:</b>   | Monthly 2025     | <b>Sample ID:</b>                       | Q1092-01 |
| <b>Client ID:</b> | 427DUP           | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte          | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|------------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Reactive Cyanide | mg/L  | +/-20               | 0.00099          | U                  | 0.00099             | U                  | 1                  | 0          |      | 01/17/2025       |

## Duplicate Sample Summary

|                                                                                             |                                                                                                    |
|---------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|
| <b>Client:</b> Aramark Uniforms<br><b>Project:</b> Monthly 2025<br><b>Client ID:</b> 428DUP | <b>SDG No.:</b> Q1119<br><b>Sample ID:</b> Q1092-02<br><b>Percent Solids for Spike Sample:</b> 100 |
|---------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------|

| Analyte          | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/AD | Qual | Analysis Date |
|------------------|-------|------------------|---------------|-----------------|------------------|-----------------|-----------------|--------|------|---------------|
| Reactive Cyanide | mg/Kg | +/-20            | 0.0087        | U               | 0.0088           | U               | 1               | 0      |      | 01/17/2025    |
| Reactive Sulfide | mg/Kg | +/-20            | 4.76          | J               | 4.76             | J               | 1               | 0      |      | 01/17/2025    |