



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** Q1119
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1119 **SAS No.:** Q1119
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	22.2	20.0	111	40 - 160	P	01/20/2025	11:06	LB134350
	Barium	88.2	100	88	40 - 160	P	01/20/2025	11:06	LB134350
	Cadmium	6.15	6.0	102	40 - 160	P	01/20/2025	11:06	LB134350
	Chromium	10.3	10.0	103	40 - 160	P	01/20/2025	11:06	LB134350
	Lead	12.2	12.0	102	40 - 160	P	01/20/2025	11:06	LB134350
	Selenium	19.4	20.0	97	40 - 160	P	01/20/2025	11:06	LB134350
	Silver	11.1	10.0	111	40 - 160	P	01/20/2025	11:06	LB134350
CRA	Mercury	0.23	0.2	113	40 - 160	CV	01/20/2025	11:15	LB134341