

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** Q1119
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1119 **SAS No.:** Q1119
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	0.38			-20	20	01/20/2025	11:10	LB134350
	Barium	2.56	6.0	43	-94	106	01/20/2025	11:10	LB134350
	Cadmium	-1.40	1.0	140	-5	7	01/20/2025	11:10	LB134350
	Chromium	57.5	52.0	111	42	62	01/20/2025	11:10	LB134350
	Lead	6.59			-12	12	01/20/2025	11:10	LB134350
	Selenium	-11.5			-20	20	01/20/2025	11:10	LB134350
	Silver	-0.92			-10	10	01/20/2025	11:10	LB134350
ICSAB01	Arsenic	108	104	104	88.4	120	01/20/2025	11:19	LB134350
	Barium	456	537	85	437	637	01/20/2025	11:19	LB134350
	Cadmium	993	972	102	826	1120	01/20/2025	11:19	LB134350
	Chromium	579	542	107	460	624	01/20/2025	11:19	LB134350
	Lead	56.1	49.0	114	37	61	01/20/2025	11:19	LB134350
	Selenium	35.5	46.0	77	26	66	01/20/2025	11:19	LB134350
	Silver	222	201	110	170	232	01/20/2025	11:19	LB134350