

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** Q1119
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1119 **SAS No.:** Q1119
Matrix: Water **Sample ID:** Q1119-01 **Client ID:** FILTER CAKEDUP
Percent Solids for Sample: NA **Duplicate ID** Q1119-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	259		270		4		P
Cadmium	ug/L	20	0.23	J	0.22	J	4		P
Chromium	ug/L	20	5.00	U	5.00	U			P
Lead	ug/L	20	56.2		56.1		0		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	0.69	J	0.63	J	9		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** Q1119
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1119 **SAS No.:** Q1119
Matrix: Water **Sample ID:** Q1119-01MS **Client ID:** FILTER CAKEMSD
Percent Solids for Sample: NA **Duplicate ID** Q1119-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	417		412		1		P
Barium	ug/L	20	380		369		3		P
Cadmium	ug/L	20	97.2		95.8		1		P
Chromium	ug/L	20	202		203		0		P
Lead	ug/L	20	530		523		1		P
Mercury	ug/L	20	34.9		36.1		3		CV
Selenium	ug/L	20	1020		1000		2		P
Silver	ug/L	20	37.3		37.4		0		P