

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms **SDG No.:** Q1119
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q1119 **SAS No.:** Q1119

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166109BS							
Arsenic	ug/L	400	403		101	80 - 120	P
Barium	ug/L	100	86.4		86	80 - 120	P
Cadmium	ug/L	100	95.7		96	80 - 120	P
Chromium	ug/L	200	202		101	80 - 120	P
Lead	ug/L	500	482		96	80 - 120	P
Selenium	ug/L	1000	998		100	80 - 120	P
Silver	ug/L	37.5	37.2		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q1119

Contract: ARAM01

Lab Code: CHEM

Case No.: Q1119

SAS No.: Q1119

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166122BS Mercury	ug/L	4.0	3.87		97	80 - 120	CV