

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ARAM01

Lab Code: CHEM Case No.: Q1119 SAS No.: Q1119 SDG No.: Q1119

Instrument ID: MSVOA_N Calibration Date/Time: 01/17/2025 09:06

Lab File ID: VN085480.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.737	0.682		-7.46	20
1,1-Dichloroethene	0.537	0.542		0.93	20
2-Butanone	0.384	0.350		-8.85	20
Carbon Tetrachloride	0.557	0.584		4.85	20
Chloroform	1.218	1.213		-0.41	20
Benzene	1.463	1.537		5.06	20
1,2-Dichloroethane	0.551	0.574		4.17	20
Trichloroethene	0.341	0.349		2.35	20
Tetrachloroethene	0.341	0.367		7.63	20
Chlorobenzene	1.095	1.136	0.3	3.74	20
1,2-Dichloroethane-d4	0.807	0.803		-0.5	20
Dibromofluoromethane	0.347	0.370		6.63	20
Toluene-d8	1.232	1.364		10.71	20
4-Bromofluorobenzene	0.422	0.475		12.56	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.