

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARAM01
 Lab Code: CHEM Case No.: Q1119 SAS No.: Q1119 SDG NO.: Q1119
 Lab File ID: VN085480.D Date Analyzed: 01/17/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:06
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	217984	8.22	352497	9.09	315106	11.86
UPPER LIMIT	435968	8.718	704994	9.594	630212	12.359
LOWER LIMIT	108992	7.718	176249	8.594	157553	11.359
EPA SAMPLE NO.						
FILTER CAKE	185520	8.22	350475	9.10	319145	11.87
VN0117WBL01	196554	8.22	358389	9.10	314818	11.86
VN0117WBS01	209125	8.22	351707	9.09	304036	11.86
VN0117WBSD01	183955	8.22	311569	9.10	273237	11.86

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.