

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1120

MATRIX: Water

METHOD: 300.0,9034,9060A,SM4500-CN C,E,SM4500-NH3,SM5210 B,SM5220 D

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples. The Matrix Spike (RW10A-20250116MS) analysis met criteria for all samples except for Chloride due to matrix interference. The Matrix Spike Duplicate (DSN002MSD) analysis met criteria for all samples except for Ammonia due to matrix interference. The Matrix Spike Duplicate (RW10A-20250116MSD) analysis met criteria for all samples except for Chloride due to matrix interference.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS: The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

QA REVIEW

Date