

DATA PACKAGE
SEMI-VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.
661 Andersen Drive
Suite 200
Pittsburgh, PA - 15220-2745
Phone No: 412-921-7090

ORDER ID : Q1121
ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) SEMI-VOLATILE DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For SVOC-SIMGroup1	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) LCS/LCSD Summary	15
9.3) Method Blank Summary	17
9.4) GS/MS Tune Summary	18
9.5) Internal Standard Area and RT Summary	20
10) Sample Data	22
10.1) RW10A-20250116	23
11) Calibration Data Summary	34
11.1) Initial Calibration Data	35
11.1.1) BN010225	35
11.2) Continued Calibration Data	223
11.2.1) BN035991.D	223
11.2.2) BN036000.D	249
12) QC Sample Data	275
12.1) Tune Raw Data	276
12.2) Method Blank Data	284
12.3) LCS Data	294
12.4) LCSD Data	318
13) Manual Integration	342
14) Analytical Runlogs	344
15) Extraction Logs	348
15.1) PB166108.pdf	348
15.2) PB166108IC.pdf	350
16) Standard Prep Logs	351
17) Shipping Document	419
17.1) Chain Of Custody	420

Table Of Contents for Q1121

17.2) Lab Certificate	421
17.3) Internal COC	422
18) Not Reviewed Data	423

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : Q1121

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q1121-01

Client Sample Number

RW10A-20250116

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 11:42 am, Jan 27, 2025

Date: 1/24/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager: Ernie Wu

Chemtech Project # Q1121

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

1 Water sample was received on 01/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PB166108BL [2-Fluorobiphenyl - 107%, Nitrobenzene-d5 - 123%], PB166108BS [Nitrobenzene-d5 - 133%], PB166108BSD [2-Fluorobiphenyl - 107%, Nitrobenzene-d5 - 126%] and RW10A-20250116 [Nitrobenzene-d5 - 118%], failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID BN035991.D met the requirements except for 2,4,6-Tribromophenol and Nitrobenzene-d5 , failure surrogates are not associated with the client list, as per criteria affected surrogates were passing; therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 11:43 am, Jan 27, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1121

MATRIX: Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.		✓	
	The Initial Calibration met the requirements .			
	The Continuous Calibration File ID BN035991.D met the requirements except for 2,4,6-Tribromophenol and Nitrobenzene-d5 , failure surrogates are not associated with the client list, as per criteria affected surrogates were passing;therefore no corrective action was taken.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for PB166108BL [2- Fluorobiphenyl - 107%, Nitrobenzene-d5 - 123%], PB166108BS [Nitrobenzene-d5 - 133%], PB166108BSD [2-Fluorobiphenyl - 107%, Nitrobenzene-d5 - 126%] and RW10A-20250116 [Nitrobenzene-d5 - 118%], failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken.			

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples .			
	The Blank Spike Duplicate met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

REVIEWED

QA REVIEW

By Sohil Jodhani, QA/QC Director at 9:56 am, Jan 27, 2025

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1121

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 01/24/2025

LAB CHRONICLE

OrderID:	Q1121	OrderDate:	1/17/2025 7:55:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	M11,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1121-01	RW10A-20250116	Water	SVOC-SIMGroup1	8270-Modified	01/16/25	01/17/25	01/20/25	01/16/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q1121
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :	RW10A-20250116							
Q1121-01	RW10A-20250116	WATER	1,4-Dioxane	0.180	J	0.07	0.2	0.2 ug/L
			Total Svoc :		0.18			
			Total Concentration:		0.18			



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Surrogate Summary

SW-846

SDG No.: Q1121

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166108BL	PB166108BL	2-Methylnaphthalene-d10	0.4	0.44	109		30	150
		Fluoranthene-d10	0.4	0.45	113		30	150
		Nitrobenzene-d5	0.4	0.49	123	*	55	111
		2-Fluorobiphenyl	0.4	0.43	107	*	53	106
		Terphenyl-d14	0.4	0.51	127		58	132
PB166108BS	PB166108BS	2-Methylnaphthalene-d10	0.4	0.43	106		30	150
		Fluoranthene-d10	0.4	0.44	111		30	150
		Nitrobenzene-d5	0.4	0.53	133	*	55	111
		2-Fluorobiphenyl	0.4	0.43	106		53	106
		Terphenyl-d14	0.4	0.45	112		58	132
PB166108BSD	PB166108BSD	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.51	126	*	55	111
		2-Fluorobiphenyl	0.4	0.43	107	*	53	106
		Terphenyl-d14	0.4	0.44	111		58	132
Q1121-01	RW10A-20250116	2-Methylnaphthalene-d10	0.4	0.39	97		30	150
		Fluoranthene-d10	0.4	0.44	110		30	150
		Nitrobenzene-d5	0.4	0.47	118	*	55	111
		2-Fluorobiphenyl	0.4	0.38	94		53	106
		Terphenyl-d14	0.4	0.45	112		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1121

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN035998.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB166108BS	1,4-Dioxane	0.4	0.38	ug/L	95				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1121

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN035999.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB166108BSD	1,4-Dioxane	0.4	0.39	ug/L	98				70	130		20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166108BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1121

SAS No.: Q1121 SDG NO.: Q1121

Lab File ID: BN035992.D

Lab Sample ID: PB166108BL

Instrument ID: BNA_N

Date Extracted: 01/17/2025

Matrix: (soil/water) Water

Date Analyzed: 01/20/2025

Level: (low/med) LOW

Time Analyzed: 11:03

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB166108BS	PB166108BS	BN035998.D	01/20/2025
RW10A-20250116	Q1121-01	BN035997.D	01/20/2025
PB166108BSD	PB166108BSD	BN035999.D	01/20/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: Q1121 SDG NO.: Q1121Lab File ID: BN035870.DDFTPP Injection Date: 01/02/2025Instrument ID: BNA_NDFTPP Injection Time: 10:49

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.9
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	39
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	43.4
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.2
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	9.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	12.1 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN035871.D	01/02/2025	11:28
SSTDICC0.2	SSTDICC0.2	BN035872.D	01/02/2025	12:04
SSTDICCC0.4	SSTDICCC0.4	BN035873.D	01/02/2025	12:40
SSTDICC0.8	SSTDICC0.8	BN035874.D	01/02/2025	13:16
SSTDICC1.6	SSTDICC1.6	BN035875.D	01/02/2025	13:52
SSTDICC3.2	SSTDICC3.2	BN035876.D	01/02/2025	14:28
SSTDICC5.0	SSTDICC5.0	BN035877.D	01/02/2025	15:04



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: Q1121 SDG NO.: Q1121Lab File ID: BN035990.DDFTPP Injection Date: 01/20/2025Instrument ID: BNA_NDFTPP Injection Time: 09:13

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.3
68	Less than 2.0% of mass 69	0.4 (0.8) 1
69	Mass 69 relative abundance	47.1
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	49
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	9.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.1 (18.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035991.D	01/20/2025	10:27
PB166108BL	PB166108BL	BN035992.D	01/20/2025	11:03
RW10A-20250116	Q1121-01	BN035997.D	01/20/2025	14:03
PB166108BS	PB166108BS	BN035998.D	01/20/2025	14:38
PB166108BSD	PB166108BSD	BN035999.D	01/20/2025	15:14
SSTDCCC0.4EC	SSTDCCC0.4	BN036000.D	01/20/2025	16:11

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG NO.: Q1121

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 01/20/2025

Lab File ID: BN035991.D Time Analyzed: 10:27

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2147	7.817	4278	10.61	2153	14.45
UPPER LIMIT	4294	8.317	8556	11.111	4306	14.952
LOWER LIMIT	1073.5	7.317	2139	10.111	1076.5	13.952
EPA SAMPLE NO.						
01 PB166108BL	2441	7.82	4766	10.61	2405	14.45
02 PB166108BS	2154	7.82	4363	10.61	2173	14.45
03 PB166108BSD	2639	7.82	5199	10.61	2411	14.45
04 RW10A-20250116	2375	7.82	4813	10.61	2482	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG NO.: Q1121

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 01/20/2025

Lab File ID: BN035991.D Time Analyzed: 10:27

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4411	17.186	3822	21.367	3850	23.666
UPPER LIMIT	8822	17.686	7644	21.867	7700	24.166
LOWER LIMIT	2205.5	16.686	1911	20.867	1925	23.166
EPA SAMPLE NO.						
01 PB166108BL	4916	17.19	4139	21.37	4174	23.67
02 PB166108BS	4379	17.19	4015	21.37	4099	23.66
03 PB166108BSD	4157	17.19	3274	21.37	3543	23.66
04 RW10A-20250116	5010	17.19	4480	21.37	4572	23.66

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/16/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	01/16/25
Client Sample ID:	RW10A-20250116	SDG No.:	Q1121
Lab Sample ID:	Q1121-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035997.D	1	01/17/25 10:10	01/20/25 14:03	PB166108

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.18	J	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.39		30 - 150		97%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.44		30 - 150		110%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.47	*	55 - 111		118%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		53 - 106		94%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		112%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2380	7.817				
1146-65-2	Naphthalene-d8	4810	10.611				
15067-26-2	Acenaphthene-d10	2480	14.452				
1517-22-2	Phenanthrene-d10	5010	17.186				
1719-03-5	Chrysene-d12	4480	21.367				
1520-96-3	Perylene-d12	4570	23.663				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035997.D
Acq On : 20 Jan 2025 14:03
Operator : RC/JU
Sample : Q1121-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW10A-20250116

Quant Time: Jan 20 14:33:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2375	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4813	0.400	ng	-0.01
13) Acenaphthene-d10	14.452	164	2482	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	5010	0.400	ng	#-0.02
29) Chrysene-d12	21.367	240	4480	0.400	ng	-0.02
35) Perylene-d12	23.663	264	4572	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.398	112	795	0.136	ng	-0.01
5) Phenol-d6	6.972	99	559	0.077	ng	-0.01
8) Nitrobenzene-d5	8.956	82	1803	0.473	ng	-0.03
11) 2-Methylnaphthalene-d10	12.197	152	2493	0.387	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	503	0.422	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4109	0.377	ng	-0.02
27) Fluoranthene-d10	19.216	212	5486	0.441	ng	-0.02
31) Terphenyl-d14	19.815	244	3985	0.446	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.318	88	415	0.176	ng	# 53
34) Bis(2-ethylhexyl)phtha...	21.295	149	580	0.090	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035997.D
Acq On : 20 Jan 2025 14:03
Operator : RC/JU
Sample : Q1121-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Quant Time: Jan 20 14:33:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Instrument :
BNA_N
ClientSampleId :
RW10A-20250116







Tgt Ion:152 Resp: 2375
Ion Ratio Lower Upper
152 100
150 153.3 117.8 176.6
115 67.9 51.0 76.4



#2
1,4-Dioxane
Concen: 0.176 ng
RT: 3.318 min Scan# 19
Delta R.T. -0.004 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion: 88 Resp: 415
Ion Ratio Lower Upper
88 100
43 113.3 32.7 49.1#
58 88.2 63.0 94.4









#4
2-Fluorophenol
Concen: 0.136 ng
RT: 5.398 min Scan# 307
Delta R.T. -0.011 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Tgt Ion:	112	Resp:	795
Ion	Ratio	Lower	Upper
112	100		
64	66.2	48.2	72.2
63	38.4	30.0	45.0



#5
Phenol-d6
Concen: 0.077 ng
RT: 6.972 min Scan# 525
Delta R.T. -0.011 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion:	99	Resp:	559
Ion	Ratio	Lower	Upper
99	100		
42	32.7	23.5	35.3
71	51.2	35.5	53.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 91
Delta R.T. -0.015 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Tgt Ion:	136	Resp:	4813
Ion	Ratio	Lower	Upper
136	100		
137	12.2	10.6	15.8
54	10.3	9.8	14.6
68	7.0	6.6	10.0



#8
Nitrobenzene-d5
Concen: 0.473 ng
RT: 8.956 min Scan# 771
Delta R.T. -0.025 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion:	82	Resp:	1803
Ion	Ratio	Lower	Upper
82	100		
128	33.6	36.9	55.3#
54	69.6	50.4	75.6









#11
2-Methylnaphthalene-d10
Concen: 0.387 ng
RT: 12.197 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Tgt Ion:152 Resp: 2493
Ion Ratio Lower Upper
152 100
151 21.7 17.0 25.6



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1441
Delta R.T. -0.010 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion:164 Resp: 2482
Ion Ratio Lower Upper
164 100
162 102.6 83.1 124.7
160 49.7 46.0 69.0









#14
2,4,6-Tribromophenol
Concen: 0.422 ng
RT: 15.933 min Scan# 11
Delta R.T. -0.024 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Tgt Ion:330 Resp: 503
Ion Ratio Lower Upper
330 100
332 95.4 77.2 115.8
141 46.3 31.6 47.4



#15
2-Fluorobiphenyl
Concen: 0.377 ng
RT: 13.073 min Scan# 1312
Delta R.T. -0.020 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion:172 Resp: 4109
Ion Ratio Lower Upper
172 100
171 37.6 30.5 45.7
170 25.1 20.8 31.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 11
Delta R.T. -0.024 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.0	7.3	10.9



#27
Fluoranthene-d10
Concen: 0.441 ng
RT: 19.216 min Scan# 1933
Delta R.T. -0.022 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion	Ratio	Lower	Upper
212	100		
106	12.8	9.0	13.6
104	7.6	5.4	8.2









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.367 min Scan# 21
Delta R.T. -0.017 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116


Tgt Ion:	240	Resp:	4480
Ion Ratio	Lower	Upper	
240	100		
120	11.9	11.1	16.7
236	28.4	24.6	36.8



#31
Terphenyl-d14
Concen: 0.446 ng
RT: 19.815 min Scan# 2062
Delta R.T. -0.022 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion:	244	Resp:	3985
Ion Ratio	Lower	Upper	
244	100		
212	10.6	10.1	15.1
122	15.8	12.2	18.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.090 ng
RT: 21.295 min Scan# 21
Delta R.T. -0.026 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03




Instrument :
BNA_N
ClientSampleId :
RW10A-20250116

Tgt Ion:149 Resp: 580
Ion Ratio Lower Upper
149 100
167 25.5 21.4 32.0
279 3.8 3.0 4.6



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.663 min Scan# 2801
Delta R.T. -0.031 min
Lab File: BN035997.D
Acq: 20 Jan 2025 14:03

Tgt Ion:264 Resp: 4572
Ion Ratio Lower Upper
264 100
260 26.7 21.7 32.5
265 65.1 33.9 50.9#





CALIBRATION SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN010225.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 02 15:39:17 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN035871.D 0.2 =BN035872.D 0.4 =BN035873.D 0.8 =BN035874.D 1.6 =BN035875.D 3.2 =BN035876.D 5.0 =BN035877.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.454	0.422	0.373	0.388	0.391	0.377	0.377	0.397	7.52
3)	n-Nitrosodimet...	0.707	0.674	0.676	0.690	0.722	0.690	0.692	0.693	2.45
4) S	2-Fluorophenol	1.031	1.009	0.952	0.958	0.997	0.956	0.968	0.981	3.13
5) S	Phenol-d6	1.351	1.255	1.180	1.197	1.215	1.163	1.170	1.219	5.44
6)	bis(2-Chloroet...	1.001	0.946	0.936	0.913	0.938	0.886	0.879	0.929	4.43
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.346	0.307	0.296	0.302	0.319	0.317	0.330	0.317	5.48
9)	Naphthalene	1.163	1.094	1.086	1.096	1.167	1.113	1.141	1.123	3.00
10)	Hexachlorobuta...	0.368	0.354	0.353	0.363	0.382	0.362	0.369	0.365	2.74
11) SURR2	Methylnaphth...	0.547	0.536	0.527	0.519	0.556	0.524	0.540	0.536	2.50
12)	2-Methylnaphth...	0.691	0.654	0.685	0.680	0.731	0.701	0.722	0.695	3.75
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.164	0.165	0.189	0.189	0.207	0.211	0.220	0.192	11.39
15) S	2-Fluorobiphenyl	1.776	1.675	1.708	1.765	1.823	1.779	1.762	1.755	2.79
16)	Acenaphthylene	1.890	1.766	1.819	1.839	1.962	1.948	1.963	1.884	4.15
17)	Acenaphthene	1.187	1.162	1.198	1.232	1.300	1.275	1.291	1.235	4.43
18)	Fluorene	1.341	1.270	1.307	1.298	1.419	1.444	1.432	1.359	5.28
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.058	0.066	0.067	0.069	0.076	0.076	0.074	0.069	9.51
21)	4-Bromophenyl-...	0.265	0.269	0.268	0.267	0.290	0.283	0.279	0.274	3.47
22)	Hexachlorobenzene	0.393	0.369	0.356	0.362	0.394	0.373	0.371	0.374	3.93
23)	Atrazine	0.169	0.191	0.176	0.171	0.198	0.190	0.193	0.184	6.36
24)	Pentachlorophenol	0.141	0.101	0.118	0.122	0.143	0.148	0.153	0.132	14.40
25)	Phenanthrene	1.131	1.132	1.142	1.144	1.228	1.193	1.200	1.167	3.36
26)	Anthracene	0.996	0.998	1.008	1.033	1.137	1.132	1.130	1.062	6.34
27) SURR	Fluoranthene-d10	0.988	0.952	0.978	0.959	1.028	1.010	1.035	0.993	3.28
28)	Fluoranthene	1.268	1.253	1.330	1.312	1.441	1.446	1.475	1.361	6.71
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.606	1.620	1.571	1.612	1.711	1.621	1.627	1.624	2.63
31) S	Terphenyl-d14	0.814	0.813	0.790	0.777	0.828	0.776	0.781	0.797	2.64
32)	Benzo(a)anthra...	1.379	1.382	1.344	1.407	1.461	1.427	1.466	1.410	3.19
33)	Chrysene	1.484	1.458	1.441	1.451	1.541	1.478	1.471	1.475	2.24
34)	Bis(2-ethylhex...	0.691	0.583	0.582	0.541	0.569	0.519	0.530	0.574	10.05
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN010225.M

36)	Indeno(1,2,3-c...	1.428	1.461	1.414	1.578	1.697	1.712	1.751	1.577	9.15
37)	Benzo(b)fluora...	1.337	1.304	1.294	1.351	1.466	1.420	1.447	1.374	5.06
38)	Benzo(k)fluora...	1.279	1.254	1.251	1.344	1.469	1.442	1.482	1.360	7.55
39) C	Benzo(a)pyrene	1.099	1.181	1.086	1.163	1.270	1.244	1.284	1.190	6.71
40)	Dibenzo(a,h)an...	1.145	1.152	1.106	1.251	1.363	1.365	1.402	1.255	9.75
41)	Benzo(g,h,i)pe...	1.335	1.316	1.245	1.407	1.490	1.501	1.530	1.403	7.72

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035871.D
 Acq On : 02 Jan 2025 11:28
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jan 02 15:48:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.836	152	1199	0.400	ng	0.00
7) Naphthalene-d8	10.629	136	2391	0.400	ng	0.00
13) Acenaphthene-d10	14.470	164	1196	0.400	ng	0.00
19) Phenanthrene-d10	17.208	188	2268	0.400	ng	0.00
29) Chrysene-d12	21.386	240	1828	0.400	ng	0.00
35) Perylene-d12	23.694	264	2148	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	309	0.105	ng	0.00
5) Phenol-d6	6.983	99	405	0.111	ng	0.00
8) Nitrobenzene-d5	8.974	82	207	0.109	ng	0.00
11) 2-Methylnaphthalene-d10	12.220	152	327	0.102	ng	0.00
14) 2,4,6-Tribromophenol	15.954	330	49	0.085	ng	0.00
15) 2-Fluorobiphenyl	13.091	172	531	0.101	ng	0.00
27) Fluoranthene-d10	19.236	212	560	0.099	ng	0.00
31) Terphenyl-d14	19.840	244	372	0.102	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.322	88	136	0.114	ng	# 93
3) n-Nitrosodimethylamine	3.632	42	212	0.102	ng	# 84
6) bis(2-Chloroethyl)ether	7.258	93	300	0.108	ng	95
9) Naphthalene	10.683	128	695	0.104	ng	# 81
10) Hexachlorobutadiene	10.981	225	220	0.101	ng	# 98
12) 2-Methylnaphthalene	12.291	142	413	0.099	ng	96
16) Acenaphthylene	14.182	152	565	0.100	ng	98
17) Acenaphthene	14.524	154	355	0.096	ng	97
18) Fluorene	15.507	166	401	0.099	ng	96
20) 4,6-Dinitro-2-methylph...	15.569	198	33	0.084	ng	# 33
21) 4-Bromophenyl-phenylether	16.401	248	150	0.096	ng	91
22) Hexachlorobenzene	16.525	284	223	0.105	ng	# 90
23) Atrazine	16.674	200	96	0.092	ng	# 78
24) Pentachlorophenol	16.860	266	80	0.107	ng	96
25) Phenanthrene	17.245	178	641	0.097	ng	94
26) Anthracene	17.332	178	565	0.094	ng	97
28) Fluoranthene	19.264	202	719	0.093	ng	98
30) Pyrene	19.626	202	734	0.099	ng	98
32) Benzo(a)anthracene	21.368	228	630	0.098	ng	# 87
33) Chrysene	21.422	228	678	0.101	ng	# 90
34) Bis(2-ethylhexyl)phtha...	21.324	149	316	0.121	ng	94
36) Indeno(1,2,3-cd)pyrene	26.047	276	767	0.091	ng	96
37) Benzo(b)fluoranthene	23.001	252	718	0.097	ng	# 69
38) Benzo(k)fluoranthene	23.045	252	687	0.094	ng	# 68
39) Benzo(a)pyrene	23.591	252	590	0.092	ng	# 61
40) Dibenzo(a,h)anthracene	26.071	278	615	0.091	ng	# 65
41) Benzo(g,h,i)perylene	26.764	276	717	0.095	ng	# 71


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035871.D
Acq On : 02 Jan 2025 11:28
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Jan 02 15:48:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.836 min Scan# 693
Delta R.T. -0.000 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 1199
Ion Ratio Lower Upper
152 100
150 157.3 117.8 176.6
115 66.5 51.0 76.4



#2
1,4-Dioxane
Concen: 0.114 ng
RT: 3.322 min Scan# 68
Delta R.T. 0.000 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion: 88 Resp: 136
Ion Ratio Lower Upper
88 100
43 29.4 32.7 49.1#
58 77.2 63.0 94.4





#3

n-Nitrosodimethylamine

Concen: 0.102 ng

RT: 3.632 min Scan# 111

Delta R.T. 0.000 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion: 42 Resp: 212

Ion Ratio Lower Upper

42 100

74 88.7 62.2 93.2

44 26.4 7.1 10.7#



#4

2-Fluorophenol

Concen: 0.105 ng

RT: 5.409 min Scan# 357

Delta R.T. 0.000 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion: 112 Resp: 309


Ion Ratio Lower Upper

112 100

64 65.0 48.2 72.2

63 40.1 30.0 45.0









#5
Phenol-d6
Concen: 0.111 ng
RT: 6.983 min Scan# 575
Delta R.T. 0.000 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 405
Ion Ratio Lower Upper
99 100
42 26.9 23.5 35.3
71 44.4 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.108 ng
RT: 7.258 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion: 93 Resp: 300
Ion Ratio Lower Upper
93 100
63 82.3 62.0 93.0
95 33.3 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.629 min Scan# 91
Delta R.T. 0.004 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	2391
Ion	Ratio	Lower	Upper
136	100		
137	14.3	10.6	15.8
54	11.7	9.8	14.6
68	8.3	6.6	10.0



#8
Nitrobenzene-d5
Concen: 0.109 ng
RT: 8.974 min Scan# 821
Delta R.T. -0.007 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion:	82	Resp:	207
Ion	Ratio	Lower	Upper
82	100		
128	50.6	36.9	55.3
54	74.4	50.4	75.6





#9

Naphthalene

Concen: 0.104 ng

RT: 10.683 min Scan# 981

Delta R.T. 0.004 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:128 Resp: 695

Ion Ratio Lower Upper

128 100

129 21.7 10.6 16.0#

127 23.1 12.8 19.2#



#10

Hexachlorobutadiene

Concen: 0.101 ng

RT: 10.981 min Scan# 1009

Delta R.T. 0.004 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:225 Resp: 220


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 62.7 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.102 ng
RT: 12.220 min Scan# 1198
Delta R.T. 0.003 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 327
Ion Ratio Lower Upper
152 100
151 21.1 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.099 ng
RT: 12.291 min Scan# 1213
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion:142 Resp: 413
Ion Ratio Lower Upper
142 100
141 89.6 71.9 107.9
115 56.6 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.470 min Scan# 1491
Delta R.T. 0.009 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 1196
Ion Ratio Lower Upper
164 100
162 101.6 83.1 124.7
160 54.0 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 0.085 ng
RT: 15.954 min Scan# 1624
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion:330 Resp: 49
Ion Ratio Lower Upper
330 100
332 106.1 77.2 115.8
141 71.4 31.6 47.4#





#15

2-Fluorobiphenyl

Concen: 0.101 ng

RT: 13.091 min Scan# 1362

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:172 Resp: 531

Ion Ratio Lower Upper

172 100

171 43.4 30.5 45.7

170 32.3 20.8 31.2#



#16

Acenaphthylene

Concen: 0.100 ng

RT: 14.182 min Scan# 1464

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:152 Resp: 565


Ion Ratio Lower Upper

152 100

151 21.8 16.3 24.5

153 13.3 10.6 15.8









#17
Acenaphthene
Concen: 0.096 ng
RT: 14.524 min Scan# 1496
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion	154	Resp	355
Ion Ratio	100		
Lower			
Upper			
153	116.6	90.5	135.7
152	62.8	48.6	73.0



#18
Fluorene
Concen: 0.099 ng
RT: 15.507 min Scan# 1588
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion	166	Resp	401
Ion Ratio <td>100</td> <td></td> <td></td>	100		
Lower			
Upper			
165	96.5	80.6	120.8
167	12.5	11.4	17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.208 min Scan# 1725

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:188 Resp: 2268

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.6 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.084 ng

RT: 15.569 min Scan# 1593

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:198 Resp: 33


Ion Ratio Lower Upper

198 100

51 143.8 64.3 96.5#

105 109.4 50.0 75.0#









#21
4-Bromophenyl-phenylether
Concen: 0.096 ng
RT: 16.401 min Scan# 1670
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:248	Resp:	150
Ion Ratio	Lower	Upper
248	100	
250	95.0	76.8 115.2
141	95.0	63.6 95.4



#22
Hexachlorobenzene
Concen: 0.105 ng
RT: 16.525 min Scan# 1670
Delta R.T. 0.010 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion:284	Resp:	223
Ion Ratio	Lower	Upper
284	100	
142	38.1	31.4 47.0
249	45.7	27.8 41.8









#23
Atrazine
Concen: 0.092 ng
RT: 16.674 min Scan# 1682
Delta R.T. 0.010 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:200 Resp: 96
Ion Ratio Lower Upper
200 100
173 56.3 35.4 53.0#
215 70.9 42.4 63.6#



#24
Pentachlorophenol
Concen: 0.107 ng
RT: 16.860 min Scan# 1697
Delta R.T. 0.010 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion:266 Resp: 80
Ion Ratio Lower Upper
266 100
264 63.7 49.9 74.9
268 67.5 50.2 75.2









#25
Phenanthrene
Concen: 0.097 ng
RT: 17.245 min Scan# 1728
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 641
Ion Ratio Lower Upper
178 100
176 23.2 15.9 23.9
179 17.8 12.9 19.3



#26
Anthracene
Concen: 0.094 ng
RT: 17.332 min Scan# 1735
Delta R.T. -0.002 min
Lab File: BN035871.D
Acq: 02 Jan 2025 11:28

Tgt Ion:178 Resp: 565
Ion Ratio Lower Upper
178 100
176 20.4 15.0 22.6
179 15.0 13.0 19.6





#27

Fluoranthene-d10

Concen: 0.099 ng

RT: 19.236 min Scan# 1986

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:212 Resp: 560

Ion Ratio Lower Upper

212 100

106 11.8 9.0 13.6

104 10.4 5.4 8.2#



#28

Fluoranthene

Concen: 0.093 ng

RT: 19.264 min Scan# 1992

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:202 Resp: 719


Ion Ratio Lower Upper

202 100

101 10.0 7.2 10.8

203 18.1 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.386 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:240 Resp: 1828

Ion Ratio Lower Upper

240 100

120 13.2 11.1 16.7

236 29.6 24.6 36.8



#30

Pyrene

Concen: 0.099 ng

RT: 19.626 min Scan# 2070

Delta R.T. -0.002 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:202 Resp: 734


Ion Ratio Lower Upper

202 100

200 21.4 17.0 25.6

203 20.2 14.6 22.0





#31

Terphenyl-d14

Concen: 0.102 ng

RT: 19.840 min Scan# 2115

Delta R.T. 0.003 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:244 Resp: 372

Ion Ratio Lower Upper

244 100

212 21.4 10.1 15.1#

122 22.9 12.2 18.4#



#32

Benzo(a)anthracene

Concen: 0.098 ng

RT: 21.368 min Scan# 2304

Delta R.T. -0.006 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:228 Resp: 630


Ion Ratio Lower Upper

228 100

226 35.6 22.7 34.1#

229 26.9 17.1 25.7#





#33

Chrysene

Concen: 0.101 ng

RT: 21.422 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 678

Ion Ratio Lower Upper

228 100

226 36.5 25.2 37.8

229 25.3 16.0 24.0#



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.121 ng

RT: 21.324 min Scan# 2299

Delta R.T. 0.003 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:149 Resp: 316


Ion Ratio Lower Upper

149 100

167 23.4 21.4 32.0

279 4.1 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.694 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:264 Resp: 2148

Ion Ratio Lower Upper

264 100

260 28.4 21.7 32.5

265 52.5 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.091 ng

RT: 26.047 min Scan# 3665

Delta R.T. -0.000 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:276 Resp: 767


Ion Ratio Lower Upper

276 100

138 26.9 18.9 28.3

277 25.4 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 0.097 ng

RT: 23.001 min Scan# 2623

Delta R.T. 0.003 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:252 Resp: 718

Ion Ratio Lower Upper

252 100

253 38.5 20.1 30.1#

125 28.1 9.9 14.9#



#38

Benzo(k)fluoranthene

Concen: 0.094 ng

RT: 23.045 min Scan# 2638

Delta R.T. -0.000 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:252 Resp: 687


Ion Ratio Lower Upper

252 100

253 39.6 20.5 30.7#

125 29.1 10.6 16.0#





#39

Benzo(a)pyrene

Concen: 0.092 ng

RT: 23.591 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:252 Resp: 590

Ion Ratio Lower Upper

252 100

253 45.5 21.5 32.3#

125 34.3 12.6 19.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.091 ng

RT: 26.071 min Scan# 3673

Delta R.T. -0.000 min

Lab File: BN035871.D

Acq: 02 Jan 2025 11:28




Tgt Ion:278 Resp: 615


Ion Ratio Lower Upper

278 100

139 34.0 15.9 23.9#

279 48.5 23.0 34.4#





#41

Benzo(g,h,i)perylene

Concen: 0.095 ng

RT: 26.764 min Scan# 3910

Delta R.T. -0.000 min

Lab File: BN035871.D


Acq: 02 Jan 2025 11:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:276 Resp: 717

Ion Ratio Lower Upper

276 100

277 41.4 22.8 34.2#

138 41.8 20.1 30.1#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035872.D
 Acq On : 02 Jan 2025 12:04
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Jan 02 15:49:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.835	152	1152	0.400	ng	0.00
7) Naphthalene-d8	10.629	136	2364	0.400	ng	0.00
13) Acenaphthene-d10	14.470	164	1186	0.400	ng	0.00
19) Phenanthrene-d10	17.208	188	2216	0.400	ng	0.00
29) Chrysene-d12	21.386	240	1774	0.400	ng	0.00
35) Perylene-d12	23.694	264	1987	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	581	0.206	ng	0.00
5) Phenol-d6	6.983	99	723	0.206	ng	0.00
8) Nitrobenzene-d5	8.974	82	363	0.194	ng	0.00
11) 2-Methylnaphthalene-d10	12.220	152	634	0.200	ng	0.00
14) 2,4,6-Tribromophenol	15.954	330	98	0.172	ng	0.00
15) 2-Fluorobiphenyl	13.091	172	993	0.191	ng	0.00
27) Fluoranthene-d10	19.236	212	1055	0.192	ng	0.00
31) Terphenyl-d14	19.840	244	721	0.204	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.322	88	243	0.212	ng	95
3) n-Nitrosodimethylamine	3.632	42	388	0.194	ng	# 86
6) bis(2-Chloroethyl)ether	7.258	93	545	0.204	ng	94
9) Naphthalene	10.682	128	1293	0.195	ng	# 93
10) Hexachlorobutadiene	10.981	225	418	0.194	ng	# 98
12) 2-Methylnaphthalene	12.291	142	773	0.188	ng	97
16) Acenaphthylene	14.182	152	1047	0.187	ng	98
17) Acenaphthene	14.524	154	689	0.188	ng	99
18) Fluorene	15.507	166	753	0.187	ng	98
20) 4,6-Dinitro-2-methylph...	15.582	198	73	0.190	ng	# 79
21) 4-Bromophenyl-phenylether	16.401	248	298	0.196	ng	# 89
22) Hexachlorobenzene	16.525	284	409	0.197	ng	99
23) Atrazine	16.674	200	212	0.208	ng	94
24) Pentachlorophenol	16.860	266	112	0.153	ng	90
25) Phenanthrene	17.245	178	1254	0.194	ng	99
26) Anthracene	17.344	178	1106	0.188	ng	99
28) Fluoranthene	19.268	202	1388	0.184	ng	99
30) Pyrene	19.626	202	1437	0.200	ng	99
32) Benzo(a)anthracene	21.368	228	1226	0.196	ng	97
33) Chrysene	21.422	228	1293	0.198	ng	95
34) Bis(2-ethylhexyl)phtha...	21.323	149	517	0.203	ng	97
36) Indeno(1,2,3-cd)pyrene	26.050	276	1452	0.185	ng	97
37) Benzo(b)fluoranthene	22.998	252	1296	0.190	ng	# 87
38) Benzo(k)fluoranthene	23.042	252	1246	0.184	ng	# 87
39) Benzo(a)pyrene	23.591	252	1173	0.199	ng	# 85
40) Dibenzo(a,h)anthracene	26.070	278	1145	0.184	ng	# 83
41) Benzo(g,h,i)perylene	26.766	276	1307	0.188	ng	# 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035872.D
Acq On : 02 Jan 2025 12:04
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Jan 02 15:49:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.835 min Scan# 693
Delta R.T. -0.001 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 1152
Ion Ratio Lower Upper
152 100
150 147.1 117.8 176.6
115 65.9 51.0 76.4



#2
1,4-Dioxane
Concen: 0.212 ng
RT: 3.322 min Scan# 68
Delta R.T. -0.000 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion: 88 Resp: 243
Ion Ratio Lower Upper
88 100
43 45.7 32.7 49.1
58 74.9 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 0.194 ng
RT: 3.632 min Scan# 111
Delta R.T. -0.000 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 388
Ion Ratio Lower Upper
42 100
74 88.7 62.2 93.2
44 17.3 7.1 10.7#



#4
2-Fluorophenol
Concen: 0.206 ng
RT: 5.409 min Scan# 357
Delta R.T. -0.000 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion: 112 Resp: 581
Ion Ratio Lower Upper
112 100
64 62.5 48.2 72.2
63 39.1 30.0 45.0









#5
Phenol-d6
Concen: 0.206 ng
RT: 6.983 min Scan# 575
Delta R.T. -0.000 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	99	Resp:	723
Ion	Ratio	Lower	Upper
99	100		
42	28.2	23.5	35.3
71	44.5	35.5	53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.204 ng
RT: 7.258 min Scan# 613
Delta R.T. -0.000 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:	93	Resp:	545
Ion	Ratio	Lower	Upper
93	100		
63	83.3	62.0	93.0
95	35.0	25.5	38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.629 min Scan# 91
Delta R.T. 0.003 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	2364
Ion	Ratio	Lower	Upper
136	100		
137	13.9	10.6	15.8
54	11.1	9.8	14.6
68	7.9	6.6	10.0



#8
Nitrobenzene-d5
Concen: 0.194 ng
RT: 8.974 min Scan# 821
Delta R.T. -0.007 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:	82	Resp:	363
Ion	Ratio	Lower	Upper
82	100		
128	43.3	36.9	55.3
54	72.6	50.4	75.6





#9

Naphthalene

Concen: 0.195 ng

RT: 10.682 min Scan# 981

Delta R.T. 0.003 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:128 Resp: 1293

Ion Ratio Lower Upper

128 100

129 16.3 10.6 16.0#

127 18.7 12.8 19.2



#10

Hexachlorobutadiene

Concen: 0.194 ng

RT: 10.981 min Scan# 1009

Delta R.T. 0.003 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:225 Resp: 418


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 62.9 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.200 ng
RT: 12.220 min Scan# 1199
Delta R.T. 0.003 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:152 Resp: 634
Ion Ratio Lower Upper
152 100
151 21.5 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.188 ng
RT: 12.291 min Scan# 1213
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:142 Resp: 773
Ion Ratio Lower Upper
142 100
141 92.0 71.9 107.9
115 51.8 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.470 min Scan# 1490
Delta R.T. 0.008 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 1186
Ion Ratio Lower Upper
164 100
162 100.5 83.1 124.7
160 53.0 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 0.172 ng
RT: 15.954 min Scan# 1624
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:330 Resp: 98
Ion Ratio Lower Upper
330 100
332 99.0 77.2 115.8
141 49.0 31.6 47.4#





#15

2-Fluorobiphenyl

Concen: 0.191 ng

RT: 13.091 min Scan# 1362

Delta R.T. -0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:172 Resp: 993

Ion Ratio Lower Upper

172 100

171 40.3 30.5 45.7

170 30.4 20.8 31.2



#16

Acenaphthylene

Concen: 0.187 ng

RT: 14.182 min Scan# 1464

Delta R.T. -0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:152 Resp: 1047


Ion Ratio Lower Upper

152 100

151 21.3 16.3 24.5

153 13.7 10.6 15.8









#17
Acenaphthene
Concen: 0.188 ng
RT: 14.524 min Scan# 1496
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 689
Ion Ratio Lower Upper
154 100
153 113.8 90.5 135.7
152 62.1 48.6 73.0



#18
Fluorene
Concen: 0.187 ng
RT: 15.507 min Scan# 1588
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:166 Resp: 753
Ion Ratio Lower Upper
166 100
165 98.4 80.6 120.8
167 14.9 11.4 17.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.208 min Scan# 1725
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.2	7.3	10.9



#20
4,6-Dinitro-2-methylphenol
Concen: 0.190 ng
RT: 15.582 min Scan# 1594
Delta R.T. 0.010 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion	Ratio	Lower	Upper
198	100		
51	101.2	64.3	96.5#
105	75.9	50.0	75.0#









#21
4-Bromophenyl-phenylether
Concen: 0.196 ng
RT: 16.401 min Scan# 1670
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:248 Resp: 298
Ion Ratio Lower Upper
248 100
250 92.4 76.8 115.2
141 97.8 63.6 95.4#



#22
Hexachlorobenzene
Concen: 0.197 ng
RT: 16.525 min Scan# 1670
Delta R.T. 0.010 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:284 Resp: 409
Ion Ratio Lower Upper
284 100
142 38.6 31.4 47.0
249 35.2 27.8 41.8









#23
Atrazine
Concen: 0.208 ng
RT: 16.674 min Scan# 1682
Delta R.T. 0.010 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 212
Ion Ratio Lower Upper
200 100
173 40.8 35.4 53.0
215 57.5 42.4 63.6



#24
Pentachlorophenol
Concen: 0.153 ng
RT: 16.860 min Scan# 1697
Delta R.T. 0.010 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:266 Resp: 112
Ion Ratio Lower Upper
266 100
264 65.2 49.9 74.9
268 75.0 50.2 75.2









#25
Phenanthrene
Concen: 0.194 ng
RT: 17.245 min Scan# 1728
Delta R.T. -0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 1254
Ion Ratio Lower Upper
178 100
176 19.5 15.9 23.9
179 16.0 12.9 19.3



#26
Anthracene
Concen: 0.188 ng
RT: 17.344 min Scan# 1736
Delta R.T. 0.010 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:178 Resp: 1106
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.5 13.0 19.6





#27

Fluoranthene-d10

Concen: 0.192 ng

RT: 19.236 min Scan# 1986

Delta R.T. -0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:212 Resp: 1055

Ion Ratio Lower Upper

212 100

106 11.9 9.0 13.6

104 8.0 5.4 8.2



#28

Fluoranthene

Concen: 0.184 ng

RT: 19.268 min Scan# 1993

Delta R.T. 0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:202 Resp: 1388


Ion Ratio Lower Upper

202 100

101 9.1 7.2 10.8

203 17.7 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.386 min Scan# 21

Delta R.T. 0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:240 Resp: 1774

Ion Ratio Lower Upper

240 100

120 14.2 11.1 16.7

236 30.2 24.6 36.8



#30

Pyrene

Concen: 0.200 ng

RT: 19.626 min Scan# 2070

Delta R.T. -0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:202 Resp: 1437


Ion Ratio Lower Upper

202 100

200 22.0 17.0 25.6

203 18.6 14.6 22.0









#31
Terphenyl-d14
Concen: 0.204 ng
RT: 19.840 min Scan# 2115
Delta R.T. 0.002 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:244 Resp: 721
Ion Ratio Lower Upper
244 100
212 15.6 10.1 15.1#
122 18.0 12.2 18.4



#32
Benzo(a)anthracene
Concen: 0.196 ng
RT: 21.368 min Scan# 2304
Delta R.T. -0.007 min
Lab File: BN035872.D
Acq: 02 Jan 2025 12:04

Tgt Ion:228 Resp: 1226
Ion Ratio Lower Upper
228 100
226 30.6 22.7 34.1
229 22.5 17.1 25.7





#33

Chrysene

Concen: 0.198 ng

RT: 21.422 min Scan# 21

Delta R.T. 0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:228 Resp: 1293

Ion Ratio Lower Upper

228 100

226 33.2 25.2 37.8

229 23.9 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.203 ng

RT: 21.323 min Scan# 2299

Delta R.T. 0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:149 Resp: 517


Ion Ratio Lower Upper

149 100

167 28.2 21.4 32.0

279 3.9 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.694 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 1987

Ion Ratio Lower Upper

264 100

260 28.2 21.7 32.5

265 51.2 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.185 ng

RT: 26.050 min Scan# 3666

Delta R.T. 0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:276 Resp: 1452


Ion Ratio Lower Upper

276 100

138 24.9 18.9 28.3

277 25.6 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 0.190 ng

RT: 22.998 min Scan# 2622

Delta R.T. -0.000 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 1296

Ion Ratio Lower Upper

252 100

253 30.8 20.1 30.1#

125 19.2 9.9 14.9#



#38

Benzo(k)fluoranthene

Concen: 0.184 ng

RT: 23.042 min Scan# 2637

Delta R.T. -0.003 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:252 Resp: 1246


Ion Ratio Lower Upper

252 100

253 31.1 20.5 30.7#

125 20.5 10.6 16.0#





#39

Benzo(a)pyrene

Concen: 0.199 ng

RT: 23.591 min Scan# 21

Delta R.T. 0.002 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 1173

Ion Ratio Lower Upper

252 100

253 32.8 21.5 32.3#

125 24.2 12.6 19.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.184 ng

RT: 26.070 min Scan# 3673

Delta R.T. -0.000 min

Lab File: BN035872.D

Acq: 02 Jan 2025 12:04




Tgt Ion:278 Resp: 1145


Ion Ratio Lower Upper

278 100

139 27.2 15.9 23.9#

279 38.6 23.0 34.4#





#41

Benzo(g,h,i)perylene

Concen: 0.188 ng

RT: 26.766 min Scan# 3910

Delta R.T. 0.002 min

Lab File: BN035872.D


Acq: 02 Jan 2025 12:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 1307

Ion Ratio Lower Upper

276 100

277 34.3 22.8 34.2#

138 32.3 20.1 30.1#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035873.D
 Acq On : 02 Jan 2025 12:40
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jan 02 15:49:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.836	152	1261	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	2582	0.400	ng	0.00
13) Acenaphthene-d10	14.462	164	1358	0.400	ng	0.00
19) Phenanthrene-d10	17.210	188	2769	0.400	ng	0.00
29) Chrysene-d12	21.384	240	2381	0.400	ng	0.00
35) Perylene-d12	23.694	264	2623	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	1200	0.388	ng	0.00
5) Phenol-d6	6.983	99	1488	0.387	ng	0.00
8) Nitrobenzene-d5	8.981	82	764	0.374	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	1360	0.393	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	256	0.393	ng	0.00
15) 2-Fluorobiphenyl	13.093	172	2320	0.389	ng	0.00
27) Fluoranthene-d10	19.238	212	2707	0.394	ng	0.00
31) Terphenyl-d14	19.837	244	1882	0.397	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.322	88	470	0.375	ng	100
3) n-Nitrosodimethylamine	3.632	42	853	0.390	ng	100
6) bis(2-Chloroethyl)ether	7.258	93	1180	0.403	ng	100
9) Naphthalene	10.679	128	2805	0.387	ng	100
10) Hexachlorobutadiene	10.978	225	912	0.388	ng	# 100
12) 2-Methylnaphthalene	12.293	142	1768	0.394	ng	100
16) Acenaphthylene	14.184	152	2470	0.386	ng	100
17) Acenaphthene	14.526	154	1627	0.388	ng	100
18) Fluorene	15.509	166	1775	0.385	ng	100
20) 4,6-Dinitro-2-methylph...	15.572	198	185	0.385	ng	100
21) 4-Bromophenyl-phenylether	16.403	248	743	0.391	ng	100
22) Hexachlorobenzene	16.515	284	986	0.381	ng	100
23) Atrazine	16.664	200	486	0.382	ng	100
24) Pentachlorophenol	16.850	266	327	0.357	ng	100
25) Phenanthrene	17.247	178	3161	0.391	ng	100
26) Anthracene	17.334	178	2792	0.380	ng	100
28) Fluoranthene	19.266	202	3684	0.391	ng	100
30) Pyrene	19.628	202	3740	0.387	ng	100
32) Benzo(a)anthracene	21.375	228	3201	0.381	ng	100
33) Chrysene	21.420	228	3430	0.391	ng	100
34) Bis(2-ethylhexyl)phtha...	21.321	149	1386	0.406	ng	100
36) Indeno(1,2,3-cd)pyrene	26.048	276	3709	0.359	ng	100
37) Benzo(b)fluoranthene	22.998	252	3395	0.377	ng	100
38) Benzo(k)fluoranthene	23.045	252	3282	0.368	ng	100
39) Benzo(a)pyrene	23.589	252	2848	0.365	ng	100
40) Dibenzo(a,h)anthracene	26.071	278	2902	0.353	ng	100
41) Benzo(g,h,i)perylene	26.764	276	3265	0.355	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035873.D
 Acq On : 02 Jan 2025 12:40
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jan 02 15:49:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.836 min Scan# 693
Delta R.T. -0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 152 Resp: 1261
Ion Ratio Lower Upper
152 100
150 147.2 117.8 176.6
115 63.7 51.0 76.4



#2
1,4-Dioxane
Concen: 0.375 ng
RT: 3.322 min Scan# 68
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion: 88 Resp: 470
Ion Ratio Lower Upper
88 100
43 40.9 32.7 49.1
58 78.7 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 0.390 ng
RT: 3.632 min Scan# 111
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 853
Ion Ratio Lower Upper
42 100
74 77.7 62.2 93.2
44 8.9 7.1 10.7



#4
2-Fluorophenol
Concen: 0.388 ng
RT: 5.409 min Scan# 357
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion: 112 Resp: 1200
Ion Ratio Lower Upper
112 100
64 60.2 48.2 72.2
63 37.5 30.0 45.0









#5
Phenol-d6
Concen: 0.387 ng
RT: 6.983 min Scan# 575
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 1488
Ion Ratio Lower Upper
99 100
42 29.4 23.5 35.3
71 44.4 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.403 ng
RT: 7.258 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion: 93 Resp: 1180
Ion Ratio Lower Upper
93 100
63 77.5 62.0 93.0
95 31.9 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	2582
Ion	Ratio	Lower	Upper
136	100		
137	13.2	10.6	15.8
54	12.2	9.8	14.6
68	8.3	6.6	10.0



#8
Nitrobenzene-d5
Concen: 0.374 ng
RT: 8.981 min Scan# 822
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:	82	Resp:	764
Ion	Ratio	Lower	Upper
82	100		
128	46.1	36.9	55.3
54	63.0	50.4	75.6









#9
Naphthalene
Concen: 0.387 ng
RT: 10.679 min Scan# 981
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	2805
Ion	Ratio	Lower	Upper
128	100		
129	13.3	10.6	16.0
127	16.0	12.8	19.2



#10
Hexachlorobutadiene
Concen: 0.388 ng
RT: 10.978 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:	225	Resp:	912
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.4	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.393 ng
RT: 12.217 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 1360
Ion Ratio Lower Upper
152 100
151 21.3 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.394 ng
RT: 12.293 min Scan# 1213
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:142 Resp: 1768
Ion Ratio Lower Upper
142 100
141 89.9 71.9 107.9
115 49.5 39.6 59.4





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.462 min Scan# 1490

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :





BNA_N

ClientSampleId :

SSTDICCC0.4

Tgt Ion:164 Resp: 1358

Ion	Ratio	Lower	Upper
164	100		
162	103.9	83.1	124.7
160	57.5	46.0	69.0



#14

2,4,6-Tribromophenol

Concen: 0.393 ng

RT: 15.956 min Scan# 1624




Delta R.T. 0.000 min


Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Tgt Ion:330 Resp: 256

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.2	115.8
141	39.5	31.6	47.4





#15

2-Fluorobiphenyl

Concen: 0.389 ng

RT: 13.093 min Scan# 1362

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:172 Resp: 2320

Ion Ratio Lower Upper

172 100

171 38.1 30.5 45.7

170 26.0 20.8 31.2



#16

Acenaphthylene

Concen: 0.386 ng

RT: 14.184 min Scan# 1464

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:152 Resp: 2470


Ion Ratio Lower Upper

152 100

151 20.4 16.3 24.5

153 13.2 10.6 15.8









#17
Acenaphthene
Concen: 0.388 ng
RT: 14.526 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 1627
Ion Ratio Lower Upper
154 100
153 113.1 90.5 135.7
152 60.8 48.6 73.0



#18
Fluorene
Concen: 0.385 ng
RT: 15.509 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:166 Resp: 1775
Ion Ratio Lower Upper
166 100
165 100.7 80.6 120.8
167 14.2 11.4 17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.210 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 2769

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.1 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.385 ng

RT: 15.572 min Scan# 1593

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:198 Resp: 185


Ion Ratio Lower Upper

198 100

51 80.4 64.3 96.5

105 62.5 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.391 ng
RT: 16.403 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	743
Ion Ratio	Lower	Upper	
248	100		
250	96.0	76.8	115.2
141	79.5	63.6	95.4



#22
Hexachlorobenzene
Concen: 0.381 ng
RT: 16.515 min Scan# 1669
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:	284	Resp:	986
Ion Ratio	Lower	Upper	
284	100		
142	39.2	31.4	47.0
249	34.8	27.8	41.8





#23

Atrazine

Concen: 0.382 ng

RT: 16.664 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:200 Resp: 486

Ion Ratio Lower Upper

200 100

173 44.2 35.4 53.0

215 53.0 42.4 63.6



#24

Pentachlorophenol

Concen: 0.357 ng

RT: 16.850 min Scan# 1696

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:266 Resp: 327


Ion Ratio Lower Upper

266 100

264 62.4 49.9 74.9

268 62.7 50.2 75.2









#25
Phenanthrene
Concen: 0.391 ng
RT: 17.247 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 3161
Ion Ratio Lower Upper
178 100
176 19.9 15.9 23.9
179 16.1 12.9 19.3



#26
Anthracene
Concen: 0.380 ng
RT: 17.334 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:178 Resp: 2792
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 16.3 13.0 19.6





#27

Fluoranthene-d10

Concen: 0.394 ng

RT: 19.238 min Scan# 1986

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 2707

Ion Ratio Lower Upper

212 100

106 11.3 9.0 13.6

104 6.8 5.4 8.2



#28

Fluoranthene

Concen: 0.391 ng

RT: 19.266 min Scan# 1992

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:202 Resp: 3684


Ion Ratio Lower Upper

202 100

101 9.0 7.2 10.8

203 17.4 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 2381

Ion Ratio Lower Upper

240 100

120 13.9 11.1 16.7

236 30.7 24.6 36.8



#30

Pyrene

Concen: 0.387 ng

RT: 19.628 min Scan# 2070

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:202 Resp: 3740


Ion Ratio Lower Upper

202 100

200 21.3 17.0 25.6

203 18.3 14.6 22.0









#31
Terphenyl-d14
Concen: 0.397 ng
RT: 19.837 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:244	Resp:	1882
Ion Ratio	Lower	Upper
244	100	
212	12.6	10.1 15.1
122	15.3	12.2 18.4



#32
Benzo(a)anthracene
Concen: 0.381 ng
RT: 21.375 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BN035873.D
Acq: 02 Jan 2025 12:40

Tgt Ion:228	Resp:	3201
Ion Ratio	Lower	Upper
228	100	
226	28.4	22.7 34.1
229	21.4	17.1 25.7





#33

Chrysene

Concen: 0.391 ng

RT: 21.420 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:228 Resp: 3430

Ion Ratio Lower Upper

228 100

226 31.5 25.2 37.8

229 20.0 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.406 ng

RT: 21.321 min Scan# 2298

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:149 Resp: 1386


Ion Ratio Lower Upper

149 100

167 26.7 21.4 32.0

279 3.8 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.694 min Scan# 2860

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:264 Resp: 2623

Ion Ratio Lower Upper

264 100

260 27.1 21.7 32.5

265 42.4 33.9 50.9



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.359 ng

RT: 26.048 min Scan# 3665

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:276 Resp: 3709


Ion Ratio Lower Upper

276 100

138 23.6 18.9 28.3

277 24.4 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 0.377 ng

RT: 22.998 min Scan# 2622

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4




Tgt Ion:252 Resp: 3395

Ion Ratio Lower Upper

252 100

253 25.1 20.1 30.1

125 12.4 9.9 14.9



#38

Benzo(k)fluoranthene

Concen: 0.368 ng

RT: 23.045 min Scan# 2638

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:252 Resp: 3282


Ion Ratio Lower Upper

252 100

253 25.6 20.5 30.7

125 13.3 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.365 ng

RT: 23.589 min Scan# 2824

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 2848

Ion Ratio Lower Upper

252 100

253 26.9 21.5 32.3

125 15.8 12.6 19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.353 ng

RT: 26.071 min Scan# 3673

Delta R.T. 0.000 min

Lab File: BN035873.D

Acq: 02 Jan 2025 12:40




Tgt Ion:278 Resp: 2902


Ion Ratio Lower Upper

278 100

139 19.9 15.9 23.9

279 28.7 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 0.355 ng

RT: 26.764 min Scan# 3910

Delta R.T. 0.000 min

Lab File: BN035873.D


Acq: 02 Jan 2025 12:40

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 3265

Ion Ratio Lower Upper

276 100

277 28.5 22.8 34.2

138 25.1 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035874.D
 Acq On : 02 Jan 2025 13:16
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Jan 02 15:50:14 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.835	152	1005	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	2000	0.400	ng	0.00
13) Acenaphthene-d10	14.462	164	985	0.400	ng	0.00
19) Phenanthrene-d10	17.210	188	1825	0.400	ng	0.00
29) Chrysene-d12	21.384	240	1508	0.400	ng	# 0.00
35) Perylene-d12	23.694	264	1682	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	1925	0.781	ng	0.00
5) Phenol-d6	6.983	99	2406	0.786	ng	0.00
8) Nitrobenzene-d5	8.981	82	1207	0.762	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	2075	0.775	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	373	0.789	ng	0.00
15) 2-Fluorobiphenyl	13.093	172	3477	0.804	ng	0.00
27) Fluoranthene-d10	19.238	212	3502	0.773	ng	0.00
31) Terphenyl-d14	19.842	244	2344	0.780	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.322	88	779	0.780	ng	Qvalue 99
3) n-Nitrosodimethylamine	3.625	42	1387	0.796	ng	98
6) bis(2-Chloroethyl)ether	7.258	93	1835	0.787	ng	99
9) Naphthalene	10.679	128	4384	0.781	ng	98
10) Hexachlorobutadiene	10.978	225	1454	0.798	ng	# 99
12) 2-Methylnaphthalene	12.293	142	2720	0.783	ng	99
16) Acenaphthylene	14.184	152	3623	0.781	ng	99
17) Acenaphthene	14.526	154	2427	0.798	ng	97
18) Fluorene	15.509	166	2558	0.764	ng	99
20) 4,6-Dinitro-2-methylph...	15.571	198	251	0.792	ng	85
21) 4-Bromophenyl-phenylether	16.403	248	973	0.778	ng	97
22) Hexachlorobenzene	16.515	284	1320	0.774	ng	99
23) Atrazine	16.676	200	624	0.743	ng	# 87
24) Pentachlorophenol	16.850	266	444	0.735	ng	97
25) Phenanthrene	17.247	178	4177	0.784	ng	99
26) Anthracene	17.334	178	3772	0.778	ng	98
28) Fluoranthene	19.266	202	4790	0.772	ng	100
30) Pyrene	19.628	202	4863	0.794	ng	100
32) Benzo(a)anthracene	21.375	228	4245	0.799	ng	99
33) Chrysene	21.419	228	4376	0.787	ng	99
34) Bis(2-ethylhexyl)phtha...	21.330	149	1633	0.755	ng	99
36) Indeno(1,2,3-cd)pyrene	26.050	276	5308	0.800	ng	99
37) Benzo(b)fluoranthene	22.998	252	4545	0.787	ng	99
38) Benzo(k)fluoranthene	23.045	252	4521	0.790	ng	97
39) Benzo(a)pyrene	23.595	252	3913	0.782	ng	96
40) Dibenzo(a,h)anthracene	26.071	278	4209	0.798	ng	98
41) Benzo(g,h,i)perylene	26.767	276	4732	0.802	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035874.D
Acq On : 02 Jan 2025 13:16
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Jan 02 15:50:14 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.835 min Scan# 69
Delta R.T. -0.001 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 152 Resp: 1005
Ion Ratio Lower Upper
152 100
150 152.7 117.8 176.6
115 68.5 51.0 76.4



#2
1,4-Dioxane
Concen: 0.780 ng
RT: 3.322 min Scan# 68
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion: 88 Resp: 779
Ion Ratio Lower Upper
88 100
43 41.3 32.7 49.1
58 78.2 63.0 94.4





#3

n-Nitrosodimethylamine

Concen: 0.796 ng

RT: 3.625 min Scan# 111

Delta R.T. -0.007 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion: 42 Resp: 1387

Ion Ratio Lower Upper

42 100

74 76.2 62.2 93.2

44 9.6 7.1 10.7



#4

2-Fluorophenol

Concen: 0.781 ng

RT: 5.409 min Scan# 357

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion: 112 Resp: 1925


Ion Ratio Lower Upper

112 100

64 61.5 48.2 72.2

63 37.6 30.0 45.0









#5
Phenol-d6
Concen: 0.786 ng
RT: 6.983 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	99	Resp:	2406
Ion	Ratio	Lower	Upper
99	100		
42	29.4	23.5	35.3
71	44.2	35.5	53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.787 ng
RT: 7.258 min Scan# 613
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion:	93	Resp:	1835
Ion	Ratio	Lower	Upper
93	100		
63	78.4	62.0	93.0
95	33.1	25.5	38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	2000
Ion	Ratio	Lower	Upper
136	100		
137	14.5	10.6	15.8
54	13.7	9.8	14.6
68	8.4	6.6	10.0



#8
Nitrobenzene-d5
Concen: 0.762 ng
RT: 8.981 min Scan# 822
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion:	82	Resp:	1207
Ion	Ratio	Lower	Upper
82	100		
128	48.1	36.9	55.3
54	61.7	50.4	75.6





#9

Naphthalene

Concen: 0.781 ng

RT: 10.679 min Scan# 981

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:128 Resp: 4384

Ion Ratio Lower Upper

128 100

129 12.6 10.6 16.0

127 14.9 12.8 19.2



#10

Hexachlorobutadiene

Concen: 0.798 ng

RT: 10.978 min Scan# 1009

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:225 Resp: 1454


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.5 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.775 ng
RT: 12.217 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 2075
Ion Ratio Lower Upper
152 100
151 21.8 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.783 ng
RT: 12.293 min Scan# 1213
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion:142 Resp: 2720
Ion Ratio Lower Upper
142 100
141 89.4 71.9 107.9
115 48.0 39.6 59.4





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.462 min Scan# 1490

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:164 Resp: 985

Ion Ratio Lower Upper

164 100

162 106.9 83.1 124.7

160 59.3 46.0 69.0



#14

2,4,6-Tribromophenol

Concen: 0.789 ng

RT: 15.956 min Scan# 1624

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:330 Resp: 373


Ion Ratio Lower Upper

330 100

332 92.2 77.2 115.8

141 40.2 31.6 47.4





#15

2-Fluorobiphenyl

Concen: 0.804 ng

RT: 13.093 min Scan# 1362

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:172 Resp: 3477

Ion Ratio Lower Upper

172 100

171 37.1 30.5 45.7

170 25.5 20.8 31.2



#16

Acenaphthylene

Concen: 0.781 ng

RT: 14.184 min Scan# 1464

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:152 Resp: 3623


Ion Ratio Lower Upper

152 100

151 20.8 16.3 24.5

153 12.8 10.6 15.8





#17

Acenaphthene

Concen: 0.798 ng

RT: 14.526 min Scan# 1496

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:154 Resp: 2427

Ion Ratio Lower Upper

154 100

153 109.9 90.5 135.7

152 57.7 48.6 73.0



#18

Fluorene

Concen: 0.764 ng

RT: 15.509 min Scan# 1588

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:166 Resp: 2558


Ion Ratio Lower Upper

166 100

165 100.2 80.6 120.8

167 13.4 11.4 17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.210 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 1825

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.0 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.792 ng

RT: 15.571 min Scan# 1593

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:198 Resp: 251


Ion Ratio Lower Upper

198 100

51 64.4 64.3 96.5

105 53.9 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.778 ng
RT: 16.403 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.2	76.8	115.2
141	84.3	63.6	95.4



#22
Hexachlorobenzene
Concen: 0.774 ng
RT: 16.515 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion	Ratio	Lower	Upper
284	100		
142	39.6	31.4	47.0
249	35.0	27.8	41.8









#23
Atrazine
Concen: 0.743 ng
RT: 16.676 min Scan# 1682
Delta R.T. 0.012 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 624
Ion Ratio Lower Upper
200 100
173 27.3 35.4 53.0#
215 51.7 42.4 63.6



#24
Pentachlorophenol
Concen: 0.735 ng
RT: 16.850 min Scan# 1696
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion:266 Resp: 444
Ion Ratio Lower Upper
266 100
264 64.6 49.9 74.9
268 64.6 50.2 75.2





#25

Phenanthrene

Concen: 0.784 ng

RT: 17.247 min Scan# 1728

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:178 Resp: 4177

Ion Ratio Lower Upper

178 100

176 19.9 15.9 23.9

179 14.8 12.9 19.3



#26

Anthracene

Concen: 0.778 ng

RT: 17.334 min Scan# 1735

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:178 Resp: 3772


Ion Ratio Lower Upper

178 100

176 19.2 15.0 22.6

179 15.3 13.0 19.6





#27

Fluoranthene-d10

Concen: 0.773 ng

RT: 19.238 min Scan# 1986

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:212 Resp: 3502

Ion Ratio Lower Upper

212 100

106 11.3 9.0 13.6

104 6.8 5.4 8.2



#28

Fluoranthene

Concen: 0.772 ng

RT: 19.266 min Scan# 1992

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:202 Resp: 4790


Ion Ratio Lower Upper

202 100

101 8.8 7.2 10.8

203 17.4 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:240 Resp: 1508

Ion Ratio Lower Upper

240 100

120 16.8 11.1 16.7#

236 31.0 24.6 36.8



#30

Pyrene

Concen: 0.794 ng

RT: 19.628 min Scan# 2070

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:202 Resp: 4863


Ion Ratio Lower Upper

202 100

200 21.3 17.0 25.6

203 17.9 14.6 22.0









#31
Terphenyl-d14
Concen: 0.780 ng
RT: 19.842 min Scan# 2116
Delta R.T. 0.005 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
244	100		
212	12.4	10.1	15.1
122	13.2	12.2	18.4



#32
Benzo(a)anthracene
Concen: 0.799 ng
RT: 21.375 min Scan# 2304
Delta R.T. -0.000 min
Lab File: BN035874.D
Acq: 02 Jan 2025 13:16

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.4	22.7	34.1
229	20.7	17.1	25.7





#33

Chrysene

Concen: 0.787 ng

RT: 21.419 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:228 Resp: 4376

Ion Ratio Lower Upper

228 100

226 32.3 25.2 37.8

229 20.4 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.755 ng

RT: 21.330 min Scan# 2299

Delta R.T. 0.009 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:149 Resp: 1633


Ion Ratio Lower Upper

149 100

167 26.9 21.4 32.0

279 3.6 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.694 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:264 Resp: 1682

Ion Ratio Lower Upper

264 100

260 27.8 21.7 32.5

265 55.1 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.800 ng

RT: 26.050 min Scan# 3666

Delta R.T. 0.003 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:276 Resp: 5308


Ion Ratio Lower Upper

276 100

138 23.9 18.9 28.3

277 25.2 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 0.787 ng

RT: 22.998 min Scan# 2622

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 4545

Ion Ratio Lower Upper

252 100

253 24.2 20.1 30.1

125 12.1 9.9 14.9



#38

Benzo(k)fluoranthene

Concen: 0.790 ng

RT: 23.045 min Scan# 2638

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:252 Resp: 4521


Ion Ratio Lower Upper

252 100

253 24.3 20.5 30.7

125 12.3 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.782 ng

RT: 23.595 min Scan# 2826

Delta R.T. 0.006 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 3913

Ion Ratio Lower Upper

252 100

253 25.5 21.5 32.3

125 13.1 12.6 19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.798 ng

RT: 26.071 min Scan# 3673

Delta R.T. -0.000 min

Lab File: BN035874.D

Acq: 02 Jan 2025 13:16




Tgt Ion:278 Resp: 4209


Ion Ratio Lower Upper

278 100

139 20.1 15.9 23.9

279 27.4 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 0.802 ng

RT: 26.767 min Scan# 3911

Delta R.T. 0.003 min

Lab File: BN035874.D


Acq: 02 Jan 2025 13:16

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:276 Resp: 4732

Ion Ratio Lower Upper

276 100

277 26.3 22.8 34.2

138 23.0 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035875.D
 Acq On : 02 Jan 2025 13:52
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jan 02 15:50:39 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.835	152	1039	0.400	ng	0.00
7) Naphthalene-d8	10.625	136	2002	0.400	ng	0.00
13) Acenaphthene-d10	14.462	164	1024	0.400	ng	0.00
19) Phenanthrene-d10	17.210	188	1916	0.400	ng	0.00
29) Chrysene-d12	21.384	240	1622	0.400	ng	0.00
35) Perylene-d12	23.691	264	1727	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	4142	1.625	ng	0.00
5) Phenol-d6	6.983	99	5048	1.595	ng	0.00
8) Nitrobenzene-d5	8.981	82	2556	1.612	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	4454	1.662	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	847	1.723	ng	0.00
15) 2-Fluorobiphenyl	13.093	172	7465	1.661	ng	0.00
27) Fluoranthene-d10	19.238	212	7879	1.657	ng	0.00
31) Terphenyl-d14	19.837	244	5373	1.662	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.321	88	1627	1.577	ng	97
3) n-Nitrosodimethylamine	3.625	42	3002	1.667	ng	# 97
6) bis(2-Chloroethyl)ether	7.250	93	3900	1.617	ng	100
9) Naphthalene	10.679	128	9344	1.663	ng	96
10) Hexachlorobutadiene	10.978	225	3062	1.678	ng	# 99
12) 2-Methylnaphthalene	12.293	142	5850	1.682	ng	99
16) Acenaphthylene	14.184	152	8035	1.666	ng	100
17) Acenaphthene	14.526	154	5325	1.684	ng	97
18) Fluorene	15.509	166	5812	1.671	ng	97
20) 4,6-Dinitro-2-methylph...	15.571	198	585	1.758	ng	# 72
21) 4-Bromophenyl-phenylether	16.403	248	2219	1.689	ng	98
22) Hexachlorobenzene	16.515	284	3022	1.687	ng	98
23) Atrazine	16.664	200	1516	1.720	ng	# 87
24) Pentachlorophenol	16.850	266	1099	1.733	ng	99
25) Phenanthrene	17.247	178	9410	1.683	ng	99
26) Anthracene	17.334	178	8714	1.713	ng	99
28) Fluoranthene	19.266	202	11040	1.694	ng	99
30) Pyrene	19.628	202	11104	1.686	ng	99
32) Benzo(a)anthracene	21.375	228	9482	1.659	ng	98
33) Chrysene	21.419	228	10000	1.672	ng	99
34) Bis(2-ethylhexyl)phtha...	21.330	149	3691	1.586	ng	98
36) Indeno(1,2,3-cd)pyrene	26.056	276	11724	1.721	ng	99
37) Benzo(b)fluoranthene	22.998	252	10128	1.707	ng	95
38) Benzo(k)fluoranthene	23.045	252	10151	1.728	ng	# 94
39) Benzo(a)pyrene	23.592	252	8775	1.709	ng	# 92
40) Dibenzo(a,h)anthracene	26.071	278	9415	1.738	ng	94
41) Benzo(g,h,i)perylene	26.764	276	10294	1.699	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035875.D
Acq On : 02 Jan 2025 13:52
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Jan 02 15:50:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.835 min Scan# 69
Delta R.T. -0.001 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 1039
Ion Ratio Lower Upper
152 100
150 149.9 117.8 176.6
115 64.8 51.0 76.4



#2
1,4-Dioxane
Concen: 1.577 ng
RT: 3.321 min Scan# 68
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion: 88 Resp: 1627
Ion Ratio Lower Upper
88 100
43 44.4 32.7 49.1
58 77.4 63.0 94.4





#3

n-Nitrosodimethylamine

Concen: 1.667 ng

RT: 3.625 min Scan# 111

Delta R.T. -0.007 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion: 42 Resp: 3002

Ion Ratio Lower Upper

42 100

74 75.3 62.2 93.2

44 5.5 7.1 10.7#



#4

2-Fluorophenol

Concen: 1.625 ng

RT: 5.409 min Scan# 357

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion: 112 Resp: 4142


Ion Ratio Lower Upper

112 100

64 60.1 48.2 72.2

63 37.0 30.0 45.0









#5
Phenol-d6
Concen: 1.595 ng
RT: 6.983 min Scan# 575
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 5048
Ion Ratio Lower Upper
99 100
42 29.5 23.5 35.3
71 45.9 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 1.617 ng
RT: 7.250 min Scan# 612
Delta R.T. -0.007 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion: 93 Resp: 3900
Ion Ratio Lower Upper
93 100
63 77.8 62.0 93.0
95 32.1 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.625 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	2002
Ion	Ratio	Lower	Upper
136	100		
137	14.9	10.6	15.8
54	13.6	9.8	14.6
68	8.5	6.6	10.0



#8
Nitrobenzene-d5
Concen: 1.612 ng
RT: 8.981 min Scan# 822
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion:	82	Resp:	2556
Ion	Ratio	Lower	Upper
82	100		
128	46.2	36.9	55.3
54	59.3	50.4	75.6









#9
Naphthalene
Concen: 1.663 ng
RT: 10.679 min Scan# 981
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 9344
Ion Ratio Lower Upper
128 100
129 11.6 10.6 16.0
127 14.1 12.8 19.2



#10
Hexachlorobutadiene
Concen: 1.678 ng
RT: 10.978 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion:225 Resp: 3062
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 1.662 ng
RT: 12.217 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 4454
Ion Ratio Lower Upper
152 100
151 21.4 17.0 25.6



#12
2-Methylnaphthalene
Concen: 1.682 ng
RT: 12.293 min Scan# 1213
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion:142 Resp: 5850
Ion Ratio Lower Upper
142 100
141 90.9 71.9 107.9
115 48.1 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.462 min Scan# 1490
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 1024
Ion Ratio Lower Upper
164 100
162 107.2 83.1 124.7
160 57.4 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 1.723 ng
RT: 15.956 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion:330 Resp: 847
Ion Ratio Lower Upper
330 100
332 94.3 77.2 115.8
141 43.6 31.6 47.4





#15

2-Fluorobiphenyl

Concen: 1.661 ng

RT: 13.093 min Scan# 1362

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:172 Resp: 7465

Ion Ratio Lower Upper

172 100

171 36.0 30.5 45.7

170 24.9 20.8 31.2



#16

Acenaphthylene

Concen: 1.666 ng

RT: 14.184 min Scan# 1464

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:152 Resp: 8035


Ion Ratio Lower Upper

152 100

151 20.2 16.3 24.5

153 13.1 10.6 15.8









#17
Acenaphthene
Concen: 1.684 ng
RT: 14.526 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
154	100		
153	109.7	90.5	135.7
152	59.1	48.6	73.0



#18
Fluorene
Concen: 1.671 ng
RT: 15.509 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.0	80.6	120.8
167	13.1	11.4	17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.210 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 1916

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.0 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 1.758 ng

RT: 15.571 min Scan# 1593

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:198 Resp: 585


Ion Ratio Lower Upper

198 100

51 51.2 64.3 96.5#

105 45.1 50.0 75.0#









#21
4-Bromophenyl-phenylether
Concen: 1.689 ng
RT: 16.403 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
248	100		
250	94.7	76.8	115.2
141	77.8	63.6	95.4



#22
Hexachlorobenzene
Concen: 1.687 ng
RT: 16.515 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.1	31.4	47.0
249	33.9	27.8	41.8









#23
Atrazine
Concen: 1.720 ng
RT: 16.664 min Scan# 1681
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 1516
Ion Ratio Lower Upper
200 100
173 34.3 35.4 53.0#
215 44.8 42.4 63.6



#24
Pentachlorophenol
Concen: 1.733 ng
RT: 16.850 min Scan# 1696
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion:266 Resp: 1099
Ion Ratio Lower Upper
266 100
264 63.3 49.9 74.9
268 64.0 50.2 75.2









#25
Phenanthrene
Concen: 1.683 ng
RT: 17.247 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 9410
Ion Ratio Lower Upper
178 100
176 19.5 15.9 23.9
179 15.2 12.9 19.3



#26
Anthracene
Concen: 1.713 ng
RT: 17.334 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion:178 Resp: 8714
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 13.0 19.6





#27

Fluoranthene-d10

Concen: 1.657 ng

RT: 19.238 min Scan# 1986

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 7879

Ion Ratio Lower Upper

212 100

106 10.8 9.0 13.6

104 6.6 5.4 8.2



#28

Fluoranthene

Concen: 1.694 ng

RT: 19.266 min Scan# 1992

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:202 Resp: 11040


Ion Ratio Lower Upper

202 100

101 8.5 7.2 10.8

203 17.3 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 1622

Ion Ratio Lower Upper

240 100

120 16.2 11.1 16.7

236 31.7 24.6 36.8



#30

Pyrene

Concen: 1.686 ng

RT: 19.628 min Scan# 2070

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:202 Resp: 11104


Ion Ratio Lower Upper

202 100

200 21.3 17.0 25.6

203 17.8 14.6 22.0









#31
Terphenyl-d14
Concen: 1.662 ng
RT: 19.837 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.0	10.1	15.1
122	14.3	12.2	18.4



#32
Benzo(a)anthracene
Concen: 1.659 ng
RT: 21.375 min Scan# 2304
Delta R.T. -0.000 min
Lab File: BN035875.D
Acq: 02 Jan 2025 13:52

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.4	22.7	34.1
229	20.4	17.1	25.7





#33

Chrysene

Concen: 1.672 ng

RT: 21.419 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 10000

Ion Ratio Lower Upper

228 100

226 31.1 25.2 37.8

229 19.5 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.586 ng

RT: 21.330 min Scan# 2299

Delta R.T. 0.009 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:149 Resp: 3691


Ion Ratio Lower Upper

149 100

167 27.8 21.4 32.0

279 3.2 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.691 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:264 Resp: 1727

Ion Ratio Lower Upper

264 100

260 29.3 21.7 32.5

265 54.3 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.721 ng

RT: 26.056 min Scan# 3668

Delta R.T. 0.009 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:276 Resp: 11724


Ion Ratio Lower Upper

276 100

138 24.1 18.9 28.3

277 24.9 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 1.707 ng

RT: 22.998 min Scan# 2622

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 10128

Ion Ratio Lower Upper

252 100

253 22.7 20.1 30.1

125 10.0 9.9 14.9



#38

Benzo(k)fluoranthene

Concen: 1.728 ng

RT: 23.045 min Scan# 2638

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:252 Resp: 10151


Ion Ratio Lower Upper

252 100

253 23.0 20.5 30.7

125 10.6 10.6 16.0#





#39

Benzo(a)pyrene

Concen: 1.709 ng

RT: 23.592 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 8775

Ion Ratio Lower Upper

252 100

253 23.0 21.5 32.3

125 11.9 12.6 19.0#



#40

Dibenzo(a,h)anthracene

Concen: 1.738 ng

RT: 26.071 min Scan# 3673

Delta R.T. -0.000 min

Lab File: BN035875.D

Acq: 02 Jan 2025 13:52




Tgt Ion:278 Resp: 9415


Ion Ratio Lower Upper

278 100

139 17.2 15.9 23.9

279 25.4 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 1.699 ng

RT: 26.764 min Scan# 3910

Delta R.T. -0.000 min

Lab File: BN035875.D


Acq: 02 Jan 2025 13:52

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 10294

Ion Ratio Lower Upper

276 100

277 24.6 22.8 34.2

138 21.8 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035876.D
 Acq On : 02 Jan 2025 14:28
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jan 02 15:51:07 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.836	152	1084	0.400	ng	0.00
7) Naphthalene-d8	10.629	136	2065	0.400	ng	0.00
13) Acenaphthene-d10	14.470	164	1017	0.400	ng	0.00
19) Phenanthrene-d10	17.208	188	1903	0.400	ng	0.00
29) Chrysene-d12	21.386	240	1710	0.400	ng	0.00
35) Perylene-d12	23.691	264	1808	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	8292	3.118	ng	0.00
5) Phenol-d6	6.983	99	10087	3.054	ng	0.00
8) Nitrobenzene-d5	8.974	82	5234	3.201	ng	0.00
11) 2-Methylnaphthalene-d10	12.220	152	8654	3.130	ng	0.00
14) 2,4,6-Tribromophenol	15.954	330	1714	3.510	ng	0.00
15) 2-Fluorobiphenyl	13.091	172	14471	3.243	ng	0.00
27) Fluoranthene-d10	19.236	212	15378	3.255	ng	0.00
31) Terphenyl-d14	19.840	244	10611	3.114	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.322	88	3270	3.037	ng	97
3) n-Nitrosodimethylamine	3.618	42	5985	3.186	ng	# 95
6) bis(2-Chloroethyl)ether	7.251	93	7684	3.054	ng	100
9) Naphthalene	10.672	128	18395	3.173	ng	95
10) Hexachlorobutadiene	10.981	225	5988	3.182	ng	# 99
12) 2-Methylnaphthalene	12.291	142	11578	3.228	ng	99
16) Acenaphthylene	14.182	152	15847	3.309	ng	100
17) Acenaphthene	14.524	154	10376	3.304	ng	97
18) Fluorene	15.507	166	11752	3.402	ng	98
20) 4,6-Dinitro-2-methylph...	15.569	198	1163	3.519	ng	# 74
21) 4-Bromophenyl-phenylether	16.401	248	4303	3.298	ng	91
22) Hexachlorobenzene	16.525	284	5671	3.188	ng	99
23) Atrazine	16.674	200	2888	3.299	ng	# 78
24) Pentachlorophenol	16.860	266	2256	3.582	ng	99
25) Phenanthrene	17.245	178	18166	3.272	ng	99
26) Anthracene	17.332	178	17227	3.409	ng	98
28) Fluoranthene	19.268	202	22007	3.400	ng	99
30) Pyrene	19.626	202	22172	3.193	ng	99
32) Benzo(a)anthracene	21.368	228	19527	3.240	ng	98
33) Chrysene	21.422	228	20218	3.207	ng	98
34) Bis(2-ethylhexyl)phtha...	21.324	149	7099	2.894	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.050	276	24767	3.474	ng	99
37) Benzo(b)fluoranthene	22.998	252	20532	3.305	ng	# 93
38) Benzo(k)fluoranthene	23.045	252	20853	3.392	ng	# 92
39) Benzo(a)pyrene	23.591	252	17995	3.347	ng	# 90
40) Dibenzo(a,h)anthracene	26.073	278	19738	3.480	ng	92
41) Benzo(g,h,i)perylene	26.766	276	21709	3.423	ng	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035876.D
Acq On : 02 Jan 2025 14:28
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: Jan 02 15:51:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.836 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2





Tgt Ion: 152 Resp: 1084
Ion Ratio Lower Upper
152 100
150 152.1 117.8 176.6
115 64.4 51.0 76.4




#2
1,4-Dioxane
Concen: 3.037 ng
RT: 3.322 min Scan# 68
Delta R.T. 0.000 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28



Tgt Ion: 88 Resp: 3270
Ion Ratio Lower Upper
88 100
43 42.6 32.7 49.1
58 76.5 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 3.186 ng
RT: 3.618 min Scan# 109
Delta R.T. -0.014 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 5985
Ion Ratio Lower Upper
42 100
74 73.7 62.2 93.2
44 5.7 7.1 10.7#



#4
2-Fluorophenol
Concen: 3.118 ng
RT: 5.409 min Scan# 357
Delta R.T. 0.000 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion: 112 Resp: 8292
Ion Ratio Lower Upper
112 100
64 60.1 48.2 72.2
63 38.0 30.0 45.0









#5
Phenol-d6
Concen: 3.054 ng
RT: 6.983 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 10087
Ion Ratio Lower Upper
99 100
42 30.1 23.5 35.3
71 45.6 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 3.054 ng
RT: 7.251 min Scan# 612
Delta R.T. -0.007 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion: 93 Resp: 7684
Ion Ratio Lower Upper
93 100
63 77.6 62.0 93.0
95 31.9 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.629 min Scan# 91
Delta R.T. 0.003 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	2065
Ion	Ratio	Lower	Upper
136	100		
137	14.3	10.6	15.8
54	11.7	9.8	14.6
68	7.5	6.6	10.0



#8
Nitrobenzene-d5
Concen: 3.201 ng
RT: 8.974 min Scan# 821
Delta R.T. -0.007 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion:	82	Resp:	5234
Ion	Ratio	Lower	Upper
82	100		
128	36.6	36.9	55.3#
54	66.6	50.4	75.6









#9
Naphthalene
Concen: 3.173 ng
RT: 10.672 min Scan# 981
Delta R.T. -0.007 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.2	10.6	16.0
127	14.0	12.8	19.2



#10
Hexachlorobutadiene
Concen: 3.182 ng
RT: 10.981 min Scan# 1009
Delta R.T. 0.003 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 3.130 ng
RT: 12.220 min Scan# 1198
Delta R.T. 0.003 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 8654
Ion Ratio Lower Upper
152 100
151 21.5 17.0 25.6



#12
2-Methylnaphthalene
Concen: 3.228 ng
RT: 12.291 min Scan# 1213
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion:142 Resp: 11578
Ion Ratio Lower Upper
142 100
141 90.0 71.9 107.9
115 47.0 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.470 min Scan# 1490
Delta R.T. 0.009 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 1017
Ion Ratio Lower Upper
164 100
162 101.0 83.1 124.7
160 56.1 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 3.510 ng
RT: 15.954 min Scan# 1624
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion:330 Resp: 1714
Ion Ratio Lower Upper
330 100
332 93.9 77.2 115.8
141 41.7 31.6 47.4





#15

2-Fluorobiphenyl

Concen: 3.243 ng

RT: 13.091 min Scan# 1362

Delta R.T. -0.002 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:172 Resp: 14471

Ion Ratio Lower Upper

172 100

171 35.8 30.5 45.7

170 24.7 20.8 31.2



#16

Acenaphthylene

Concen: 3.309 ng

RT: 14.182 min Scan# 1464

Delta R.T. -0.002 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:152 Resp: 15847


Ion Ratio Lower Upper

152 100

151 20.3 16.3 24.5

153 13.1 10.6 15.8









#17
Acenaphthene
Concen: 3.304 ng
RT: 14.524 min Scan# 1496
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 10376
Ion Ratio Lower Upper
154 100
153 110.5 90.5 135.7
152 58.7 48.6 73.0



#18
Fluorene
Concen: 3.402 ng
RT: 15.507 min Scan# 1588
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion:166 Resp: 11752
Ion Ratio Lower Upper
166 100
165 99.1 80.6 120.8
167 12.9 11.4 17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.208 min Scan# 11

Delta R.T. -0.002 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 1903

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.6 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 3.519 ng

RT: 15.569 min Scan# 1593

Delta R.T. -0.002 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:198 Resp: 1163


Ion Ratio Lower Upper

198 100

51 51.4 64.3 96.5#

105 48.6 50.0 75.0#









#21
4-Bromophenyl-phenylether
Concen: 3.298 ng
RT: 16.401 min Scan# 1670
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	90.5	76.8	115.2
141	91.8	63.6	95.4



#22
Hexachlorobenzene
Concen: 3.188 ng
RT: 16.525 min Scan# 1670
Delta R.T. 0.010 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.4	31.4	47.0
249	34.7	27.8	41.8









#23
Atrazine
Concen: 3.299 ng
RT: 16.674 min Scan# 1682
Delta R.T. 0.010 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 2888
Ion Ratio Lower Upper
200 100
173 21.3 35.4 53.0#
215 45.1 42.4 63.6



#24
Pentachlorophenol
Concen: 3.582 ng
RT: 16.860 min Scan# 1697
Delta R.T. 0.010 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion:266 Resp: 2256
Ion Ratio Lower Upper
266 100
264 62.5 49.9 74.9
268 63.4 50.2 75.2









#25
Phenanthrene
Concen: 3.272 ng
RT: 17.245 min Scan# 1728
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.9	23.9
179	15.3	12.9	19.3



#26
Anthracene
Concen: 3.409 ng
RT: 17.332 min Scan# 1735
Delta R.T. -0.002 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.0	22.6
179	15.2	13.0	19.6





#27

Fluoranthene-d10

Concen: 3.255 ng

RT: 19.236 min Scan# 1986

Delta R.T. -0.002 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 15378

Ion Ratio Lower Upper

212 100

106 11.0 9.0 13.6

104 6.7 5.4 8.2



#28

Fluoranthene

Concen: 3.400 ng

RT: 19.268 min Scan# 1993

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:202 Resp: 22007


Ion Ratio Lower Upper

202 100

101 8.5 7.2 10.8

203 17.3 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.386 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 1710

Ion Ratio Lower Upper

240 100

120 12.8 11.1 16.7

236 30.0 24.6 36.8



#30

Pyrene

Concen: 3.193 ng

RT: 19.626 min Scan# 2070

Delta R.T. -0.002 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:202 Resp: 22172


Ion Ratio Lower Upper

202 100

200 21.4 17.0 25.6

203 17.6 14.6 22.0









#31
Terphenyl-d14
Concen: 3.114 ng
RT: 19.840 min Scan# 2115
Delta R.T. 0.003 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 10611
Ion Ratio Lower Upper
244 100
212 10.7 10.1 15.1
122 12.1 12.2 18.4#



#32
Benzo(a)anthracene
Concen: 3.240 ng
RT: 21.368 min Scan# 2304
Delta R.T. -0.006 min
Lab File: BN035876.D
Acq: 02 Jan 2025 14:28

Tgt Ion:228 Resp: 19527
Ion Ratio Lower Upper
228 100
226 27.9 22.7 34.1
229 19.4 17.1 25.7





#33

Chrysene

Concen: 3.207 ng

RT: 21.422 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 20218

Ion Ratio Lower Upper

228 100

226 30.6 25.2 37.8

229 19.3 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 2.894 ng

RT: 21.324 min Scan# 2299

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:149 Resp: 7099


Ion Ratio Lower Upper

149 100

167 27.8 21.4 32.0

279 2.9 3.0 4.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.691 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 1808

Ion Ratio Lower Upper

264 100

260 30.7 21.7 32.5

265 56.5 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.474 ng

RT: 26.050 min Scan# 3666

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:276 Resp: 24767


Ion Ratio Lower Upper

276 100

138 24.1 18.9 28.3

277 25.1 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 3.305 ng

RT: 22.998 min Scan# 2622

Delta R.T. -0.000 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 20532

Ion Ratio Lower Upper

252 100

253 22.0 20.1 30.1

125 9.5 9.9 14.9#



#38

Benzo(k)fluoranthene

Concen: 3.392 ng

RT: 23.045 min Scan# 2638

Delta R.T. -0.000 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:252 Resp: 20853


Ion Ratio Lower Upper

252 100

253 22.2 20.5 30.7

125 9.5 10.6 16.0#





#39

Benzo(a)pyrene

Concen: 3.347 ng

RT: 23.591 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 17995

Ion Ratio Lower Upper

252 100

253 22.0 21.5 32.3

125 11.2 12.6 19.0#



#40

Dibenzo(a,h)anthracene

Concen: 3.480 ng

RT: 26.073 min Scan# 3674

Delta R.T. 0.003 min

Lab File: BN035876.D

Acq: 02 Jan 2025 14:28




Tgt Ion:278 Resp: 19738


Ion Ratio Lower Upper

278 100

139 16.8 15.9 23.9

279 24.0 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 3.423 ng

RT: 26.766 min Scan# 3910

Delta R.T. 0.003 min

Lab File: BN035876.D


Acq: 02 Jan 2025 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 21709

Ion Ratio Lower Upper

276 100

277 23.9 22.8 34.2

138 21.3 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035877.D
 Acq On : 02 Jan 2025 15:04
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Jan 02 15:51:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.835	152	1152	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	2131	0.400	ng	0.00
13) Acenaphthene-d10	14.462	164	1092	0.400	ng	0.00
19) Phenanthrene-d10	17.210	188	2104	0.400	ng	0.00
29) Chrysene-d12	21.384	240	1949	0.400	ng	0.00
35) Perylene-d12	23.694	264	2023	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.409	112	13943	4.933	ng	0.00
5) Phenol-d6	6.983	99	16841	4.798	ng	0.00
8) Nitrobenzene-d5	8.981	82	8790	5.209	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	14382	5.041	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	3002	5.726	ng	0.00
15) 2-Fluorobiphenyl	13.093	172	24046	5.018	ng	0.00
27) Fluoranthene-d10	19.238	212	27229	5.214	ng	0.00
31) Terphenyl-d14	19.837	244	19033	4.901	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.322	88	5423	4.740	ng	98
3) n-Nitrosodimethylamine	3.618	42	9967	4.993	ng	# 97
6) bis(2-Chloroethyl)ether	7.250	93	12664	4.736	ng	99
9) Naphthalene	10.679	128	30381	5.079	ng	95
10) Hexachlorobutadiene	10.978	225	9822	5.057	ng	# 99
12) 2-Methylnaphthalene	12.293	142	19239	5.198	ng	98
16) Acenaphthylene	14.184	152	26794	5.210	ng	100
17) Acenaphthene	14.526	154	17619	5.226	ng	96
18) Fluorene	15.509	166	19548	5.269	ng	99
20) 4,6-Dinitro-2-methylph...	15.571	198	1944	5.320	ng	# 66
21) 4-Bromophenyl-phenylether	16.403	248	7340	5.088	ng	97
22) Hexachlorobenzene	16.515	284	9749	4.956	ng	99
23) Atrazine	16.676	200	5084	5.253	ng	# 77
24) Pentachlorophenol	16.850	266	4029	5.786	ng	98
25) Phenanthrene	17.247	178	31568	5.142	ng	99
26) Anthracene	17.334	178	29722	5.320	ng	98
28) Fluoranthene	19.266	202	38787	5.420	ng	99
30) Pyrene	19.628	202	39641	5.009	ng	99
32) Benzo(a)anthracene	21.375	228	35712	5.199	ng	97
33) Chrysene	21.419	228	35830	4.987	ng	99
34) Bis(2-ethylhexyl)phtha...	21.330	149	12924	4.623	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.050	276	44269	5.549	ng	98
37) Benzo(b)fluoranthene	22.998	252	36590	5.265	ng	# 93
38) Benzo(k)fluoranthene	23.045	252	37476	5.447	ng	# 92
39) Benzo(a)pyrene	23.592	252	32476	5.398	ng	# 89
40) Dibenzo(a,h)anthracene	26.071	278	35448	5.585	ng	92
41) Benzo(g,h,i)perylene	26.770	276	38678	5.450	ng	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035877.D
Acq On : 02 Jan 2025 15:04
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jan 02 15:51:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.835 min Scan# 69
Delta R.T. -0.001 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 1152
Ion Ratio Lower Upper
152 100
150 151.1 117.8 176.6
115 65.8 51.0 76.4



#2
1,4-Dioxane
Concen: 4.740 ng
RT: 3.322 min Scan# 68
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion: 88 Resp: 5423
Ion Ratio Lower Upper
88 100
43 42.6 32.7 49.1
58 77.3 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 4.993 ng
RT: 3.618 min Scan# 109
Delta R.T. -0.015 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 9967
Ion Ratio Lower Upper
42 100
74 75.3 62.2 93.2
44 5.7 7.1 10.7#



#4
2-Fluorophenol
Concen: 4.933 ng
RT: 5.409 min Scan# 357
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion: 112 Resp: 13943
Ion Ratio Lower Upper
112 100
64 61.0 48.2 72.2
63 38.4 30.0 45.0









#5
Phenol-d6
Concen: 4.798 ng
RT: 6.983 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 16841
Ion Ratio Lower Upper
99 100
42 29.5 23.5 35.3
71 45.5 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 4.736 ng
RT: 7.250 min Scan# 612
Delta R.T. -0.007 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion: 93 Resp: 12664
Ion Ratio Lower Upper
93 100
63 78.3 62.0 93.0
95 32.4 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	2131
Ion	Ratio	Lower	Upper
136	100		
137	14.2	10.6	15.8
54	13.0	9.8	14.6
68	9.3	6.6	10.0



#8
Nitrobenzene-d5
Concen: 5.209 ng
RT: 8.981 min Scan# 822
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:	82	Resp:	8790
Ion	Ratio	Lower	Upper
82	100		
128	43.7	36.9	55.3
54	58.5	50.4	75.6









#9
Naphthalene
Concen: 5.079 ng
RT: 10.679 min Scan# 981
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 30381
Ion Ratio Lower Upper
128 100
129 11.1 10.6 16.0
127 13.8 12.8 19.2



#10
Hexachlorobutadiene
Concen: 5.057 ng
RT: 10.978 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:225 Resp: 9822
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 5.041 ng
RT: 12.217 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 14382
Ion Ratio Lower Upper
152 100
151 21.3 17.0 25.6



#12
2-Methylnaphthalene
Concen: 5.198 ng
RT: 12.293 min Scan# 1213
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:142 Resp: 19239
Ion Ratio Lower Upper
142 100
141 89.7 71.9 107.9
115 46.5 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.462 min Scan# 1490
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 1092
Ion Ratio Lower Upper
164 100
162 107.2 83.1 124.7
160 59.0 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 5.726 ng
RT: 15.956 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:330 Resp: 3002
Ion Ratio Lower Upper
330 100
332 95.6 77.2 115.8
141 42.7 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 5.018 ng
RT: 13.093 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.3	30.5	45.7
170	24.4	20.8	31.2



#16
Acenaphthylene
Concen: 5.210 ng
RT: 14.184 min Scan# 1464
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.3	24.5
153	12.9	10.6	15.8









#17
Acenaphthene
Concen: 5.226 ng
RT: 14.526 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 17619
Ion Ratio Lower Upper
154 100
153 109.7 90.5 135.7
152 57.3 48.6 73.0



#18
Fluorene
Concen: 5.269 ng
RT: 15.509 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:166 Resp: 19548
Ion Ratio Lower Upper
166 100
165 99.6 80.6 120.8
167 12.9 11.4 17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.210 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 2104

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.5 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 5.320 ng

RT: 15.571 min Scan# 1593

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04




Tgt Ion:198 Resp: 1944


Ion Ratio Lower Upper

198 100

51 44.9 64.3 96.5#

105 42.5 50.0 75.0#









#21
4-Bromophenyl-phenylether
Concen: 5.088 ng
RT: 16.403 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.1	76.8	115.2
141	75.0	63.6	95.4



#22
Hexachlorobenzene
Concen: 4.956 ng
RT: 16.515 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion	Ratio	Lower	Upper
284	100		
142	37.8	31.4	47.0
249	34.7	27.8	41.8









#23
Atrazine
Concen: 5.253 ng
RT: 16.676 min Scan# 1682
Delta R.T. 0.012 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:200 Resp: 5084
Ion Ratio Lower Upper
200 100
173 19.6 35.4 53.0#
215 45.8 42.4 63.6



#24
Pentachlorophenol
Concen: 5.786 ng
RT: 16.850 min Scan# 1696
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:266 Resp: 4029
Ion Ratio Lower Upper
266 100
264 63.7 49.9 74.9
268 63.9 50.2 75.2









#25
Phenanthrene
Concen: 5.142 ng
RT: 17.247 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 31568
Ion Ratio Lower Upper
178 100
176 19.8 15.9 23.9
179 15.0 12.9 19.3



#26
Anthracene
Concen: 5.320 ng
RT: 17.334 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:178 Resp: 29722
Ion Ratio Lower Upper
178 100
176 19.2 15.0 22.6
179 15.2 13.0 19.6





#27

Fluoranthene-d10

Concen: 5.214 ng

RT: 19.238 min Scan# 1986

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:212 Resp: 27229

Ion Ratio Lower Upper

212 100

106 11.2 9.0 13.6

104 6.7 5.4 8.2



#28

Fluoranthene

Concen: 5.420 ng

RT: 19.266 min Scan# 1992

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04




Tgt Ion:202 Resp: 38787


Ion Ratio Lower Upper

202 100

101 8.7 7.2 10.8

203 17.2 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:240 Resp: 1949

Ion Ratio Lower Upper

240 100

120 16.1 11.1 16.7

236 30.7 24.6 36.8



#30

Pyrene

Concen: 5.009 ng

RT: 19.628 min Scan# 2070

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04




Tgt Ion:202 Resp: 39641


Ion Ratio Lower Upper

202 100

200 21.4 17.0 25.6

203 17.6 14.6 22.0









#31
Terphenyl-d14
Concen: 4.901 ng
RT: 19.837 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.3	10.1	15.1
122	13.1	12.2	18.4



#32
Benzo(a)anthracene
Concen: 5.199 ng
RT: 21.375 min Scan# 2304
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.9	22.7	34.1
229	19.8	17.1	25.7





#33

Chrysene

Concen: 4.987 ng

RT: 21.419 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 35830

Ion Ratio Lower Upper

228 100

226 31.3 25.2 37.8

229 19.2 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 4.623 ng

RT: 21.330 min Scan# 2299

Delta R.T. 0.009 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04




Tgt Ion:149 Resp: 12924


Ion Ratio Lower Upper

149 100

167 27.2 21.4 32.0

279 2.8 3.0 4.6#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.694 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:264 Resp: 2023
Ion Ratio Lower Upper
264 100
260 28.0 21.7 32.5
265 56.4 33.9 50.9#



#36
Indeno(1,2,3-cd)pyrene
Concen: 5.549 ng
RT: 26.050 min Scan# 3666
Delta R.T. 0.003 min
Lab File: BN035877.D
Acq: 02 Jan 2025 15:04

Tgt Ion:276 Resp: 44269
Ion Ratio Lower Upper
276 100
138 24.6 18.9 28.3
277 24.9 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 5.265 ng

RT: 22.998 min Scan# 2622

Delta R.T. -0.000 min

Lab File: BN035877.D


Acq: 02 Jan 2025 15:04

Instrument :



BNA_N

ClientSampleId :

SSTDICC5.0



Tgt Ion:252	Resp:	36590
Ion	Ratio	Lower Upper
252	100	
253	21.9	20.1 30.1
125	9.7	9.9 14.9



#38

Benzo(k)fluoranthene


Concen: 5.447 ng

RT: 23.045 min Scan# 2638



Delta R.T. -0.000 min


Lab File: BN035877.D

Acq: 02 Jan 2025 15:04



Tgt Ion:252	Resp:	37476
Ion	Ratio	Lower Upper
252	100	
253	21.9	20.5 30.7
125	9.5	10.6 16.0





#39

Benzo(a)pyrene

Concen: 5.398 ng

RT: 23.592 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 32476

Ion Ratio Lower Upper

252 100

253 21.4 21.5 32.3#

125 10.7 12.6 19.0#



#40

Dibenzo(a,h)anthracene

Concen: 5.585 ng

RT: 26.071 min Scan# 3673

Delta R.T. -0.000 min

Lab File: BN035877.D

Acq: 02 Jan 2025 15:04




Tgt Ion:278 Resp: 35448


Ion Ratio Lower Upper

278 100

139 17.0 15.9 23.9

279 23.7 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 5.450 ng

RT: 26.770 min Scan# 3912

Delta R.T. 0.006 min

Lab File: BN035877.D


Acq: 02 Jan 2025 15:04

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:276 Resp: 38678

Ion Ratio Lower Upper

276 100

277 23.7 22.8 34.2

138 21.5 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035878.D
 Acq On : 02 Jan 2025 15:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010225

Quant Time: Jan 02 16:21:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.835	152	1111	0.400	ng	0.00
7) Naphthalene-d8	10.625	136	2204	0.400	ng	0.00
13) Acenaphthene-d10	14.462	164	1019	0.400	ng	0.00
19) Phenanthrene-d10	17.210	188	1951	0.400	ng	0.00
29) Chrysene-d12	21.384	240	1648	0.400	ng	# 0.00
35) Perylene-d12	23.697	264	1887	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.401	112	933	0.342	ng	0.00
5) Phenol-d6	6.983	99	1123	0.332	ng	0.00
8) Nitrobenzene-d5	8.981	82	694	0.398	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	1148	0.389	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	171	0.350	ng	0.00
15) 2-Fluorobiphenyl	13.093	172	1929	0.431	ng	0.00
27) Fluoranthene-d10	19.238	212	1850	0.382	ng	0.00
31) Terphenyl-d14	19.837	244	1291	0.393	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.321	88	520	0.471	ng	97
3) n-Nitrosodimethylamine	3.625	42	806	0.419	ng	# 97
6) bis(2-Chloroethyl)ether	7.250	93	1074	0.416	ng	97
9) Naphthalene	10.679	128	2466	0.399	ng	99
10) Hexachlorobutadiene	10.978	225	778	0.387	ng	# 98
12) 2-Methylnaphthalene	12.293	142	1561	0.408	ng	98
16) Acenaphthylene	14.184	152	2167	0.452	ng	100
17) Acenaphthene	14.526	154	1341	0.426	ng	98
18) Fluorene	15.509	166	1412	0.408	ng	96
20) 4,6-Dinitro-2-methylph...	15.571	198	139	0.410	ng	97
21) 4-Bromophenyl-phenylether	16.403	248	532	0.398	ng	98
22) Hexachlorobenzene	16.515	284	743	0.407	ng	99
23) Atrazine	16.664	200	346	0.386	ng	97
24) Pentachlorophenol	16.862	266	253	0.392	ng	95
25) Phenanthrene	17.247	178	2381	0.418	ng	99
26) Anthracene	17.334	178	2196	0.424	ng	98
28) Fluoranthene	19.266	202	2530	0.381	ng	99
30) Pyrene	19.628	202	2613	0.391	ng	99
32) Benzo(a)anthracene	21.375	228	2353	0.405	ng	100
33) Chrysene	21.419	228	2440	0.402	ng	97
34) Bis(2-ethylhexyl)phtha...	21.330	149	930	0.393	ng	100
36) Indeno(1,2,3-cd)pyrene	26.050	276	2972	0.399	ng	98
37) Benzo(b)fluoranthene	23.001	252	2413	0.372	ng	# 94
38) Benzo(k)fluoranthene	23.048	252	2442	0.381	ng	96
39) Benzo(a)pyrene	23.594	252	2250	0.401	ng	97
40) Dibenzo(a,h)anthracene	26.074	278	2353	0.397	ng	98
41) Benzo(g,h,i)perylene	26.767	276	2378	0.359	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035878.D
Acq On : 02 Jan 2025 15:51
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN010225

Quant Time: Jan 02 16:21:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.835 min Scan# 693
Delta R.T. -0.001 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion:152 Resp: 1111
Ion Ratio Lower Upper
152 100
150 151.7 117.8 176.6
115 63.5 51.0 76.4



#2
1,4-Dioxane
Concen: 0.471 ng
RT: 3.321 min Scan# 68
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion: 88 Resp: 520
Ion Ratio Lower Upper
88 100
43 40.0 32.7 49.1
58 75.0 63.0 94.4





#3

n-Nitrosodimethylamine

Concen: 0.419 ng

RT: 3.625 min Scan# 111

Delta R.T. -0.007 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion: 42 Resp: 806

Ion Ratio Lower Upper

42 100

74 79.4 62.2 93.2

44 12.0 7.1 10.7#



#4

2-Fluorophenol

Concen: 0.342 ng

RT: 5.401 min Scan# 356

Delta R.T. -0.007 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion: 112 Resp: 933


Ion Ratio Lower Upper

112 100

64 60.5 48.2 72.2

63 36.9 30.0 45.0









#5
Phenol-d6
Concen: 0.332 ng
RT: 6.983 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion: 99 Resp: 1123
Ion Ratio Lower Upper
99 100
42 29.0 23.5 35.3
71 42.0 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.416 ng
RT: 7.250 min Scan# 612
Delta R.T. -0.007 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion: 93 Resp: 1074
Ion Ratio Lower Upper
93 100
63 80.4 62.0 93.0
95 33.2 25.5 38.3



#7
Naphthalene-d8

Concen: 0.400 ng

RT: 10.625 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225

Tgt Ion:136 Resp: 2204





Ion Ratio Lower Upper

136 100

137 14.0 10.6 15.8

54 13.1 9.8 14.6

68 8.4 6.6 10.0



#8

Nitrobenzene-d5

Concen: 0.398 ng

RT: 8.981 min Scan# 822

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion: 82 Resp: 694


Ion Ratio Lower Upper

82 100

128 48.5 36.9 55.3

54 59.7 50.4 75.6









#9
Naphthalene
Concen: 0.399 ng
RT: 10.679 min Scan# 981
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion:128 Resp: 2466
Ion Ratio Lower Upper
128 100
129 14.0 10.6 16.0
127 15.9 12.8 19.2



#10
Hexachlorobutadiene
Concen: 0.387 ng
RT: 10.978 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion:225 Resp: 778
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.1 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.389 ng
RT: 12.217 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion:152 Resp: 1148
Ion Ratio Lower Upper
152 100
151 21.7 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.408 ng
RT: 12.293 min Scan# 1213
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion:142 Resp: 1561
Ion Ratio Lower Upper
142 100
141 87.7 71.9 107.9
115 48.8 39.6 59.4





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.462 min Scan# 1490

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:164 Resp: 1019

Ion Ratio Lower Upper

164 100

162 108.5 83.1 124.7

160 62.0 46.0 69.0



#14

2,4,6-Tribromophenol

Concen: 0.350 ng

RT: 15.956 min Scan# 1624

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:330 Resp: 171


Ion Ratio Lower Upper

330 100

332 93.0 77.2 115.8

141 40.4 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 0.431 ng
RT: 13.093 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion:	172	Resp:	1929
Ion	Ratio	Lower	Upper
172	100		
171	37.3	30.5	45.7
170	26.0	20.8	31.2



#16
Acenaphthylene
Concen: 0.452 ng
RT: 14.184 min Scan# 1464
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion:	152	Resp:	2167
Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.3	24.5
153	13.0	10.6	15.8









#17
Acenaphthene
Concen: 0.426 ng
RT: 14.526 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion	154	Resp	1341
Ion Ratio	Lower	Upper	
154	100		
153	110.5	90.5	135.7
152	60.5	48.6	73.0



#18
Fluorene
Concen: 0.408 ng
RT: 15.509 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion	166	Resp	1412
Ion Ratio	Lower	Upper	
166	100		
165	96.0	80.6	120.8
167	14.0	11.4	17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.210 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:188 Resp: 1951

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.7 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.410 ng

RT: 15.571 min Scan# 1593

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:198 Resp: 139


Ion Ratio Lower Upper

198 100

51 82.4 64.3 96.5

105 66.2 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.398 ng
RT: 16.403 min Scan# 1660
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion:248 Resp: 532
Ion Ratio Lower Upper
248 100
250 93.9 76.8 115.2
141 80.5 63.6 95.4



#22
Hexachlorobenzene
Concen: 0.407 ng
RT: 16.515 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion:284 Resp: 743
Ion Ratio Lower Upper
284 100
142 38.1 31.4 47.0
249 35.3 27.8 41.8





#23

Atrazine

Concen: 0.386 ng

RT: 16.664 min Scan# 1681

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:200 Resp: 346

Ion Ratio Lower Upper

200 100

173 44.8 35.4 53.0

215 50.0 42.4 63.6



#24

Pentachlorophenol

Concen: 0.392 ng

RT: 16.862 min Scan# 1697

Delta R.T. 0.012 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:266 Resp: 253


Ion Ratio Lower Upper

266 100

264 56.1 49.9 74.9

268 60.9 50.2 75.2









#25
Phenanthrene
Concen: 0.418 ng
RT: 17.247 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.2	15.9	23.9
179	15.5	12.9	19.3



#26
Anthracene
Concen: 0.424 ng
RT: 17.334 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.0	22.6
179	15.2	13.0	19.6





#27

Fluoranthene-d10

Concen: 0.382 ng

RT: 19.238 min Scan# 1986

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:212 Resp: 1850

Ion Ratio Lower Upper

212 100

106 11.5 9.0 13.6

104 7.1 5.4 8.2



#28

Fluoranthene

Concen: 0.381 ng

RT: 19.266 min Scan# 1992

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:202 Resp: 2530


Ion Ratio Lower Upper

202 100

101 9.2 7.2 10.8

203 17.6 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N





ClientSampleId :

ICVBN010225

Tgt Ion:240 Resp: 1648

Ion Ratio Lower Upper

240	100		
120	19.6	11.1	16.7#
236	31.7	24.6	36.8



#30

Pyrene

Concen: 0.391 ng

RT: 19.628 min Scan# 2070

Delta R.T. -0.000 min




Lab File: BN035878.D


Acq: 02 Jan 2025 15:51

Tgt Ion:202 Resp: 2613

Ion Ratio Lower Upper

202	100		
200	21.6	17.0	25.6
203	18.5	14.6	22.0









#31
Terphenyl-d14
Concen: 0.393 ng
RT: 19.837 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51




Instrument :
BNA_N
ClientSampleId :
ICVBN010225


Tgt Ion	Ratio	Lower	Upper
244	100		
212	13.8	10.1	15.1
122	16.8	12.2	18.4



#32
Benzo(a)anthracene
Concen: 0.405 ng
RT: 21.375 min Scan# 2304
Delta R.T. -0.000 min
Lab File: BN035878.D
Acq: 02 Jan 2025 15:51

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.8	22.7	34.1
229	21.4	17.1	25.7





#33

Chrysene

Concen: 0.402 ng

RT: 21.419 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:228 Resp: 2440

Ion Ratio Lower Upper

228 100

226 33.4 25.2 37.8

229 20.8 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.393 ng

RT: 21.330 min Scan# 2299

Delta R.T. 0.009 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:149 Resp: 930


Ion Ratio Lower Upper

149 100

167 26.9 21.4 32.0

279 3.4 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.697 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:264 Resp: 1887

Ion Ratio Lower Upper

264 100

260 28.0 21.7 32.5

265 54.3 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.399 ng

RT: 26.050 min Scan# 3666

Delta R.T. 0.003 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:276 Resp: 2972


Ion Ratio Lower Upper

276 100

138 25.6 18.9 28.3

277 24.0 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 0.372 ng

RT: 23.001 min Scan# 2623

Delta R.T. 0.003 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:252 Resp: 2413

Ion Ratio Lower Upper

252 100

253 27.2 20.1 30.1

125 15.7 9.9 14.9#



#38

Benzo(k)fluoranthene

Concen: 0.381 ng

RT: 23.048 min Scan# 2639

Delta R.T. 0.003 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:252 Resp: 2442


Ion Ratio Lower Upper

252 100

253 26.7 20.5 30.7

125 15.7 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.401 ng

RT: 23.594 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:252 Resp: 2250

Ion Ratio Lower Upper

252 100

253 28.5 21.5 32.3

125 17.0 12.6 19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.397 ng

RT: 26.074 min Scan# 3674

Delta R.T. 0.003 min

Lab File: BN035878.D

Acq: 02 Jan 2025 15:51




Tgt Ion:278 Resp: 2353


Ion Ratio Lower Upper

278 100

139 21.6 15.9 23.9

279 29.2 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 0.359 ng

RT: 26.767 min Scan# 3911

Delta R.T. 0.003 min

Lab File: BN035878.D


Acq: 02 Jan 2025 15:51

Instrument :

BNA_N

ClientSampleId :

ICVBN010225





Tgt Ion:276 Resp: 2378

Ion Ratio Lower Upper

276 100

277 28.9 22.8 34.2

138 26.7 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035878.D
 Acq On : 02 Jan 2025 15:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010225

Quant Time: Jan 02 16:21:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.397	0.468	-17.9	111	0.00
3	n-Nitrosodimethylamine	0.693	0.725	-4.6	94	0.00
4 S	2-Fluorophenol	0.981	0.840	14.4	78	0.00
5 S	Phenol-d6	1.219	1.011	17.1	75	0.00
6	bis(2-Chloroethyl)ether	0.929	0.967	-4.1	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.317	0.315	0.6	91	0.00
9	Naphthalene	1.123	1.119	0.4	88	0.00
10	Hexachlorobutadiene	0.365	0.353	3.3	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.536	0.521	2.8	84	0.00
12	2-Methylnaphthalene	0.695	0.708	-1.9	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.192	0.168	12.5	67	0.00
15 S	2-Fluorobiphenyl	1.755	1.893	-7.9	83	0.00
16	Acenaphthylene	1.884	2.127	-12.9	88	0.00
17	Acenaphthene	1.235	1.316	-6.6	82	0.00
18	Fluorene	1.359	1.386	-2.0	80	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
20	4,6-Dinitro-2-methylphenol	0.069	0.071	-2.9	75	0.00
21	4-Bromophenyl-phenylether	0.274	0.273	0.4	72	0.00
22	Hexachlorobenzene	0.374	0.381	-1.9	75	0.00
23	Atrazine	0.184	0.177	3.8	71	0.00
24	Pentachlorophenol	0.132	0.130	1.5	77	0.01
25	Phenanthrene	1.167	1.220	-4.5	75	0.00
26	Anthracene	1.062	1.126	-6.0	79	0.00
27 SURR	Fluoranthene-d10	0.993	0.948	4.5	68	0.00
28	Fluoranthene	1.361	1.297	4.7	69	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	69	0.00
30	Pyrene	1.624	1.586	2.3	70	0.00
31 S	Terphenyl-d14	0.797	0.783	1.8	69	0.00
32	Benzo(a)anthracene	1.410	1.428	-1.3	74	0.00
33	Chrysene	1.475	1.481	-0.4	71	0.00
34	Bis(2-ethylhexyl)phthalate	0.574	0.564	1.7	67	0.00
35 I	Perylene-d12	1.000	1.000	0.0	72	0.00
36	Indeno(1,2,3-cd)pyrene	1.577	1.575	0.1	80	0.00
37	Benzo(b)fluoranthene	1.374	1.279	6.9	71	0.00
38	Benzo(k)fluoranthene	1.360	1.294	4.9	74	0.00
39 C	Benzo(a)pyrene	1.190	1.192	-0.2	79	0.00
40	Dibenzo(a,h)anthracene	1.255	1.247	0.6	81	0.00
41	Benzo(g,h,i)perylene	1.403	1.260	10.2	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
 Data File : BN035878.D
 Acq On : 02 Jan 2025 15:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN010225

Quant Time: Jan 02 16:21:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.471	-17.7	111	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	94	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	78	0.00
5 S	Phenol-d6	0.400	0.332	17.0	75	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	91	0.00
9	Naphthalene	0.400	0.399	0.3	88	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	84	0.00
12	2-Methylnaphthalene	0.400	0.408	-2.0	88	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.400	0.350	12.5	67	0.00
15 S	2-Fluorobiphenyl	0.400	0.431	-7.7	83	0.00
16	Acenaphthylene	0.400	0.452	-13.0	88	0.00
17	Acenaphthene	0.400	0.426	-6.5	82	0.00
18	Fluorene	0.400	0.408	-2.0	80	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	70	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.410	-2.5	75	0.00
21	4-Bromophenyl-phenylether	0.400	0.398	0.5	72	0.00
22	Hexachlorobenzene	0.400	0.407	-1.7	75	0.00
23	Atrazine	0.400	0.386	3.5	71	0.00
24	Pentachlorophenol	0.400	0.392	2.0	77	0.01
25	Phenanthrene	0.400	0.418	-4.5	75	0.00
26	Anthracene	0.400	0.424	-6.0	79	0.00
27 SURR	Fluoranthene-d10	0.400	0.382	4.5	68	0.00
28	Fluoranthene	0.400	0.381	4.8	69	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	69	0.00
30	Pyrene	0.400	0.391	2.3	70	0.00
31 S	Terphenyl-d14	0.400	0.393	1.8	69	0.00
32	Benzo(a)anthracene	0.400	0.405	-1.3	74	0.00
33	Chrysene	0.400	0.402	-0.5	71	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.393	1.8	67	0.00
35 I	Perylene-d12	0.400	0.400	0.0	72	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	80	0.00
37	Benzo(b)fluoranthene	0.400	0.372	7.0	71	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	74	0.00
39 C	Benzo(a)pyrene	0.400	0.401	-0.3	79	0.00
40	Dibenzo(a,h)anthracene	0.400	0.397	0.8	81	0.00
41	Benzo(g,h,i)perylene	0.400	0.359	10.3	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG No.: Q1121

Instrument ID: BNA_N Calibration Date/Time: 01/20/2025 10:27

Lab File ID: BN035991.D Init. Calib. Date(s): 01/02/2025 01/02/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:28 15:04

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.536	0.576		7.5	20.0
Fluoranthene-d10	0.993	1.092		10.0	20.0
2-Fluorophenol	0.981	1.122		14.4	20.0
Phenol-d6	1.219	1.348		10.6	20.0
Nitrobenzene-d5	0.317	0.415		30.9	20.0
2-Fluorobiphenyl	1.755	1.847		5.2	20.0
2,4,6-Tribromophenol	0.192	0.257		33.9	20.0
Terphenyl-d14	0.797	0.890		11.7	20.0
1,4-Dioxane	0.397	0.460		15.9	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN035991.D
 Acq On : 20 Jan 2025 10:27
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Jan 20 11:14:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2147	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4278	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2153	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	4411	0.400	ng	-0.02
29) Chrysene-d12	21.367	240	3822	0.400	ng	-0.02
35) Perylene-d12	23.666	264	3850	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.398	112	2408	0.457	ng	-0.01
5) Phenol-d6	6.972	99	2894	0.442	ng	-0.01
8) Nitrobenzene-d5	8.956	82	1775	0.524	ng	-0.03
11) 2-Methylnaphthalene-d10	12.198	152	2463	0.430	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	553	0.535	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	3977	0.421	ng	-0.02
27) Fluoranthene-d10	19.220	212	4816	0.440	ng	-0.02
31) Terphenyl-d14	19.815	244	3403	0.447	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.318	88	988	0.463	ng	95
3) n-Nitrosodimethylamine	3.621	42	1749	0.470	ng	# 95
6) bis(2-Chloroethyl)ether	7.239	93	2415	0.485	ng	97
9) Naphthalene	10.654	128	5273	0.439	ng	97
10) Hexachlorobutadiene	10.953	225	1637	0.420	ng	# 98
12) 2-Methylnaphthalene	12.274	142	3238	0.436	ng	99
16) Acenaphthylene	14.164	152	4352	0.429	ng	100
17) Acenaphthene	14.506	154	2979	0.448	ng	99
18) Fluorene	15.500	166	3705	0.507	ng	99
20) 4,6-Dinitro-2-methylph...	15.560	198	370	0.483	ng	# 86
21) 4-Bromophenyl-phenylether	16.392	248	1338	0.442	ng	# 76
22) Hexachlorobenzene	16.504	284	1726	0.419	ng	96
23) Atrazine	16.653	200	912	0.449	ng	# 85
24) Pentachlorophenol	16.839	266	699	0.479	ng	98
25) Phenanthrene	17.223	178	5645	0.439	ng	99
26) Anthracene	17.323	178	4967	0.424	ng	98
28) Fluoranthene	19.248	202	6520	0.435	ng	99
30) Pyrene	19.611	202	6583	0.424	ng	100
32) Benzo(a)anthracene	21.349	228	5656	0.420	ng	99
33) Chrysene	21.403	228	5943	0.422	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	2945	0.537	ng	99
36) Indeno(1,2,3-cd)pyrene	26.002	276	5925	0.390	ng	98
37) Benzo(b)fluoranthene	22.970	252	5761	0.436	ng	98
38) Benzo(k)fluoranthene	23.017	252	5712	0.436	ng	99
39) Benzo(a)pyrene	23.561	252	4835	0.422	ng	98
40) Dibenzo(a,h)anthracene	26.017	278	4653	0.385	ng	99
41) Benzo(g,h,i)perylene	26.718	276	5237	0.388	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035991.D
Acq On : 20 Jan 2025 10:27
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Jan 20 11:14:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 2147
Ion Ratio Lower Upper
152 100
150 158.2 117.8 176.6
115 65.3 51.0 76.4




#2
1,4-Dioxane
Concen: 0.463 ng
RT: 3.318 min Scan# 19
Delta R.T. -0.004 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion: 88 Resp: 988
Ion Ratio Lower Upper
88 100
43 45.3 32.7 49.1
58 82.6 63.0 94.4



Abundance Scan 111 (3.632 min): BN035873.D\data.ms (-10



#3

n-Nitrosodimethylamine

Concen: 0.470 ng

RT: 3.621 min Scan# 61

Delta R.T. -0.011 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion: 42 Resp: 1749

Ion Ratio Lower Upper


42 100

74 81.2 62.2 93.2


44 5.8 7.1 10.7#



Abundance Scan 61 (3.621 min): BN035991.D\data.ms (-42)



Abundance Scan 357 (5.409 min): BN035873.D\data.ms (-34



#4

2-Fluorophenol

Concen: 0.457 ng

RT: 5.398 min Scan# 307

Delta R.T. -0.011 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27



Tgt Ion: 112 Resp: 2408

Ion Ratio Lower Upper


112 100


64 62.3 48.2 72.2

63 38.3 30.0 45.0



Abundance Scan 307 (5.398 min): BN035991.D\data.ms (-28









#5
Phenol-d6
Concen: 0.442 ng
RT: 6.972 min Scan# 51
Delta R.T. -0.011 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 2894
Ion Ratio Lower Upper
99 100
42 30.5 23.5 35.3
71 44.6 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.485 ng
RT: 7.239 min Scan# 562
Delta R.T. -0.018 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion: 93 Resp: 2415
Ion Ratio Lower Upper
93 100
63 81.1 62.0 93.0
95 32.5 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 91
Delta R.T. -0.015 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	4278
Ion	Ratio	Lower	Upper
136	100		
137	12.2	10.6	15.8
54	10.1	9.8	14.6
68	6.4	6.6	10.0



#8
Nitrobenzene-d5
Concen: 0.524 ng
RT: 8.956 min Scan# 771
Delta R.T. -0.025 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:	82	Resp:	1775
Ion	Ratio	Lower	Upper
82	100		
128	33.7	36.9	55.3
54	72.1	50.4	75.6









#9
Naphthalene
Concen: 0.439 ng
RT: 10.654 min Scan# 91
Delta R.T. -0.025 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 5273
Ion Ratio Lower Upper
128 100
129 12.0 10.6 16.0
127 14.9 12.8 19.2



#10
Hexachlorobutadiene
Concen: 0.420 ng
RT: 10.953 min Scan# 958
Delta R.T. -0.025 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:225 Resp: 1637
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.0 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.430 ng
RT: 12.198 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 2463
Ion Ratio Lower Upper
152 100
151 21.6 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.436 ng
RT: 12.274 min Scan# 1161
Delta R.T. -0.019 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:142 Resp: 3238
Ion Ratio Lower Upper
142 100
141 91.3 71.9 107.9
115 50.5 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 14
Delta R.T. -0.010 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 2153
Ion Ratio Lower Upper
164 100
162 104.1 83.1 124.7
160 50.0 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 0.535 ng
RT: 15.933 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:330 Resp: 553
Ion Ratio Lower Upper
330 100
332 96.6 77.2 115.8
141 42.3 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 0.421 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.020 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	3977
Ion Ratio	Lower	Upper	
172	100		
171	37.5	30.5	45.7
170	25.2	20.8	31.2



#16
Acenaphthylene
Concen: 0.429 ng
RT: 14.164 min Scan# 1414
Delta R.T. -0.020 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:	152	Resp:	4352
Ion Ratio	Lower	Upper	
152	100		
151	20.7	16.3	24.5
153	13.4	10.6	15.8









#17
Acenaphthene
Concen: 0.448 ng
RT: 14.506 min Scan# 1496
Delta R.T. -0.020 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
154	100		
153	112.1	90.5	135.7
152	58.9	48.6	73.0



#18
Fluorene
Concen: 0.507 ng
RT: 15.500 min Scan# 1539
Delta R.T. -0.010 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion	Ratio	Lower	Upper
166	100		
165	100.1	80.6	120.8
167	13.4	11.4	17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. -0.024 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:188 Resp: 4411

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.8 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.483 ng

RT: 15.560 min Scan# 1544

Delta R.T. -0.011 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27




Tgt Ion:198 Resp: 370


Ion Ratio Lower Upper

198 100

51 70.6 64.3 96.5

105 48.2 50.0 75.0#









#21
4-Bromophenyl-phenylether
Concen: 0.442 ng
RT: 16.392 min Scan# 1611
Delta R.T. -0.011 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	103.5	76.8	115.2
141	40.6	63.6	95.4



#22
Hexachlorobenzene
Concen: 0.419 ng
RT: 16.504 min Scan# 1620
Delta R.T. -0.011 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion	Ratio	Lower	Upper
284	100		
142	43.7	31.4	47.0
249	35.3	27.8	41.8









#23
Atrazine
Concen: 0.449 ng
RT: 16.653 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 912
Ion Ratio Lower Upper
200 100
173 27.1 35.4 53.0#
215 49.1 42.4 63.6



#24
Pentachlorophenol
Concen: 0.479 ng
RT: 16.839 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:266 Resp: 699
Ion Ratio Lower Upper
266 100
264 63.8 49.9 74.9
268 64.7 50.2 75.2









#25
Phenanthrene
Concen: 0.439 ng
RT: 17.223 min Scan# 1686
Delta R.T. -0.024 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178	Resp:	5645
Ion Ratio	Lower	Upper
178	100	
176	20.2	15.9 23.9
179	15.4	12.9 19.3



#26
Anthracene
Concen: 0.424 ng
RT: 17.323 min Scan# 1686
Delta R.T. -0.011 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:178	Resp:	4967
Ion Ratio	Lower	Upper
178	100	
176	19.3	15.0 22.6
179	15.2	13.0 19.6









#27
Fluoranthene-d10
Concen: 0.440 ng
RT: 19.220 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	12.1	9.0	13.6
104	7.5	5.4	8.2



#28
Fluoranthene
Concen: 0.435 ng
RT: 19.248 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion	Ratio	Lower	Upper
202	100		
101	9.8	7.2	10.8
203	17.3	13.9	20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.367 min Scan# 21
Delta R.T. -0.017 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
240	100		
120	13.3	11.1	16.7
236	29.7	24.6	36.8



#30
Pyrene
Concen: 0.424 ng
RT: 19.611 min Scan# 2018
Delta R.T. -0.017 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.0	25.6
203	18.1	14.6	22.0









#31
Terphenyl-d14
Concen: 0.447 ng
RT: 19.815 min Scan# 2062
Delta R.T. -0.022 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.7	10.1	15.1
122	16.4	12.2	18.4



#32
Benzo(a)anthracene
Concen: 0.420 ng
RT: 21.349 min Scan# 2253
Delta R.T. -0.026 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.1	22.7	34.1
229	20.3	17.1	25.7





#33

Chrysene

Concen: 0.422 ng

RT: 21.403 min Scan# 21

Delta R.T. -0.017 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 5943

Ion Ratio Lower Upper

228 100

226 31.0 25.2 37.8

229 20.1 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.537 ng

RT: 21.295 min Scan# 2247

Delta R.T. -0.026 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27




Tgt Ion:149 Resp: 2945


Ion Ratio Lower Upper

149 100

167 27.0 21.4 32.0

279 3.7 3.0 4.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.666 min Scan# 21
Delta R.T. -0.028 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 3850
Ion Ratio Lower Upper
264 100
260 26.7 21.7 32.5
265 57.1 33.9 50.9#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.390 ng
RT: 26.002 min Scan# 3601
Delta R.T. -0.046 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:276 Resp: 5925
Ion Ratio Lower Upper
276 100
138 24.7 18.9 28.3
277 25.2 19.5 29.3









#37
Benzo(b)fluoranthene
Concen: 0.436 ng
RT: 22.970 min Scan# 2564
Delta R.T. -0.028 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 5761
Ion Ratio Lower Upper
252 100
253 25.5 20.1 30.1
125 13.6 9.9 14.9



#38
Benzo(k)fluoranthene
Concen: 0.436 ng
RT: 23.017 min Scan# 2580
Delta R.T. -0.028 min
Lab File: BN035991.D
Acq: 20 Jan 2025 10:27

Tgt Ion:252 Resp: 5712
Ion Ratio Lower Upper
252 100
253 25.9 20.5 30.7
125 14.0 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.422 ng

RT: 23.561 min Scan# 21

Delta R.T. -0.028 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 4835

Ion Ratio Lower Upper

252 100

253 27.7 21.5 32.3

125 17.2 12.6 19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.385 ng

RT: 26.017 min Scan# 3606

Delta R.T. -0.054 min

Lab File: BN035991.D

Acq: 20 Jan 2025 10:27




Tgt Ion:278 Resp: 4653


Ion Ratio Lower Upper

278 100

139 19.7 15.9 23.9

279 28.3 23.0 34.4





#41

Benzo(g,h,i)perylene

Concen: 0.388 ng

RT: 26.718 min Scan# 3846

Delta R.T. -0.046 min

Lab File: BN035991.D


Acq: 20 Jan 2025 10:27

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:276 Resp: 5237

Ion Ratio Lower Upper

276 100

277 25.7 22.8 34.2

138 24.1 20.1 30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035991.D
Acq On : 20 Jan 2025 10:27
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 11:14:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	170#	-0.02
2	1,4-Dioxane	0.397	0.460	-15.9	210#	0.00
3	n-Nitrosodimethylamine	0.693	0.815	-17.6	205#	-0.01
4 S	2-Fluorophenol	0.981	1.122	-14.4	201#	-0.01
5 S	Phenol-d6	1.219	1.348	-10.6	194#	-0.01
6	bis(2-Chloroethyl)ether	0.929	1.125	-21.1	205#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	166#	-0.01
8 S	Nitrobenzene-d5	0.317	0.415	-30.9#	232#	-0.03
9	Naphthalene	1.123	1.233	-9.8	188#	-0.03
10	Hexachlorobutadiene	0.365	0.383	-4.9	179#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.536	0.576	-7.5	181#	-0.02
12	2-Methylnaphthalene	0.695	0.757	-8.9	183#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	159#	0.00
14 S	2,4,6-Tribromophenol	0.192	0.257	-33.9#	216#	-0.02
15 S	2-Fluorobiphenyl	1.755	1.847	-5.2	171#	-0.02
16	Acenaphthylene	1.884	2.021	-7.3	176#	-0.02
17	Acenaphthene	1.235	1.384	-12.1	183#	-0.02
18	Fluorene	1.359	1.721	-26.6#	209#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	159#	-0.02
20	4,6-Dinitro-2-methylphenol	0.069	0.084	-21.7	200#	-0.01
21	4-Bromophenyl-phenylether	0.274	0.303	-10.6	180#	-0.01
22	Hexachlorobenzene	0.374	0.391	-4.5	175#	-0.01
23	Atrazine	0.184	0.207	-12.5	188#	-0.01
24	Pentachlorophenol	0.132	0.158	-19.7	214#	-0.01
25	Phenanthrene	1.167	1.280	-9.7	179#	-0.02
26	Anthracene	1.062	1.126	-6.0	178#	-0.01
27 SURR	Fluoranthene-d10	0.993	1.092	-10.0	178#	-0.02
28	Fluoranthene	1.361	1.478	-8.6	177#	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	161#	-0.02
30	Pyrene	1.624	1.722	-6.0	176#	-0.02
31 S	Terphenyl-d14	0.797	0.890	-11.7	181#	-0.02
32	Benzo(a)anthracene	1.410	1.480	-5.0	177#	-0.03
33	Chrysene	1.475	1.555	-5.4	173#	-0.02
34	Bis(2-ethylhexyl)phthalate	0.574	0.771	-34.3#	212#	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	147	-0.03
36	Indeno(1,2,3-cd)pyrene	1.577	1.539	2.4	160#	-0.05
37	Benzo(b)fluoranthene	1.374	1.496	-8.9	170#	-0.03
38	Benzo(k)fluoranthene	1.360	1.484	-9.1	174#	-0.03
39 C	Benzo(a)pyrene	1.190	1.256	-5.5	170#	-0.03
40	Dibenzo(a,h)anthracene	1.255	1.209	3.7	160#	-0.05
41	Benzo(g,h,i)perylene	1.403	1.360	3.1	160#	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035991.D
Acq On : 20 Jan 2025 10:27
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 11:14:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	170	-0.02
2	1,4-Dioxane	0.400	0.463	-15.8	210	0.00
3	n-Nitrosodimethylamine	0.400	0.470	-17.5	205	-0.01
4 S	2-Fluorophenol	0.400	0.457	-14.2	201	-0.01
5 S	Phenol-d6	0.400	0.442	-10.5	194	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.485	-21.2	205	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	166	-0.01
8 S	Nitrobenzene-d5	0.400	0.524	-31.0#	232	-0.03
9	Naphthalene	0.400	0.439	-9.7	188	-0.03
10	Hexachlorobutadiene	0.400	0.420	-5.0	179	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.430	-7.5	181	-0.02
12	2-Methylnaphthalene	0.400	0.436	-9.0	183	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	159	0.00
14 S	2,4,6-Tribromophenol	0.400	0.535	-33.8#	216	-0.02
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	171	-0.02
16	Acenaphthylene	0.400	0.429	-7.2	176	-0.02
17	Acenaphthene	0.400	0.448	-12.0	183	-0.02
18	Fluorene	0.400	0.507	-26.7#	209	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	159	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.483	-20.7	200	-0.01
21	4-Bromophenyl-phenylether	0.400	0.442	-10.5	180	-0.01
22	Hexachlorobenzene	0.400	0.419	-4.7	175	-0.01
23	Atrazine	0.400	0.449	-12.2	188	-0.01
24	Pentachlorophenol	0.400	0.479	-19.7	214	-0.01
25	Phenanthrene	0.400	0.439	-9.7	179	-0.02
26	Anthracene	0.400	0.424	-6.0	178	-0.01
27 SURR	Fluoranthene-d10	0.400	0.440	-10.0	178	-0.02
28	Fluoranthene	0.400	0.435	-8.7	177	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	161	-0.02
30	Pyrene	0.400	0.424	-6.0	176	-0.02
31 S	Terphenyl-d14	0.400	0.447	-11.7	181	-0.02
32	Benzo(a)anthracene	0.400	0.420	-5.0	177	-0.03
33	Chrysene	0.400	0.422	-5.5	173	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.537	-34.3#	212	-0.03
35 I	Perylene-d12	0.400	0.400	0.0	147	-0.03
36	Indeno(1,2,3-cd)pyrene	0.400	0.390	2.5	160	-0.05
37	Benzo(b)fluoranthene	0.400	0.436	-9.0	170	-0.03
38	Benzo(k)fluoranthene	0.400	0.436	-9.0	174	-0.03
39 C	Benzo(a)pyrene	0.400	0.422	-5.5	170	-0.03
40	Dibenzo(a,h)anthracene	0.400	0.385	3.8	160	-0.05
41	Benzo(g,h,i)perylene	0.400	0.388	3.0	160	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG No.: Q1121

Instrument ID: BNA_N Calibration Date/Time: 01/20/2025 16:11

Lab File ID: BN036000.D Init. Calib. Date(s): 01/02/2025 01/02/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:28 15:04

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.536	0.580		8.2	50.0
Fluoranthene-d10	0.993	1.025		3.2	50.0
2-Fluorophenol	0.981	1.150		17.2	50.0
Phenol-d6	1.219	1.348		10.6	50.0
Nitrobenzene-d5	0.317	0.430		35.6	50.0
2-Fluorobiphenyl	1.755	1.950		11.1	50.0
2,4,6-Tribromophenol	0.192	0.234		21.9	50.0
Terphenyl-d14	0.797	0.905		13.6	50.0
1,4-Dioxane	0.397	0.483		21.7	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN036000.D
 Acq On : 20 Jan 2025 16:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/21/2025
 Supervised By :mohammad ahmed 01/21/2025

Quant Time: Jan 20 17:29:00 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2210	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4309	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2121	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	3983	0.400	ng	-0.02
29) Chrysene-d12	21.367	240	3167	0.400	ng	-0.02
35) Perylene-d12	23.666	264	3290	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.390	112	2542	0.469	ng	-0.02
5) Phenol-d6	6.972	99	2979	0.442	ng	-0.01
8) Nitrobenzene-d5	8.956	82	1855	0.544	ng	-0.03
11) 2-Methylnaphthalene-d10	12.198	152	2501	0.434	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	497	0.488	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4137	0.444	ng	-0.02
27) Fluoranthene-d10	19.221	212	4081	0.413	ng	-0.02
31) Terphenyl-d14	19.815	244	2866	0.454	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.318	88	1068	0.487	ng	99
3) n-Nitrosodimethylamine	3.621	42	1932	0.504	ng	# 98
6) bis(2-Chloroethyl)ether	7.232	93	2472	0.482	ng	97
9) Naphthalene	10.654	128	5411	0.447	ng	97
10) Hexachlorobutadiene	10.953	225	1718	0.437	ng	# 99
12) 2-Methylnaphthalene	12.269	142	3275	0.438	ng	99
16) Acenaphthylene	14.164	152	4225	0.423	ng	99
17) Acenaphthene	14.506	154	2843	0.434	ng	99
18) Fluorene	15.500	166	3587	0.498	ng	98
20) 4,6-Dinitro-2-methylph...	15.560	198	256	0.370	ng	94
21) 4-Bromophenyl-phenylether	16.379	248	1236	0.453	ng	# 88
22) Hexachlorobenzene	16.504	284	1595	0.428	ng	97
23) Atrazine	16.653	200	839	0.458	ng	# 87
24) Pentachlorophenol	16.839	266	597	0.453	ng	97
25) Phenanthrene	17.223	178	5041	0.434	ng	99
26) Anthracene	17.323	178	4488	0.424	ng	99
28) Fluoranthene	19.248	202	5508	0.407	ng	99
30) Pyrene	19.611	202	5550	0.432	ng	99
32) Benzo(a)anthracene	21.349	228	4679	0.419	ng	98
33) Chrysene	21.403	228	4915	0.421	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	2577	0.567	ng	98
36) Indeno(1,2,3-cd)pyrene	26.002	276	5439	0.419	ng	99
37) Benzo(b)fluoranthene	22.973	252	4943m	0.437	ng	
38) Benzo(k)fluoranthene	23.020	252	4806	0.430	ng	98
39) Benzo(a)pyrene	23.567	252	4125	0.422	ng	96
40) Dibenzo(a,h)anthracene	26.019	278	4327	0.419	ng	100
41) Benzo(g,h,i)perylene	26.718	276	4707	0.408	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN036000.D
 Acq On : 20 Jan 2025 16:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025


Quant Time: Jan 20 17:29:00 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jan 02 15:39:17 2025


Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:152 Resp: 2210
Ion Ratio Lower Upper
152 100
150 150.6 117.8 176.6
115 63.5 51.0 76.4




Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025




#2
1,4-Dioxane
Concen: 0.487 ng
RT: 3.318 min Scan# 19
Delta R.T. -0.004 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion: 88 Resp: 1068
Ion Ratio Lower Upper
88 100
43 43.3 32.7 49.1
58 78.6 63.0 94.4



Abundance Scan 111 (3.632 min): BN035873.D\data.ms (-10



#3

n-Nitrosodimethylamine

Concen: 0.504 ng

RT: 3.621 min Scan# 61

Delta R.T. -0.011 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion: 42 Resp: 193

Ion Ratio Lower Upper

42 100

74 76.5 62.2 93.2



44 5.8 7.1 10.7

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/21/2025


Supervised By :mohammad ahmed 01/21/2025



Abundance Scan 61 (3.621 min): BN036000.D\data.ms (-42)



Abundance Scan 357 (5.409 min): BN035873.D\data.ms (-34



#4

2-Fluorophenol

Concen: 0.469 ng

RT: 5.390 min Scan# 306

Delta R.T. -0.018 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11



Tgt Ion:112 Resp: 2542

Ion Ratio Lower Upper


112 100


64 62.7 48.2 72.2

63 39.1 30.0 45.0




Abundance Scan 306 (5.390 min): BN036000.D\data.ms (-28





#5
Phenol-d6
Concen: 0.442 ng
RT: 6.972 min Scan# 51
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 99 Resp: 2979
Ion Ratio Lower Upper
99 100
42 31.4 23.5 35.3
71 45.4 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.482 ng
RT: 7.232 min Scan# 561
Delta R.T. -0.026 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion: 93 Resp: 2472
Ion Ratio Lower Upper
93 100
63 80.8 62.0 93.0
95 32.5 25.5 38.3





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 911
Delta R.T. -0.015 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 136 Resp: 4309
Ion Ratio Lower Upper
136 100
137 12.6 10.6 15.8
54 9.5 9.8 14.6
68 6.3 6.6 10.0#




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#8
Nitrobenzene-d5
Concen: 0.544 ng
RT: 8.956 min Scan# 771
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion: 82 Resp: 1855
Ion Ratio Lower Upper
82 100
128 36.0 36.9 55.3#
54 69.0 50.4 75.6









#9
Naphthalene
Concen: 0.447 ng
RT: 10.654 min Scan# 911
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:128 Resp: 541
Ion Ratio Lower Upper
128 100
129 12.5 10.6 16.0
127 14.5 12.8 19.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#10
Hexachlorobutadiene
Concen: 0.437 ng
RT: 10.953 min Scan# 958
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:225 Resp: 1718
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.5 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.434 ng
RT: 12.198 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 250
Ion Ratio Lower Upper
152 100
151 21.9 17.0 25.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#12
2-Methylnaphthalene
Concen: 0.438 ng
RT: 12.269 min Scan# 1160
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:142 Resp: 3275
Ion Ratio Lower Upper
142 100
141 90.2 71.9 107.9
115 51.2 39.6 59.4





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1441
Delta R.T. -0.010 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 212
Ion Ratio Lower Upper
164 100
162 101.5 83.1 124.7
160 51.7 46.0 69.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025




#14
2,4,6-Tribromophenol
Concen: 0.488 ng
RT: 15.933 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11



Tgt Ion:330 Resp: 497
Ion Ratio Lower Upper
330 100
332 101.8 77.2 115.8
141 45.3 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 0.444 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.020 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:	172	Resp:	413
Ion Ratio	Lower	Upper	
172	100		
171	36.7	30.5	45.7
170	25.1	20.8	31.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#16
Acenaphthylene
Concen: 0.423 ng
RT: 14.164 min Scan# 1414
Delta R.T. -0.020 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:	152	Resp:	4225
Ion Ratio	Lower	Upper	
152	100		
151	20.8	16.3	24.5
153	12.8	10.6	15.8





#17
Acenaphthene
Concen: 0.434 ng
RT: 14.506 min Scan# 1496
Delta R.T. -0.020 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:154 Resp: 284
Ion Ratio Lower Upper
154 100
153 112.8 90.5 135.7
152 59.8 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#18
Fluorene
Concen: 0.498 ng
RT: 15.500 min Scan# 1539
Delta R.T. -0.010 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:166 Resp: 3587
Ion Ratio Lower Upper
166 100
165 98.9 80.6 120.8
167 13.1 11.4 17.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 11
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:188 Resp: 398
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.6 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.370 ng
RT: 15.560 min Scan# 1544
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:198 Resp: 256
Ion Ratio Lower Upper
198 100
51 75.8 64.3 96.5
105 57.7 50.0 75.0





#21
4-Bromophenyl-phenylether
Concen: 0.453 ng
RT: 16.379 min Scan# 1610
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDC00.4EC



Tgt Ion:248 Resp: 1230
Ion Ratio Lower Upper
248 100
250 93.2 76.8 115.2
141 100.0 63.6 95.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#22
Hexachlorobenzene
Concen: 0.428 ng
RT: 16.504 min Scan# 1620
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:284 Resp: 1595
Ion Ratio Lower Upper
284 100
142 42.0 31.4 47.0
249 35.5 27.8 41.8





#23
Atrazine
Concen: 0.458 ng
RT: 16.653 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:200 Resp: 839
Ion Ratio Lower Upper
200 100
173 29.2 35.4 53.0
215 49.4 42.4 63.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#24
Pentachlorophenol
Concen: 0.453 ng
RT: 16.839 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:266 Resp: 597
Ion Ratio Lower Upper
266 100
264 61.3 49.9 74.9
268 66.2 50.2 75.2









#25
Phenanthrene
Concen: 0.434 ng
RT: 17.223 min Scan# 1686
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:178 Resp: 504
Ion Ratio Lower Upper
178 100
176 19.7 15.9 23.9
179 15.7 12.9 19.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#26
Anthracene
Concen: 0.424 ng
RT: 17.323 min Scan# 1686
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:178 Resp: 4488
Ion Ratio Lower Upper
178 100
176 19.3 15.0 22.6
179 15.7 13.0 19.6





#27
Fluoranthene-d10
Concen: 0.413 ng
RT: 19.221 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:212 Resp: 408
Ion Ratio Lower Upper
212 100
106 11.9 9.0 13.6
104 7.2 5.4 8.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#28
Fluoranthene
Concen: 0.407 ng
RT: 19.248 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:202 Resp: 5508
Ion Ratio Lower Upper
202 100
101 9.4 7.2 10.8
203 17.2 13.9 20.9





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.367 min Scan# 2136
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:240 Resp: 316
Ion Ratio Lower Upper
240 100
120 15.5 11.1 16.7
236 30.4 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#30
Pyrene
Concen: 0.432 ng
RT: 19.611 min Scan# 2018
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:202 Resp: 5550
Ion Ratio Lower Upper
202 100
200 21.5 17.0 25.6
203 17.9 14.6 22.0





#31
Terphenyl-d14
Concen: 0.454 ng
RT: 19.815 min Scan# 2062
Delta R.T. -0.022 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:244 Resp: 2860
Ion Ratio Lower Upper
244 100
212 11.8 10.1 15.1
122 15.6 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#32
Benzo(a)anthracene
Concen: 0.419 ng
RT: 21.349 min Scan# 2253
Delta R.T. -0.026 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:228 Resp: 4679
Ion Ratio Lower Upper
228 100
226 29.3 22.7 34.1
229 20.3 17.1 25.7





#33

Chrysene

Concen: 0.421 ng

RT: 21.403 min Scan# 21

Delta R.T. -0.017 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:	228	Resp:	491
Ion Ratio	Lower	Upper	
228	100		
226	32.3	25.2	37.8
229	20.2	16.0	24.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.567 ng




RT: 21.295 min Scan# 2247


Delta R.T. -0.026 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11


Tgt Ion:	149	Resp:	2577
Ion Ratio	Lower	Upper	
149	100		
167	27.8	21.4	32.0
279	3.5	3.0	4.6





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.666 min Scan# 2802
Delta R.T. -0.028 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:264 Resp: 3290
Ion Ratio Lower Upper
264 100
260 25.9 21.7 32.5
265 67.1 33.9 50.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.419 ng
RT: 26.002 min Scan# 3601
Delta R.T. -0.046 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11


Tgt Ion:276 Resp: 5439
Ion Ratio Lower Upper
276 100
138 24.6 18.9 28.3
277 24.7 19.5 29.3





#37
Benzo(b)fluoranthene
Concen: 0.437 ng m
RT: 22.973 min Scan# 2565
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:252 Resp: 494
Ion Ratio Lower Upper
252 100
253 26.8 20.1 30.1
125 15.3 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#38
Benzo(k)fluoranthene
Concen: 0.430 ng
RT: 23.020 min Scan# 2581
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:252 Resp: 4806
Ion Ratio Lower Upper
252 100
253 26.7 20.5 30.7
125 14.3 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.422 ng

RT: 23.567 min Scan# 21

Delta R.T. -0.022 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4EC

Tgt Ion:	252	Resp:	412
Ion Ratio	Lower	Upper	
252	100		
253	29.4	21.5	32.3
125	17.0	12.6	19.0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.419 ng




RT: 26.019 min Scan# 3607


Delta R.T. -0.052 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11


Tgt Ion:	278	Resp:	4327
Ion Ratio	Lower	Upper	
278	100		
139	20.3	15.9	23.9
279	28.7	23.0	34.4





#41
Benzo(g,h,i)perylene
Concen: 0.408 ng
RT: 26.718 min Scan# 3846
Delta R.T. -0.046 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 276 Resp: 470
Ion Ratio Lower Upper
276 100
277 27.1 22.8 34.2
138 24.7 20.1 30.1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN036000.D
Acq On : 20 Jan 2025 16:11
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 17:29:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	175#	-0.02
2	1,4-Dioxane	0.397	0.483	-21.7	227#	0.00
3	n-Nitrosodimethylamine	0.693	0.874	-26.1#	226#	-0.01
4 S	2-Fluorophenol	0.981	1.150	-17.2	212#	-0.02
5 S	Phenol-d6	1.219	1.348	-10.6	200#	-0.01
6	bis(2-Chloroethyl)ether	0.929	1.119	-20.5	209#	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	167#	-0.01
8 S	Nitrobenzene-d5	0.317	0.430	-35.6#	243#	-0.03
9	Naphthalene	1.123	1.256	-11.8	193#	-0.03
10	Hexachlorobutadiene	0.365	0.399	-9.3	188#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.536	0.580	-8.2	184#	-0.02
12	2-Methylnaphthalene	0.695	0.760	-9.4	185#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	156#	0.00
14 S	2,4,6-Tribromophenol	0.192	0.234	-21.9	194#	-0.02
15 S	2-Fluorobiphenyl	1.755	1.950	-11.1	178#	-0.02
16	Acenaphthylene	1.884	1.992	-5.7	171#	-0.02
17	Acenaphthene	1.235	1.340	-8.5	175#	-0.02
18	Fluorene	1.359	1.691	-24.4	202#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	144	-0.02
20	4,6-Dinitro-2-methylphenol	0.069	0.064	7.2	138	-0.01
21	4-Bromophenyl-phenylether	0.274	0.310	-13.1	166#	-0.02
22	Hexachlorobenzene	0.374	0.400	-7.0	162#	-0.01
23	Atrazine	0.184	0.211	-14.7	173#	-0.01
24	Pentachlorophenol	0.132	0.150	-13.6	183#	-0.01
25	Phenanthrene	1.167	1.266	-8.5	159#	-0.02
26	Anthracene	1.062	1.127	-6.1	161#	-0.01
27 SURR	Fluoranthene-d10	0.993	1.025	-3.2	151#	-0.02
28	Fluoranthene	1.361	1.383	-1.6	150	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	133	-0.02
30	Pyrene	1.624	1.752	-7.9	148	-0.02
31 S	Terphenyl-d14	0.797	0.905	-13.6	152#	-0.02
32	Benzo(a)anthracene	1.410	1.477	-4.8	146	-0.03
33	Chrysene	1.475	1.552	-5.2	143	-0.02
34	Bis(2-ethylhexyl)phthalate	0.574	0.814	-41.8#	186#	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	125	-0.03
36	Indeno(1,2,3-cd)pyrene	1.577	1.653	-4.8	147	-0.05
37	Benzo(b)fluoranthene	1.374	1.502	-9.3	146	-0.03
38	Benzo(k)fluoranthene	1.360	1.461	-7.4	146	-0.03
39 C	Benzo(a)pyrene	1.190	1.254	-5.4	145	-0.02
40	Dibenzo(a,h)anthracene	1.255	1.315	-4.8	149	-0.05
41	Benzo(g,h,i)perylene	1.403	1.431	-2.0	144	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN036000.D
Acq On : 20 Jan 2025 16:11
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 20 17:29:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	175	-0.02
2	1,4-Dioxane	0.400	0.487	-21.7	227	0.00
3	n-Nitrosodimethylamine	0.400	0.504	-26.0#	226	-0.01
4 S	2-Fluorophenol	0.400	0.469	-17.2	212	-0.02
5 S	Phenol-d6	0.400	0.442	-10.5	200	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.482	-20.5	209	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	167	-0.01
8 S	Nitrobenzene-d5	0.400	0.544	-36.0#	243	-0.03
9	Naphthalene	0.400	0.447	-11.7	193	-0.03
10	Hexachlorobutadiene	0.400	0.437	-9.2	188	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.434	-8.5	184	-0.02
12	2-Methylnaphthalene	0.400	0.438	-9.5	185	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	156	0.00
14 S	2,4,6-Tribromophenol	0.400	0.488	-22.0	194	-0.02
15 S	2-Fluorobiphenyl	0.400	0.444	-11.0	178	-0.02
16	Acenaphthylene	0.400	0.423	-5.7	171	-0.02
17	Acenaphthene	0.400	0.434	-8.5	175	-0.02
18	Fluorene	0.400	0.498	-24.5	202	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	144	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.370	7.5	138	-0.01
21	4-Bromophenyl-phenylether	0.400	0.453	-13.2	166	-0.02
22	Hexachlorobenzene	0.400	0.428	-7.0	162	-0.01
23	Atrazine	0.400	0.458	-14.5	173	-0.01
24	Pentachlorophenol	0.400	0.453	-13.2	183	-0.01
25	Phenanthrene	0.400	0.434	-8.5	159	-0.02
26	Anthracene	0.400	0.424	-6.0	161	-0.01
27 SURR	Fluoranthene-d10	0.400	0.413	-3.2	151	-0.02
28	Fluoranthene	0.400	0.407	-1.7	150	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	133	-0.02
30	Pyrene	0.400	0.432	-8.0	148	-0.02
31 S	Terphenyl-d14	0.400	0.454	-13.5	152	-0.02
32	Benzo(a)anthracene	0.400	0.419	-4.7	146	-0.03
33	Chrysene	0.400	0.421	-5.2	143	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.567	-41.7#	186	-0.03
35 I	Perylene-d12	0.400	0.400	0.0	125	-0.03
36	Indeno(1,2,3-cd)pyrene	0.400	0.419	-4.7	147	-0.05
37	Benzo(b)fluoranthene	0.400	0.437	-9.2	146	-0.03
38	Benzo(k)fluoranthene	0.400	0.430	-7.5	146	-0.03
39 C	Benzo(a)pyrene	0.400	0.422	-5.5	145	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.419	-4.7	149	-0.05
41	Benzo(g,h,i)perylene	0.400	0.408	-2.0	144	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



QC SAMPLE DATA


- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035870.D
Acq On : 02 Jan 2025 10:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jan 02 15:39:17 2025



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	35733	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.0	34920	PASS
70	69	0.00	2	0.7	232	PASS
127	198	10	80	43.4	38907	PASS
197	198	0.00	2	0.3	256	PASS
198	198	100	100	100.0	89616	PASS
199	198	5	9	6.2	5523	PASS
275	198	10	60	25.6	22899	PASS
365	198	1	100	3.3	2996	PASS
441	198	0.01	100	9.5	8544	PASS
442	442	50	100	100.0	56448	PASS
443	442	15	24	19.3	10868	PASS

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/2/2025	BNA_N	<u>BN035870.D</u>
Compound Name	Response	Retention Time
DDT	243994	20.695
DDD	3023	20.307
DDE	79	19.748
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3102	247096	1.26

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035870.D
Acq On : 02 Jan 2025 10:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jan 02 12:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN035870.D\data.ms

(70) Pentachlorophenol (C)

16.854min (-0.012) 32111.49 ng




response 87193

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.79
264.00	61.60	56.74
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010225\
Data File : BN035870.D
Acq On : 02 Jan 2025 10:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jan 02 12:15:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN035870.D\data.ms

(77) Benzidine

19.448min (-0.006) 0.00 ng

response 567773


Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.41
183.00	13.20	12.73
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035990.D
Acq On : 20 Jan 2025 09:13
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jan 02 15:39:17 2025



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.3	67213	PASS
68	69	0.00	2	0.8	492	PASS
69	198	0.00	100	47.1	61771	PASS
70	69	0.00	2	0.7	439	PASS
127	198	10	80	49.0	64256	PASS
197	198	0.00	2	0.2	293	PASS
198	198	100	100	100.0	131045	PASS
199	198	5	9	6.4	8433	PASS
275	198	10	60	24.1	31595	PASS
365	198	1	100	4.1	5388	PASS
441	198	0.01	100	9.1	11885	PASS
442	442	50	100	100.0	70829	PASS
443	442	15	24	18.6	13182	PASS

Instrument :
BNA_N
ClientSampleId :
DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/20/2025	BNA_N	<u>BN035990.D</u>
Compound Name	Response	Retention Time
DDT	436464	20.675
DDD	13351	20.233
DDE	932	19.72
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
14283	450747	3.17

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035990.D
Acq On : 20 Jan 2025 09:13
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jan 20 11:00:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN035990.D\data.ms

(70) Pentachlorophenol (C)

16.840min (-0.006) 16779.47 ng




response 143269

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.53
264.00	61.60	62.56
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035990.D
Acq On : 20 Jan 2025 09:13
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jan 20 11:00:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN035990.D\data.ms

(77) Benzidine

19.428min (-0.006) 0.00 ng

response 652716

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.72
183.00	13.20	12.07
0.00	0.00	0.00

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	PB166108BL	SDG No.:	Q1121
Lab Sample ID:	PB166108BL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035992.D	1	01/17/25 10:10	01/20/25 11:03	PB166108

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.44		30 - 150		109%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150		113%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.49	*	55 - 111		123%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43	*	53 - 106		107%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.51		58 - 132		127%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2440	7.817				
1146-65-2	Naphthalene-d8	4770	10.611				
15067-26-2	Acenaphthene-d10	2410	14.452				
1517-22-2	Phenanthrene-d10	4920	17.186				
1719-03-5	Chrysene-d12	4140	21.367				
1520-96-3	Perylene-d12	4170	23.666				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035992.D
Acq On : 20 Jan 2025 11:03
Operator : RC/JU
Sample : PB166108BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166108BL

Quant Time: Jan 20 11:31:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2441	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4766	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2405	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	4916	0.400	ng	#-0.02
29) Chrysene-d12	21.367	240	4139	0.400	ng	-0.02
35) Perylene-d12	23.666	264	4174	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.397	112	2193	0.366	ng	-0.01
5) Phenol-d6	6.972	99	2602	0.350	ng	-0.01
8) Nitrobenzene-d5	8.966	82	1865	0.494	ng	-0.01
11) 2-Methylnaphthalene-d10	12.197	152	2795	0.438	ng	-0.02
14) 2,4,6-Tribromophenol	15.932	330	398	0.345	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4524	0.429	ng	-0.02
27) Fluoranthene-d10	19.220	212	5493	0.450	ng	-0.02
31) Terphenyl-d14	19.815	244	4187	0.508	ng	-0.02


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035992.D
Acq On : 20 Jan 2025 11:03
Operator : RC/JU
Sample : PB166108BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166108BL


Quant Time: Jan 20 11:31:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Instrument :
BNA_N
ClientSampleId :
PB166108BL







Tgt Ion:152 Resp: 2441
Ion Ratio Lower Upper
152 100
150 154.4 117.8 176.6
115 67.2 51.0 76.4



#4
2-Fluorophenol
Concen: 0.366 ng
RT: 5.397 min Scan# 307
Delta R.T. -0.011 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Tgt Ion:112 Resp: 2193
Ion Ratio Lower Upper
112 100
64 62.2 48.2 72.2
63 39.1 30.0 45.0









#5
Phenol-d6
Concen: 0.350 ng
RT: 6.972 min Scan# 51
Delta R.T. -0.011 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03




Instrument :
BNA_N
ClientSampleId :
PB166108BL


Tgt Ion:	99	Resp:	2602
Ion	Ratio	Lower	Upper
99	100		
42	29.4	23.5	35.3
71	45.2	35.5	53.3



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 926
Delta R.T. -0.015 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Tgt Ion:	136	Resp:	4766
Ion	Ratio	Lower	Upper
136	100		
137	12.0	10.6	15.8
54	10.5	9.8	14.6
68	5.9	6.6	10.0









#8
Nitrobenzene-d5
Concen: 0.494 ng
RT: 8.966 min Scan# 71
Delta R.T. -0.015 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03




Instrument :
BNA_N
ClientSampleId :
PB166108BL


Tgt Ion	Ratio	Lower	Upper
82	100		
128	43.2	36.9	55.3
54	62.7	50.4	75.6



#11
2-Methylnaphthalene-d10
Concen: 0.438 ng
RT: 12.197 min Scan# 1146
Delta R.T. -0.019 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Tgt Ion	Ratio	Lower	Upper
152	100		
151	21.5	17.0	25.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 14
Delta R.T. -0.010 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03




Instrument :
BNA_N
ClientSampleId :
PB166108BL


Tgt Ion:164 Resp: 2405
Ion Ratio Lower Upper
164 100
162 103.1 83.1 124.7
160 53.5 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 0.345 ng
RT: 15.932 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Tgt Ion:330 Resp: 398
Ion Ratio Lower Upper
330 100
332 91.2 77.2 115.8
141 43.0 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 0.429 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.021 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03




Instrument :
BNA_N
ClientSampleId :
PB166108BL


Tgt Ion:	172	Resp:	4524
Ion Ratio	Lower	Upper	
172	100		
171	36.3	30.5	45.7
170	25.6	20.8	31.2



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 1675
Delta R.T. -0.024 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Tgt Ion:	188	Resp:	4916
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	11.3	7.3	10.9





#27

Fluoranthene-d10

Concen: 0.450 ng

RT: 19.220 min Scan# 19

Delta R.T. -0.018 min

Lab File: BN035992.D

Acq: 20 Jan 2025 11:03

Instrument :

BNA_N

ClientSampleId :

PB166108BL





Tgt Ion:212 Resp: 5493

Ion Ratio Lower Upper

212 100

106 11.7 9.0 13.6

104 7.2 5.4 8.2



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.367 min Scan# 2255

Delta R.T. -0.017 min

Lab File: BN035992.D

Acq: 20 Jan 2025 11:03




Tgt Ion:240 Resp: 4139


Ion Ratio Lower Upper

240 100

120 13.3 11.1 16.7

236 29.2 24.6 36.8









#31
Terphenyl-d14
Concen: 0.508 ng
RT: 19.815 min Scan# 20
Delta R.T. -0.022 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03




Instrument :
BNA_N
ClientSampleId :
PB166108BL

Tgt Ion:244	Resp:	4187
Ion	Ratio	Lower Upper
244	100	
212	11.3	10.1 15.1
122	15.2	12.2 18.4



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.666 min Scan# 2802
Delta R.T. -0.028 min
Lab File: BN035992.D
Acq: 20 Jan 2025 11:03

Tgt Ion:264	Resp:	4174
Ion	Ratio	Lower Upper
264	100	
260	27.0	21.7 32.5
265	51.5	33.9 50.9#



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	PB166108BS	SDG No.:	Q1121
Lab Sample ID:	PB166108BS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035998.D	1	01/17/25 10:10	01/20/25 14:38	PB166108

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.38		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.43		30 - 150		106%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.44		30 - 150		111%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.53	*	55 - 111		133%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43		53 - 106		106%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		112%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2150	7.817				
1146-65-2	Naphthalene-d8	4360	10.611				
15067-26-2	Acenaphthene-d10	2170	14.452				
1517-22-2	Phenanthrene-d10	4380	17.186				
1719-03-5	Chrysene-d12	4020	21.367				
1520-96-3	Perylene-d12	4100	23.663				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN035998.D
 Acq On : 20 Jan 2025 14:38
 Operator : RC/JU
 Sample : PB166108BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB166108BS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/21/2025
 Supervised By :mohammad ahmed 01/21/2025

Quant Time: Jan 20 15:10:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2154	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4363	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2173	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	4379	0.400	ng	#-0.02
29) Chrysene-d12	21.367	240	4015	0.400	ng	-0.02
35) Perylene-d12	23.663	264	4099	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.398	112	2207	0.418	ng	-0.01
5) Phenol-d6	6.972	99	2534	0.386	ng	-0.01
8) Nitrobenzene-d5	8.956	82	1840	0.533	ng	-0.03
11) 2-Methylnaphthalene-d10	12.198	152	2490	0.426	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	469	0.450	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4059	0.426	ng	-0.02
27) Fluoranthene-d10	19.216	212	4814	0.443	ng	-0.02
31) Terphenyl-d14	19.815	244	3584	0.448	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.311	88	810	0.379	ng	# 74
3) n-Nitrosodimethylamine	3.614	42	1712	0.459	ng	94
6) bis(2-Chloroethyl)ether	7.239	93	2482	0.496	ng	97
9) Naphthalene	10.654	128	5221	0.426	ng	98
10) Hexachlorobutadiene	10.953	225	1557	0.392	ng	# 100
12) 2-Methylnaphthalene	12.274	142	3281	0.433	ng	99
16) Acenaphthylene	14.164	152	4595	0.449	ng	100
17) Acenaphthene	14.506	154	2953	0.440	ng	99
18) Fluorene	15.500	166	3538	0.479	ng	98
20) 4,6-Dinitro-2-methylph...	15.560	198	290	0.381	ng	97
21) 4-Bromophenyl-phenylether	16.380	248	1238	0.412	ng	# 84
22) Hexachlorobenzene	16.504	284	1631	0.398	ng	97
23) Atrazine	16.653	200	1013	0.503	ng	# 85
24) Pentachlorophenol	16.839	266	1173	0.809	ng	97
25) Phenanthrene	17.224	178	5634	0.441	ng	99
26) Anthracene	17.323	178	5206	0.448	ng	98
28) Fluoranthene	19.248	202	6317	0.424	ng	98
30) Pyrene	19.611	202	6564	0.403	ng	99
32) Benzo(a)anthracene	21.349	228	6038	0.427	ng	98
33) Chrysene	21.403	228	6127	0.414	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	3249	0.564	ng	99
36) Indeno(1,2,3-cd)pyrene	26.002	276	6731	0.416	ng	98
37) Benzo(b)fluoranthene	22.970	252	6002m	0.426	ng	
38) Benzo(k)fluoranthene	23.014	252	6168	0.442	ng	98
39) Benzo(a)pyrene	23.561	252	5550	0.455	ng	98
40) Dibenzo(a,h)anthracene	26.017	278	5340	0.415	ng	100
41) Benzo(g,h,i)perylene	26.712	276	5293	0.368	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035998.D
Acq On : 20 Jan 2025 14:38
Operator : RC/JU
Sample : PB166108BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB166108BS

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025


Quant Time: Jan 20 15:10:48 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Jan 02 15:39:17 2025


Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion: 152 Resp: 2154
Ion Ratio Lower Upper
152 100
150 155.3 117.8 176.6
115 65.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#2
1,4-Dioxane
Concen: 0.379 ng
RT: 3.311 min Scan# 18
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion: 88 Resp: 810
Ion Ratio Lower Upper
88 100
43 77.8 32.7 49.1#
58 86.3 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 0.459 ng
RT: 3.614 min Scan# 60
Delta R.T. -0.018 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion: 42 Resp: 171
Ion Ratio Lower Upper
42 100
74 83.6 62.2 93.2
44 10.3 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#4
2-Fluorophenol
Concen: 0.418 ng
RT: 5.398 min Scan# 307
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38


Tgt Ion:112 Resp: 2207
Ion Ratio Lower Upper
112 100
64 63.1 48.2 72.2
63 38.9 30.0 45.0





#5
Phenol-d6
Concen: 0.386 ng
RT: 6.972 min Scan# 511
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion: 99 Resp: 2534
Ion Ratio Lower Upper
99 100
42 30.3 23.5 35.3
71 45.9 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.496 ng
RT: 7.239 min Scan# 562
Delta R.T. -0.018 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38


Tgt Ion: 93 Resp: 2482
Ion Ratio Lower Upper
93 100
63 80.3 62.0 93.0
95 32.4 25.5 38.3





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 911
Delta R.T. -0.015 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion: 136 Resp: 436
Ion Ratio Lower Upper
136 100
137 12.3 10.6 15.8
54 9.5 9.8 14.6
68 6.6 6.6 10.0#




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#8
Nitrobenzene-d5
Concen: 0.533 ng
RT: 8.956 min Scan# 771
Delta R.T. -0.025 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion: 82 Resp: 1840
Ion Ratio Lower Upper
82 100
128 34.7 36.9 55.3#
54 67.9 50.4 75.6









#9
Naphthalene
Concen: 0.426 ng
RT: 10.654 min Scan# 91
Delta R.T. -0.025 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion:128 Resp: 522
Ion Ratio Lower Upper
128 100
129 12.5 10.6 16.0
127 15.2 12.8 19.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#10
Hexachlorobutadiene
Concen: 0.392 ng
RT: 10.953 min Scan# 958
Delta R.T. -0.025 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:225 Resp: 1557
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.426 ng
RT: 12.198 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion:152 Resp: 2490
Ion Ratio Lower Upper
152 100
151 21.6 17.0 25.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#12
2-Methylnaphthalene
Concen: 0.433 ng
RT: 12.274 min Scan# 1161
Delta R.T. -0.019 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38


Tgt Ion:142 Resp: 3281
Ion Ratio Lower Upper
142 100
141 90.2 71.9 107.9
115 48.7 39.6 59.4





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1490
Delta R.T. -0.009 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS




Tgt Ion:164 Resp: 217
Ion Ratio Lower Upper
164 100
162 102.0 83.1 124.7
160 51.1 46.0 69.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025




#14
2,4,6-Tribromophenol
Concen: 0.450 ng
RT: 15.933 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38



Tgt Ion:330 Resp: 469
Ion Ratio Lower Upper
330 100
332 98.3 77.2 115.8
141 47.3 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 0.426 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.020 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion:172 Resp: 4059
Ion Ratio Lower Upper
172 100
171 36.2 30.5 45.7
170 25.1 20.8 31.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#16
Acenaphthylene
Concen: 0.449 ng
RT: 14.164 min Scan# 1414
Delta R.T. -0.020 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:152 Resp: 4595
Ion Ratio Lower Upper
152 100
151 20.6 16.3 24.5
153 13.3 10.6 15.8









#17
Acenaphthene
Concen: 0.440 ng
RT: 14.506 min Scan# 1496
Delta R.T. -0.020 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion:154 Resp: 295
Ion Ratio Lower Upper
154 100
153 112.3 90.5 135.7
152 59.1 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#18
Fluorene
Concen: 0.479 ng
RT: 15.500 min Scan# 1539
Delta R.T. -0.009 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:166 Resp: 3538
Ion Ratio Lower Upper
166 100
165 102.9 80.6 120.8
167 13.7 11.4 17.0






#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 11
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :




PB166108BS




Tgt Ion:188 Resp: 4379
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.3 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025




#20
4,6-Dinitro-2-methylphenol
Concen: 0.381 ng
RT: 15.560 min Scan# 1544
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38



Tgt Ion:198 Resp: 290
Ion Ratio Lower Upper
198 100
51 77.9 64.3 96.5
105 59.3 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.412 ng
RT: 16.380 min Scan# 1610
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion:248 Resp: 1238
Ion Ratio Lower Upper
248 100
250 91.1 76.8 115.2
141 105.2 63.6 95.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#22
Hexachlorobenzene
Concen: 0.398 ng
RT: 16.504 min Scan# 1620
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:284 Resp: 1631
Ion Ratio Lower Upper
284 100
142 41.8 31.4 47.0
249 36.2 27.8 41.8









#23
Atrazine
Concen: 0.503 ng
RT: 16.653 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS

Tgt Ion:200 Resp: 101
Ion Ratio Lower Upper
200 100
173 27.6 35.4 53.0
215 48.7 42.4 63.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#24
Pentachlorophenol
Concen: 0.809 ng
RT: 16.839 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38


Tgt Ion:266 Resp: 1173
Ion Ratio Lower Upper
266 100
264 64.5 49.9 74.9
268 65.8 50.2 75.2





#25
Phenanthrene
Concen: 0.441 ng
RT: 17.224 min Scan# 1686
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion:178 Resp: 5634
Ion Ratio Lower Upper
178 100
176 19.9 15.9 23.9
179 15.3 12.9 19.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#26
Anthracene
Concen: 0.448 ng
RT: 17.323 min Scan# 1686
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:178 Resp: 5206
Ion Ratio Lower Upper
178 100
176 19.4 15.0 22.6
179 15.3 13.0 19.6





#27

Fluoranthene-d10

Concen: 0.443 ng

RT: 19.216 min Scan# 1940

Delta R.T. -0.022 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38





Instrument :

BNA_N

ClientSampleId :

PB166108BS

Tgt Ion:	212	Resp:	4814
Ion	Ratio	Lower	Upper
212	100		
106	12.3	9.0	13.6
104	7.5	5.4	8.2

Manual Integrations
APPROVEDReviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025

#28

Fluoranthene

Concen: 0.424 ng




RT: 19.248 min Scan# 1940


Delta R.T. -0.017 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Tgt Ion:	202	Resp:	6317
Ion	Ratio	Lower	Upper
202	100		
101	10.2	7.2	10.8
203	17.0	13.9	20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.367 min Scan# 21

Delta R.T. -0.017 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38





Instrument :

BNA_N

ClientSampleId :

PB166108BS

Tgt Ion:240	Resp:	401
Ion Ratio	Lower	Upper
240	100	
120	12.7	11.1 16.7
236	28.7	24.6 36.8

Manual Integrations
APPROVEDReviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025

#30

Pyrene

Concen: 0.403 ng




RT: 19.611 min Scan# 2018


Delta R.T. -0.017 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38


Tgt Ion:202	Resp:	6564
Ion Ratio	Lower	Upper
202	100	
200	21.8	17.0 25.6
203	17.8	14.6 22.0





#31
Terphenyl-d14
Concen: 0.448 ng
RT: 19.815 min Scan# 2062
Delta R.T. -0.022 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38



Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion:244 Resp: 3584
Ion Ratio Lower Upper
244 100
212 11.0 10.1 15.1
122 14.4 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#32
Benzo(a)anthracene
Concen: 0.427 ng
RT: 21.349 min Scan# 2253
Delta R.T. -0.026 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:228 Resp: 6038
Ion Ratio Lower Upper
228 100
226 28.0 22.7 34.1
229 19.9 17.1 25.7





#33

Chrysene

Concen: 0.414 ng

RT: 21.403 min Scan# 21

Delta R.T. -0.017 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS

Tgt Ion:228 Resp: 612

Ion Ratio Lower Upper

228 100

226 31.4 25.2 37.8





229 19.7 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.564 ng

RT: 21.295 min Scan# 2247

Delta R.T. -0.026 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38




Tgt Ion:149 Resp: 3249


Ion Ratio Lower Upper

149 100

167 26.5 21.4 32.0


279 3.3 3.0 4.6





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.663 min Scan# 21
Delta R.T. -0.031 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion:264 Resp: 4099
Ion Ratio Lower Upper
264 100
260 27.1 21.7 32.5
265 66.0 33.9 50.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.416 ng
RT: 26.002 min Scan# 3601
Delta R.T. -0.046 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:276 Resp: 6731
Ion Ratio Lower Upper
276 100
138 25.0 18.9 28.3
277 25.4 19.5 29.3





#37

Benzo(b)fluoranthene

Concen: 0.426 ng m

RT: 22.970 min Scan# 2579

Delta R.T. -0.028 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS





Tgt Ion:252	Resp:	600
Ion Ratio	Lower	Upper
252	100	
253	27.2	20.1 30.1
125	15.0	9.9 14.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025



#38

Benzo(k)fluoranthene

Concen: 0.442 ng




RT: 23.014 min Scan# 2579


Delta R.T. -0.031 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Tgt Ion:252	Resp:	6168
Ion Ratio	Lower	Upper
252	100	
253	26.3	20.5 30.7
125	14.6	10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.455 ng

RT: 23.561 min Scan# 21

Delta R.T. -0.028 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :





PB166108BS

Tgt Ion:	252	Resp:	5550
Ion Ratio	Lower	Upper	
252	100		
253	27.3	21.5	32.3
125	17.1	12.6	19.0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025

Supervised By :mohammad ahmed 01/21/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.415 ng




RT: 26.017 min Scan# 3606


Delta R.T. -0.054 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38


Tgt Ion:	278	Resp:	5340
Ion Ratio	Lower	Upper	
278	100		
139	19.8	15.9	23.9
279	28.5	23.0	34.4





#41
Benzo(g,h,i)perylene
Concen: 0.368 ng
RT: 26.712 min Scan# 3844
Delta R.T. -0.051 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38



Instrument :
BNA_N
ClientSampleId :
PB166108BS



Tgt Ion: 276 Resp: 529
Ion Ratio Lower Upper
276 100
277 27.1 22.8 34.2
138 24.3 20.1 30.1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2025
Supervised By :mohammad ahmed 01/21/2025



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	PB166108BSD	SDG No.:	Q1121
Lab Sample ID:	PB166108BSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035999.D	1	01/17/25 10:10	01/20/25 15:14	PB166108

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.39		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.40		30 - 150		100%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		95%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.51	*	55 - 111		126%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43	*	53 - 106		107%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.44		58 - 132		111%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2640	7.817				
1146-65-2	Naphthalene-d8	5200	10.611				
15067-26-2	Acenaphthene-d10	2410	14.452				
1517-22-2	Phenanthrene-d10	4160	17.186				
1719-03-5	Chrysene-d12	3270	21.367				
1520-96-3	Perylene-d12	3540	23.66				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN035999.D
 Acq On : 20 Jan 2025 15:14
 Operator : RC/JU
 Sample : PB166108BSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB166108BSD

Quant Time: Jan 20 15:40:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2639	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	5199	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2411	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	4157	0.400	ng	-0.02
29) Chrysene-d12	21.367	240	3274	0.400	ng	-0.02
35) Perylene-d12	23.660	264	3543	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.398	112	2508	0.387	ng	-0.01
5) Phenol-d6	6.972	99	2821	0.351	ng	-0.01
8) Nitrobenzene-d5	8.956	82	2084	0.506	ng	-0.03
11) 2-Methylnaphthalene-d10	12.198	152	2798	0.402	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	446	0.385	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4533	0.428	ng	-0.02
27) Fluoranthene-d10	19.216	212	3942	0.382	ng	-0.02
31) Terphenyl-d14	19.815	244	2884	0.442	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.318	88	1031	0.393	ng	# 80
3) n-Nitrosodimethylamine	3.621	42	2072	0.453	ng	97
6) bis(2-Chloroethyl)ether	7.239	93	2858	0.467	ng	96
9) Naphthalene	10.654	128	6025	0.413	ng	97
10) Hexachlorobutadiene	10.953	225	1855	0.391	ng	# 99
12) 2-Methylnaphthalene	12.269	142	3657	0.405	ng	99
16) Acenaphthylene	14.164	152	4921	0.433	ng	99
17) Acenaphthene	14.506	154	3151	0.423	ng	99
18) Fluorene	15.500	166	3576	0.437	ng	98
20) 4,6-Dinitro-2-methylph...	15.560	198	250	0.346	ng	92
21) 4-Bromophenyl-phenylether	16.379	248	1245	0.437	ng	90
22) Hexachlorobenzene	16.504	284	1608	0.414	ng	96
23) Atrazine	16.653	200	882	0.461	ng	# 84
24) Pentachlorophenol	16.839	266	997	0.725	ng	99
25) Phenanthrene	17.223	178	5073	0.418	ng	99
26) Anthracene	17.323	178	4668	0.423	ng	99
28) Fluoranthene	19.248	202	5208	0.368	ng	99
30) Pyrene	19.611	202	5319	0.400	ng	99
32) Benzo(a)anthracene	21.349	228	4804	0.416	ng	98
33) Chrysene	21.403	228	4961	0.411	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	2475	0.527	ng	97
36) Indeno(1,2,3-cd)pyrene	25.996	276	6118	0.438	ng	98
37) Benzo(b)fluoranthene	22.967	252	4965	0.408	ng	# 95
38) Benzo(k)fluoranthene	23.014	252	5132	0.426	ng	97
39) Benzo(a)pyrene	23.561	252	4695	0.446	ng	98
40) Dibenzo(a,h)anthracene	26.011	278	4789	0.431	ng	100
41) Benzo(g,h,i)perylene	26.712	276	4823	0.388	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035999.D
Acq On : 20 Jan 2025 15:14
Operator : RC/JU
Sample : PB166108BSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166108BSD

Quant Time: Jan 20 15:40:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:152 Resp: 2639
Ion Ratio Lower Upper
152 100
150 155.1 117.8 176.6
115 65.6 51.0 76.4



#2
1,4-Dioxane
Concen: 0.393 ng
RT: 3.318 min Scan# 19
Delta R.T. -0.004 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion: 88 Resp: 1031
Ion Ratio Lower Upper
88 100
43 61.8 32.7 49.1#
58 89.9 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 0.453 ng
RT: 3.621 min Scan# 61
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion: 42 Resp: 2072
Ion Ratio Lower Upper
42 100
74 80.5 62.2 93.2
44 7.2 7.1 10.7



#4
2-Fluorophenol
Concen: 0.387 ng
RT: 5.398 min Scan# 307
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion: 112 Resp: 2508
Ion Ratio Lower Upper
112 100
64 63.0 48.2 72.2
63 39.1 30.0 45.0









#5
Phenol-d6
Concen: 0.351 ng
RT: 6.972 min Scan# 51
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion: 99 Resp: 2821
Ion Ratio Lower Upper
99 100
42 31.7 23.5 35.3
71 46.8 35.5 53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.467 ng
RT: 7.239 min Scan# 562
Delta R.T. -0.018 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion: 93 Resp: 2858
Ion Ratio Lower Upper
93 100
63 81.4 62.0 93.0
95 32.7 25.5 38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 91
Delta R.T. -0.015 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:	136	Resp:	5199
Ion	Ratio	Lower	Upper
136	100		
137	12.1	10.6	15.8
54	9.2	9.8	14.6#
68	6.1	6.6	10.0#



#8
Nitrobenzene-d5
Concen: 0.506 ng
RT: 8.956 min Scan# 771
Delta R.T. -0.025 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:	82	Resp:	2084
Ion	Ratio	Lower	Upper
82	100		
128	33.5	36.9	55.3#
54	70.7	50.4	75.6









#9
Naphthalene
Concen: 0.413 ng
RT: 10.654 min Scan# 91
Delta R.T. -0.025 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:	128	Resp:	6025
Ion	Ratio	Lower	Upper
128	100		
129	12.1	10.6	16.0
127	14.9	12.8	19.2



#10
Hexachlorobutadiene
Concen: 0.391 ng
RT: 10.953 min Scan# 958
Delta R.T. -0.025 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:	225	Resp:	1855
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.5	77.3









#11
2-Methylnaphthalene-d10
Concen: 0.402 ng
RT: 12.198 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:152 Resp: 2798
Ion Ratio Lower Upper
152 100
151 21.1 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.405 ng
RT: 12.269 min Scan# 1160
Delta R.T. -0.024 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:142 Resp: 3657
Ion Ratio Lower Upper
142 100
141 90.5 71.9 107.9
115 50.7 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 14
Delta R.T. -0.010 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.4	83.1	124.7
160	52.5	46.0	69.0



#14
2,4,6-Tribromophenol
Concen: 0.385 ng
RT: 15.933 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.2	115.8
141	50.2	31.6	47.4









#15
2-Fluorobiphenyl
Concen: 0.428 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.020 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:	172	Resp:	4533
Ion Ratio	Lower	Upper	
172	100		
171	36.3	30.5	45.7
170	25.1	20.8	31.2



#16
Acenaphthylene
Concen: 0.433 ng
RT: 14.164 min Scan# 1414
Delta R.T. -0.020 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:	152	Resp:	4921
Ion Ratio	Lower	Upper	
152	100		
151	20.6	16.3	24.5
153	12.8	10.6	15.8









#17
Acenaphthene
Concen: 0.423 ng
RT: 14.506 min Scan# 1446
Delta R.T. -0.020 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.6	90.5	135.7
152	59.7	48.6	73.0



#18
Fluorene
Concen: 0.437 ng
RT: 15.500 min Scan# 1539
Delta R.T. -0.010 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.0	80.6	120.8
167	13.6	11.4	17.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 11
Delta R.T. -0.024 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:188 Resp: 4157
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.4 7.3 10.9



#20
4,6-Dinitro-2-methylphenol
Concen: 0.346 ng
RT: 15.560 min Scan# 1544
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:198 Resp: 250
Ion Ratio Lower Upper
198 100
51 71.9 64.3 96.5
105 57.1 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.437 ng
RT: 16.379 min Scan# 1610
Delta R.T. -0.024 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion	Ratio	Lower	Upper
248	100		
250	92.3	76.8	115.2
141	94.7	63.6	95.4



#22
Hexachlorobenzene
Concen: 0.414 ng
RT: 16.504 min Scan# 1620
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion	Ratio	Lower	Upper
284	100		
142	43.2	31.4	47.0
249	35.3	27.8	41.8









#23
Atrazine
Concen: 0.461 ng
RT: 16.653 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:200 Resp: 882
Ion Ratio Lower Upper
200 100
173 27.3 35.4 53.0#
215 47.6 42.4 63.6



#24
Pentachlorophenol
Concen: 0.725 ng
RT: 16.839 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:266 Resp: 997
Ion Ratio Lower Upper
266 100
264 61.8 49.9 74.9
268 63.0 50.2 75.2









#25
Phenanthrene
Concen: 0.418 ng
RT: 17.223 min Scan# 1686
Delta R.T. -0.024 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.9	15.9	23.9
179	15.4	12.9	19.3



#26
Anthracene
Concen: 0.423 ng
RT: 17.323 min Scan# 1686
Delta R.T. -0.011 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.0	22.6
179	15.3	13.0	19.6









#27
Fluoranthene-d10
Concen: 0.382 ng
RT: 19.216 min Scan# 1940
Delta R.T. -0.022 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion	Ratio	Lower	Upper
212	100		
106	11.8	9.0	13.6
104	6.9	5.4	8.2



#28
Fluoranthene
Concen: 0.368 ng
RT: 19.248 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion	Ratio	Lower	Upper
202	100		
101	9.5	7.2	10.8
203	17.3	13.9	20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.367 min Scan# 21
Delta R.T. -0.017 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion	Ratio	Lower	Upper
240	100		
120	12.3	11.1	16.7
236	29.0	24.6	36.8



#30
Pyrene
Concen: 0.400 ng
RT: 19.611 min Scan# 2018
Delta R.T. -0.017 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.0	25.6
203	17.6	14.6	22.0









#31
Terphenyl-d14
Concen: 0.442 ng
RT: 19.815 min Scan# 2062
Delta R.T. -0.022 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:244	Resp:	2884
Ion Ratio	Lower	Upper
244	100	
212	11.2	10.1 15.1
122	14.8	12.2 18.4



#32
Benzo(a)anthracene
Concen: 0.416 ng
RT: 21.349 min Scan# 2253
Delta R.T. -0.026 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:228	Resp:	4804
Ion Ratio	Lower	Upper
228	100	
226	29.3	22.7 34.1
229	20.5	17.1 25.7





#33

Chrysene

Concen: 0.411 ng

RT: 21.403 min Scan# 21

Delta R.T. -0.017 min

Lab File: BN035999.D

Acq: 20 Jan 2025 15:14

Instrument :

BNA_N

ClientSampleId :

PB166108BSD





Tgt Ion:228 Resp: 4961

Ion Ratio Lower Upper

228 100

226 30.9 25.2 37.8

229 19.7 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.527 ng

RT: 21.295 min Scan# 2247

Delta R.T. -0.026 min

Lab File: BN035999.D

Acq: 20 Jan 2025 15:14




Tgt Ion:149 Resp: 2475


Ion Ratio Lower Upper

149 100

167 28.7 21.4 32.0

279 3.6 3.0 4.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.660 min Scan# 21
Delta R.T. -0.034 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:264 Resp: 3543
Ion Ratio Lower Upper
264 100
260 27.1 21.7 32.5
265 66.8 33.9 50.9#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.438 ng
RT: 25.996 min Scan# 3599
Delta R.T. -0.052 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:276 Resp: 6118
Ion Ratio Lower Upper
276 100
138 24.6 18.9 28.3
277 25.3 19.5 29.3









#37
Benzo(b)fluoranthene
Concen: 0.408 ng
RT: 22.967 min Scan# 21
Delta R.T. -0.031 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14




Instrument :
BNA_N
ClientSampleId :
PB166108BSD


Tgt Ion:252 Resp: 4965
Ion Ratio Lower Upper
252 100
253 26.8 20.1 30.1
125 15.1 9.9 14.9#



#38
Benzo(k)fluoranthene
Concen: 0.426 ng
RT: 23.014 min Scan# 2579
Delta R.T. -0.031 min
Lab File: BN035999.D
Acq: 20 Jan 2025 15:14

Tgt Ion:252 Resp: 5132
Ion Ratio Lower Upper
252 100
253 26.7 20.5 30.7
125 15.0 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.446 ng

RT: 23.561 min Scan# 21

Delta R.T. -0.028 min

Lab File: BN035999.D

Acq: 20 Jan 2025 15:14





Instrument :

BNA_N

ClientSampleId :

PB166108BSD

Tgt Ion:	252	Resp:	4695
Ion	Ratio	Lower	Upper
252	100		
253	27.8	21.5	32.3
125	16.6	12.6	19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.431 ng




RT: 26.011 min Scan# 3604


Delta R.T. -0.060 min

Lab File: BN035999.D

Acq: 20 Jan 2025 15:14

Tgt Ion:	278	Resp:	4789
Ion	Ratio	Lower	Upper
278	100		
139	19.7	15.9	23.9
279	28.7	23.0	34.4





#41

Benzo(g,h,i)perylene

Concen: 0.388 ng

RT: 26.712 min Scan# 3844

Delta R.T. -0.052 min

Lab File: BN035999.D


Acq: 20 Jan 2025 15:14

Instrument :

BNA_N

ClientSampleId :

PB166108BSD





Tgt Ion:276 Resp: 4823

Ion Ratio Lower Upper

276 100

277 26.2 22.8 34.2

138 23.2 20.1 30.1





Manual Integration Report

Sequence:	BN010225	Instrument	BNA_n
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

Manual Integration Report

Sequence:

BN012025

Instrument

BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB166108BS	BN035998.D	Benzo(b)fluoranthene	yogesh	1/21/2025 7:29:52 AM	mohammad	1/21/2025 7:42:00 AM	Peak Integrated by Software
SSTDCCC0.4	BN036000.D	Benzo(b)fluoranthene	yogesh	1/21/2025 7:29:55 AM	mohammad	1/21/2025 7:42:00 AM	Peak Integrated by Software

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN010225

Review By	Jagrut	Review On	1/3/2025 3:37:20 PM
Supervise By	mohammad	Supervise On	1/6/2025 8:20:47 AM
SubDirectory	BN010225	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn010225
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657		
CCC	SP6661		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN035870.D	02 Jan 2025 10:49	RC/JU	Ok
2	SSTDICC0.1	BN035871.D	02 Jan 2025 11:28	RC/JU	Ok
3	SSTDICC0.2	BN035872.D	02 Jan 2025 12:04	RC/JU	Ok
4	SSTDICCC0.4	BN035873.D	02 Jan 2025 12:40	RC/JU	Ok
5	SSTDICC0.8	BN035874.D	02 Jan 2025 13:16	RC/JU	Ok
6	SSTDICC1.6	BN035875.D	02 Jan 2025 13:52	RC/JU	Ok
7	SSTDICC3.2	BN035876.D	02 Jan 2025 14:28	RC/JU	Ok
8	SSTDICC5.0	BN035877.D	02 Jan 2025 15:04	RC/JU	Ok
9	SSTDICV0.4	BN035878.D	02 Jan 2025 15:51	RC/JU	Ok
10	PB165828BL	BN035879.D	02 Jan 2025 16:41	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN012025

Review By	yogesh	Review On	1/21/2025 7:30:11 AM
Supervise By	mohammad	Supervise On	1/21/2025 7:42:00 AM
SubDirectory	BN012025	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn010225
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657		
CCC	SP6661		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN035990.D	20 Jan 2025 09:13	RC/JU	Ok
2	SSTDCCC0.4	BN035991.D	20 Jan 2025 10:27	RC/JU	Ok
3	PB166108BL	BN035992.D	20 Jan 2025 11:03	RC/JU	Ok
4	Q1112-02	BN035993.D	20 Jan 2025 11:39	RC/JU	Ok
5	Q1112-04	BN035994.D	20 Jan 2025 12:15	RC/JU	Ok,M
6	Q1113-02	BN035995.D	20 Jan 2025 12:51	RC/JU	Ok,M
7	Q1113-06	BN035996.D	20 Jan 2025 13:27	RC/JU	Ok
8	Q1121-01	BN035997.D	20 Jan 2025 14:03	RC/JU	Ok
9	PB166108BS	BN035998.D	20 Jan 2025 14:38	RC/JU	Ok,M
10	PB166108BSD	BN035999.D	20 Jan 2025 15:14	RC/JU	Ok
11	SSTDCCC0.4	BN036000.D	20 Jan 2025 16:11	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN010225

Review By	Jagrut	Review On	1/3/2025 3:37:20 PM
Supervise By	mohammad	Supervise On	1/6/2025 8:20:47 AM
SubDirectory	BN010225	HP Acquire Method	BNA_N, 8270_HP Processing Method bn010225

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657
CCC	SP6661
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN035870.D	02 Jan 2025 10:49		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN035871.D	02 Jan 2025 11:28		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN035872.D	02 Jan 2025 12:04		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN035873.D	02 Jan 2025 12:40	The Calibration is Good For DOD	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN035874.D	02 Jan 2025 13:16		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN035875.D	02 Jan 2025 13:52		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN035876.D	02 Jan 2025 14:28		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN035877.D	02 Jan 2025 15:04		RC/JU	Ok
9	SSTDICV0.4	ICVBN010225	BN035878.D	02 Jan 2025 15:51		RC/JU	Ok
10	PB165828BL	PB165828BL	BN035879.D	02 Jan 2025 16:41	Analyzed for Contamination check	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN012025

Review By	yogesh	Review On	1/21/2025 7:30:11 AM
Supervise By	mohammad	Supervise On	1/21/2025 7:42:00 AM
SubDirectory	BN012025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn010225

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6663,SP6662,SP6661,SP6660,SP6659,SP6658,SP6657
CCC	SP6661
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN035990.D	20 Jan 2025 09:13		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN035991.D	20 Jan 2025 10:27		RC/JU	Ok
3	PB166108BL	PB166108BL	BN035992.D	20 Jan 2025 11:03		RC/JU	Ok
4	Q1112-02	TOWER-1	BN035993.D	20 Jan 2025 11:39		RC/JU	Ok
5	Q1112-04	TOWER-2	BN035994.D	20 Jan 2025 12:15		RC/JU	Ok,M
6	Q1113-02	DSN002	BN035995.D	20 Jan 2025 12:51		RC/JU	Ok,M
7	Q1113-06	DSN003	BN035996.D	20 Jan 2025 13:27		RC/JU	Ok
8	Q1121-01	RW10A-20250116	BN035997.D	20 Jan 2025 14:03		RC/JU	Ok
9	PB166108BS	PB166108BS	BN035998.D	20 Jan 2025 14:38		RC/JU	Ok,M
10	PB166108BSD	PB166108BSD	BN035999.D	20 Jan 2025 15:14		RC/JU	Ok
11	SSTDCCC0.4	SSTDCCC0.4EC	BN036000.D	20 Jan 2025 16:11		RC/JU	Ok,M

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-20

Clean Up SOP #: N/A

Matrix : Water

Weigh By: N/A

Balance check: N/A

Balance ID: N/A

pH Strip Lot#: E3574

Extraction By: RS

Filter By: RS

pH Meter ID: N/A

Hood ID: 4,5,6,7

Extraction Start Date : 01/17/2025

Extraction Start Time : 10:10

Extraction End Date : 01/17/2025

Extraction End Time : 15:10

Concentration By: EH

Supervisor By : rajesh

Extraction Method: ☒ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Standardized Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6616
Surrogate	1.0ML	0.4 PPM	SP6666
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3871
Baked Na2SO4	N/A	EP2577
10N NaOH	N/A	EP2559
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted <2 with 1:1 H2SO4 & >11 with 10 N NaOH.

KD Bath ID: Water bath -01,02

KD Bath Temperature: 60 °C

Envap ID: NEVAP-02

Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
01/17/25	RP (Ext Lab)	RC/SVOC
15:15	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 01/17/2025

Sample ID	Client Sample ID	Test	g / (ml)	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB166108BL	SBLK108	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			SEP-01
PB166108BS	SLCS108	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			2
PB166108BS	SLCSD108	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1			3
Q1112-02	TOWER-1	SVOC-SIMGrou p1	980	6	RUPESH	ritesh	1	B		4
Q1112-04	TOWER-2	SVOC-SIMGrou p1	990	6	RUPESH	ritesh	1	B		5
Q1113-02	DSN002	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1	B		6
Q1113-06	DSN003	SVOC-SIMGrou p1	1000	6	RUPESH	ritesh	1	B		7
Q1121-01	RW10A-20250116	SVOC-SIMGrou p1	980	6	RUPESH	ritesh	1	C		8

* Extracts relinquished on the same date as received.

Q1121-SVOC-SIMGroup1

2
1/17/25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1121

WorkList ID : 186987

Department : Extraction

Date : 01-17-2025 10:07:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1112-02	TOWER-1	Water	SVOC-SIMGroup1	Cool 4 deg C	PSEG04	M11	01/16/2025	8270-Modified
Q1112-04	TOWER-2	Water	SVOC-SIMGroup1	Cool 4 deg C	PSEG04	M11	01/16/2025	8270-Modified
Q1113-02	DSN002	Water	SVOC-SIMGroup1	Cool 4 deg C	PSEG04	M11	01/16/2025	8270-Modified
Q1113-06	DSN003	Water	SVOC-SIMGroup1	Cool 4 deg C	PSEG04	M11	01/16/2025	8270-Modified
Q1121-01	RW10A-20250116	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	M11	01/16/2025	8270-Modified

Date/Time 01/17/25 10:08
 Raw Sample Received by: RS (Ext Lab)
 Raw Sample Relinquished by: JDCSM

Date/Time 01/17/25 10:35
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: RS (Ext Lab)
 350 of 468

Prep Standard - Chemical Standard Summary

Order ID : Q1121
Test : SVOC-SIMGroup1
Prepbatch ID : PB166108,
Sequence ID/Qc Batch ID: BN012025,

Standard ID :
EP2559,EP2565,EP2577,SP6616,SP6629,SP6656,SP6657,SP6658,SP6659,SP6660,SP6661,SP6662,SP6663,SP666
6,SP6682,SP6683,SP6684,SP6717,

Chemical ID :
1ul/100ul
sample,E3551,E3657,E3788,E3791,E3817,E3818,E3828,E3871,M5173,S10103,S10246,S11011,S11074,S11097,S114
94,S11792,S11831,S12077,S12079,S12105,S12113,S12126,S12142,S12189,S12207,S12208,S12314,S12328,S12453
,S12469,S12517,W3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2559	11/14/2024	05/14/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 11/14/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2577	01/06/2025	07/01/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 01/06/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6616	09/06/2024	02/12/2025	Rahul Chavli	None	None	mohammad ahmed 09/11/2024

FROM 0.00160ml of S11011 + 0.02000ml of S11792 + 0.04000ml of S12105 + 0.04000ml of S12126 + 0.04000ml of S12453 +
99.85840ml of E3788 = Final Quantity: 100.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6629	09/12/2024	03/04/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								10/14/2024

FROM 0.10000ml of S12314 + 4.90000ml of E3791 = Final Quantity: 5.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6656	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.03350ml of S10103 + 0.05000ml of S11494 + 0.05000ml of S12079 + 0.12500ml of S11831 + 0.12500ml of S12113 + 0.20000ml of S12077 + 0.25000ml of S11097 + 24.16650ml of E3817 = Final Quantity: 25.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6657	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
11/28/2024								

FROM 0.50000ml of E3817 + 0.01000ml of SP6629 + 0.50000ml of SP6656 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6658	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
11/28/2024								

FROM 0.68000ml of E3817 + 0.01000ml of SP6629 + 0.32000ml of SP6656 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6659	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.84000ml of E3817 + 0.01000ml of SP6629 + 0.16000ml of SP6656 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6660	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.92000ml of E3817 + 0.01000ml of SP6629 + 0.08000ml of SP6656 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6661	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.96000ml of E3817 + 0.01000ml of SP6629 + 0.04000ml of SP6656 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6662	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.50000ml of E3817 + 0.01000ml of SP6629 + 0.50000ml of SP6661 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6663	10/24/2024	02/08/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.75000ml of E3817 + 0.01000ml of SP6629 + 0.25000ml of SP6661 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6666	10/29/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								11/28/2024

FROM 0.00800ml of S12189 + 0.01600ml of S12207 + 0.04000ml of S11831 + 199.93600ml of E3818 = Final Quantity: 200.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6682	11/15/2024	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								12/03/2024

FROM 0.10000ml of S12328 + 4.90000ml of E3828 = Final Quantity: 5.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6683	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
								12/03/2024

FROM SOURCE
0.00630ml of S12189 + 0.01280ml of S12208 + 0.03200ml of S11074 + 0.03200ml of S11831 + 0.06400ml of S12142 +
0.06400ml of S12469 + 0.06400ml of S12517 + 19.72490ml of E3828 = Final Quantity: 20.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6684	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
SOURCE FROM 0.87500ml of E3828 + 0.01000ml of SP6682 + 0.12500ml of SP6683 = Final Quantity: 1.010 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTTP 8270E	SP6717	01/15/2025	03/31/2025	Rahul Chavli	None	None	Yogesh Patel
FROM 1.00000ml of S10246 + 19.00000ml of E3871 = Final Quantity: 20.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	04/23/2025	08/13/2024 / Rajesh	08/13/2024 / Rajesh	E3788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G2362009	03/09/2025	09/09/2024 / Rajesh	09/03/2024 / Rajesh	E3791

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24H2762011	04/09/2025	10/09/2024 / Rajesh	10/09/2024 / Rajesh	E3817

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	04/23/2025	10/23/2024 / Rajesh	10/09/2024 / Rajesh	E3818

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G0862003	05/09/2025	11/09/2024 / Rajesh	11/04/2024 / Rajesh	E3828

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	07/14/2025	01/14/2025 / Rajesh	12/27/2024 / Rajesh	E3871

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	02/08/2025	08/08/2024 / Jagrut	12/09/2021 / Christian	S10103

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0182667	03/31/2025	01/15/2025 / Rahul	03/18/2022 / Christian	S10246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0193449	02/20/2025	08/20/2024 / yogesh	01/13/2023 / Christian	S11011

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0187043	05/15/2025	11/15/2024 / Jagrut	02/06/2023 / Christian	S11074

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	495831	02/08/2025	08/08/2024 / Jagrut	02/07/2023 / Christian	S11097

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	02/08/2025	08/08/2024 / Jagrut	08/11/2023 / Yogesh	S11494

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/21/2025	08/21/2024 / Jagrut	11/21/2023 / Rahul	S11792

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	04/11/2025	10/11/2024 / Jagrut	11/21/2023 / rahul	S11831

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	02/08/2025	08/08/2024 / Jagrut	01/31/2024 / Rahul	S12077

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	04/24/2025	10/24/2024 / Jagrut	01/31/2024 / Rahul	S12079

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12105

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	02/09/2025	08/09/2024 / Jagrut	03/08/2024 / Rahul	S12113

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12126

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	04/30/2025	11/14/2024 / anahy	03/15/2024 / Rahul	S12142

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml, methanol, 5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12189

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0206381	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12207

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0206381	05/15/2025	11/15/2024 / Jagrut	03/15/2024 / Rahul	S12208

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0206540	03/04/2025	09/04/2024 / anahy	05/30/2024 / Rahul	S12314

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0206540	05/13/2025	11/13/2024 / anahy	05/30/2024 / Rahul	S12328

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12453

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12469

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12517

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By: _____

Erica Castiglione
Chemist



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Receivalon
02/07/23 by CG

Sheet 511096
to
511099

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	495831	≤ -10 °C	Methylene Chloride	10/30/2027	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1003 ± 17.27
acenaphthylene	208-96-8	97.6	14.290.1P	999.8 ± 17.22
aniline	62-53-3	99.9	64.7.1P	995 ± 17.13
anthracene	120-12-7	99.5	15.7.1P	1001 ± 17.24
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 17.21
benzo[a]anthracene	56-55-3	100	16.7.3P	1001 ± 17.24
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1001 ± 19.91
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 17.92
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 19.88
benzo[a]pyrene	50-32-8	97	20.286.2P	999.1 ± 26.35
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 17.24
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	999.7 ± 17.89
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1001 ± 17.23
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.13P	999.5 ± 17.89
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 17.21
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 19.86
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1P	999.1 ± 17.2
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 19.58
carbazole	86-74-8	99.4	239.7.2P	1000 ± 17.22

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Brianna Smith

Certified By: _____

Brianna Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 495831

Expiration Date: 10/30/2027

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	1000 ± 17.22
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	1000 ± 17.22
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	1002 ± 17.25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:



Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
03/10/22
by
CG
S10242
to
S10247

Catalog No. : 31615 **Lot No.:** A0182667
Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2025 **Storage:** 10°C or colder
Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol CAS # 87-86-5 (Lot 211229RSR) Purity 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 (Lot Q117-147) Purity 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine CAS # 92-87-5 (Lot 211228JLM) Purity 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022

Balance: B345965662

Marlene Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
02/08/23

b1
CG

S 11071
to
S 11075

Catalog No. : 31853 Lot No.: A0187043
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : July 31, 2027 Storage: 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	1,4-Dioxane CAS # 123-91-1 Purity 99% (Lot SHBN5929)	2,019.0 µg/mL	+/- 11.8486 µg/mL Gravimetric +/- 43.2570 µg/mL Unstressed +/- 44.5129 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko - Operations Tech I

Date Mixed: 07-Jul-2022

Balance: 1128360905

Marlina Cowan - Operations Tech II ARM QC

Date Passed: 12-Jul-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0193449
Description : Custom Pentachlorophenol Standard
Custom Pentachlorophenol Standard 25,000µg/mL, Methanol,
1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : January 31, 2026 **Storage:** 10°C or colder
Ship: Ambient

Received on
01/13/23
by
CG
S11011
+0
S11015

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP221012	99%	25,050.0 µg/mL	+/- 778.6378

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Jan-2023

Balance: B442140311

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 8/13/24

E 3788

Ken Koehnlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24G2362009
Manufactured Date: 2024-06-10
Expiration Date: 2025-09-09
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μ eq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24F10024

E 3791

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700

Page 1 of 1

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24H2762011
Manufactured Date: 2024-06-05
Expiration Date: 2025-09-04
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	5
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3817

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected forwater)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 10/9/24

E 3818

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™



Material No.: 9266-A4

Batch No.: 24J0862003

Manufactured Date: 2024-09-12

Expiration Date: 2025-12-12

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3828

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0%
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	0.0
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	0.01%

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3871

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓
511498 } 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976

Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

S12075 } RC
↓
S12079 } 02/01/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

S12075 } RC
↓
S12079 } 02/01/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL	
Compound			CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane			123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988	1
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998	2
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200	3
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573	4
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180	5
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574	6
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837	7
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290	8
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829	9
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937	10
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505	11
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838	12
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909	13
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442	14
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512	15
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230	16
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876	17
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887	18
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265	
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422	
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968	
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998	
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221	
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240	
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553	
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251	
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220	
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130	
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857	
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220	
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958	
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553	

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988	1
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998	2
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200	3
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573	4
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180	5
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574	6
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837	7
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290	8
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829	9
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937	10
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505	11
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838	12
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909	13
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442	14
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512	15
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230	16
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876	17
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887	18
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265	
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422	
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968	
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998	
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221	
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240	
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553	
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251	
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220	
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130	
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857	
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220	
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958	
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553	

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : January 31, 2032 **Storage:** 10°C or colder
Ship: Ambient

512187 } RC/
↓
512206 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope A. Riglin

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024

Balance Serial # 1128360905

Christie Mills

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : December 31, 2029 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

S12207 } RC/
↓
S12221 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-S (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : December 31, 2029 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

S12207 } RC/
↓
S12221 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-S (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12312 } RC/
↓ 05/30/24
S12331

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12312 } RC/
↓ 05/30/24
S12331

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024

Balance Serial # 1128360905

Jennifer J. Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. **Q1121**
QUOTE NO.
COC Number **2041597**

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: **Tetra Tech Inc.**
ADDRESS: **4433 Corporation Lane Suite 300**
CITY: **Virginia Beach** STATE: **VA** ZIP: **23462**
ATTENTION: **Ernie Wu**
PHONE: **757-466-4901** FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **NWIRP Beth page**
PROJECT NO.: **1126-08005-WE13** LOCATION: **Bethpage, NY**
PROJECT MANAGER: **Ernie Wu**
e-mail: **ernie.wu@tetratech.com**
PHONE: **757-466-4901** FAX:

CLIENT BILLING INFORMATION

BILL TO: **See Contract** PO#:
ADDRESS:
CITY STATE ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): **Standard TAT** DAYS*
EDD: **Standard TAT** DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☒ Other **See Contract**
☐ EDD FORMAT

VOCs 1,4-Dioxane 8270511
1 2 3 4 5 6 7 8 9

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS		
			COMP	GRAB	DATE	TIME		A/E	E									← Specify Preservatives A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER
								1	2	3	4	5	6	7	8	9			
1.	RW10A-20250116	GW		X	1-16-25	1040	3	2	1										
2.																			
3.																			
4.																			
5.																			
6.																			
7.																			
8.																			
9.																			
10.																			

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. gh Wu	DATE/TIME: 1-16-25/1530	RECEIVED BY: [Signature] 1530	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 2.8°C
RELINQUISHED BY SAMPLER: 2. [Signature]	DATE/TIME:	RECEIVED BY: [Signature]	Comments:
RELINQUISHED BY SAMPLER: 3. [Signature]	DATE/TIME: 1-16-25	RECEIVED BY: 3.	

Page **1** of **1**

CLIENT: ☐ Hand Delivered ☐ Other
CHEMTECH: ☒ Picked Up ☐ Field Sampling

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1121	TETR06	Order Date : 1/17/2025 7:55:00 AM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : NWIRP Bethpage 112G080	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 1/16/2025 6:10:00 PM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1121-01	RW10A-20250116	Water	01/16/2025	10:40	VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time : 1/17/25 0940

Received By :

Date / Time : 01/17/25 9:40 AM

Storage Area : VOA Refridgerator Room

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN035998.D
 Acq On : 20 Jan 2025 14:38
 Operator : RC/JU
 Sample : PB166108BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB166108BS

Quant Time: Jan 20 15:10:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2154	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4363	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2173	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	4379	0.400	ng	#-0.02
29) Chrysene-d12	21.367	240	4015	0.400	ng	-0.02
35) Perylene-d12	23.663	264	4099	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.398	112	2207	0.418	ng	-0.01
5) Phenol-d6	6.972	99	2534	0.386	ng	-0.01
8) Nitrobenzene-d5	8.956	82	1840	0.533	ng	-0.03
11) 2-Methylnaphthalene-d10	12.198	152	2490	0.426	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	469	0.450	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4059	0.426	ng	-0.02
27) Fluoranthene-d10	19.216	212	4814	0.443	ng	-0.02
31) Terphenyl-d14	19.815	244	3584	0.448	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.311	88	810	0.379	ng	# 74
3) n-Nitrosodimethylamine	3.614	42	1712	0.459	ng	94
6) bis(2-Chloroethyl)ether	7.239	93	2482	0.496	ng	97
9) Naphthalene	10.654	128	5221	0.426	ng	98
10) Hexachlorobutadiene	10.953	225	1557	0.392	ng	# 100
12) 2-Methylnaphthalene	12.274	142	3281	0.433	ng	99
16) Acenaphthylene	14.164	152	4595	0.449	ng	100
17) Acenaphthene	14.506	154	2953	0.440	ng	99
18) Fluorene	15.500	166	3538	0.479	ng	98
20) 4,6-Dinitro-2-methylph...	15.560	198	290	0.381	ng	97
21) 4-Bromophenyl-phenylether	16.380	248	1238	0.412	ng	# 84
22) Hexachlorobenzene	16.504	284	1631	0.398	ng	97
23) Atrazine	16.653	200	1013	0.503	ng	# 85
24) Pentachlorophenol	16.839	266	1173	0.809	ng	97
25) Phenanthrene	17.224	178	5634	0.441	ng	99
26) Anthracene	17.323	178	5206	0.448	ng	98
28) Fluoranthene	19.248	202	6317	0.424	ng	98
30) Pyrene	19.611	202	6564	0.403	ng	99
32) Benzo(a)anthracene	21.349	228	6038	0.427	ng	98
33) Chrysene	21.403	228	6127	0.414	ng	100
34) Bis(2-ethylhexyl)phtha...	21.295	149	3249	0.564	ng	99
36) Indeno(1,2,3-cd)pyrene	26.002	276	6731	0.416	ng	98
37) Benzo(b)fluoranthene	23.014	252	6168	0.438	ng	97
38) Benzo(k)fluoranthene	23.014	252	6168	0.442	ng	98
39) Benzo(a)pyrene	23.561	252	5550	0.455	ng	98
40) Dibenzo(a,h)anthracene	26.017	278	5340	0.415	ng	100
41) Benzo(g,h,i)perylene	26.712	276	5293	0.368	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN035998.D
Acq On : 20 Jan 2025 14:38
Operator : RC/JU
Sample : PB166108BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166108BS


Quant Time: Jan 20 15:10:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.817 min Scan# 64
Delta R.T. -0.019 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38


Instrument :
BNA_N
ClientSampleId :
PB166108BS





Tgt Ion:152 Resp: 2154
Ion Ratio Lower Upper
152 100
150 155.3 117.8 176.6
115 65.5 51.0 76.4




#2
1,4-Dioxane
Concen: 0.379 ng
RT: 3.311 min Scan# 18
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38



Tgt Ion: 88 Resp: 810
Ion Ratio Lower Upper
88 100
43 77.8 32.7 49.1#
58 86.3 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 0.459 ng
RT: 3.614 min Scan# 60
Delta R.T. -0.018 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:	42	Resp:	1712
Ion	Ratio	Lower	Upper
42	100		
74	83.6	62.2	93.2
44	10.3	7.1	10.7



#4
2-Fluorophenol
Concen: 0.418 ng
RT: 5.398 min Scan# 307
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:	112	Resp:	2207
Ion	Ratio	Lower	Upper
112	100		
64	63.1	48.2	72.2
63	38.9	30.0	45.0









#5
Phenol-d6
Concen: 0.386 ng
RT: 6.972 min Scan# 51
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:	99	Resp:	2534
Ion	Ratio	Lower	Upper
99	100		
42	30.3	23.5	35.3
71	45.9	35.5	53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.496 ng
RT: 7.239 min Scan# 562
Delta R.T. -0.018 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:	93	Resp:	2482
Ion	Ratio	Lower	Upper
93	100		
63	80.3	62.0	93.0
95	32.4	25.5	38.3





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.611 min Scan# 91

Delta R.T. -0.015 min

Lab File: BN035998.D


Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS



Tgt Ion:136 Resp: 4363




Ion Ratio Lower Upper

136 100

137 12.3 10.6 15.8

54 9.5 9.8 14.6#

68 6.6 6.6 10.0#



#8

Nitrobenzene-d5

Concen: 0.533 ng

RT: 8.956 min Scan# 771

Delta R.T. -0.025 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38




Tgt Ion: 82 Resp: 1840


Ion Ratio Lower Upper

82 100

128 34.7 36.9 55.3#

54 67.9 50.4 75.6





#9

Naphthalene

Concen: 0.426 ng

RT: 10.654 min Scan# 90

Delta R.T. -0.025 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS





Tgt Ion:128 Resp: 5221

Ion Ratio Lower Upper

128 100

129 12.5 10.6 16.0

127 15.2 12.8 19.2



#10

Hexachlorobutadiene

Concen: 0.392 ng

RT: 10.953 min Scan# 958

Delta R.T. -0.025 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38




Tgt Ion:225 Resp: 1557


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.0 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.426 ng
RT: 12.198 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:152 Resp: 2490
Ion Ratio Lower Upper
152 100
151 21.6 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.433 ng
RT: 12.274 min Scan# 1161
Delta R.T. -0.019 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:142 Resp: 3281
Ion Ratio Lower Upper
142 100
141 90.2 71.9 107.9
115 48.7 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 14
Delta R.T. -0.009 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:164 Resp: 2173
Ion Ratio Lower Upper
164 100
162 102.0 83.1 124.7
160 51.1 46.0 69.0



#14
2,4,6-Tribromophenol
Concen: 0.450 ng
RT: 15.933 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:330 Resp: 469
Ion Ratio Lower Upper
330 100
332 98.3 77.2 115.8
141 47.3 31.6 47.4









#15
2-Fluorobiphenyl
Concen: 0.426 ng
RT: 13.073 min Scan# 1410
Delta R.T. -0.020 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:	172	Resp:	4059
Ion Ratio	Lower	Upper	
172	100		
171	36.2	30.5	45.7
170	25.1	20.8	31.2



#16
Acenaphthylene
Concen: 0.449 ng
RT: 14.164 min Scan# 1414
Delta R.T. -0.020 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:	152	Resp:	4595
Ion Ratio	Lower	Upper	
152	100		
151	20.6	16.3	24.5
153	13.3	10.6	15.8









#17
Acenaphthene
Concen: 0.440 ng
RT: 14.506 min Scan# 1446
Delta R.T. -0.020 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:154 Resp: 2953
Ion Ratio Lower Upper
154 100
153 112.3 90.5 135.7
152 59.1 48.6 73.0



#18
Fluorene
Concen: 0.479 ng
RT: 15.500 min Scan# 1539
Delta R.T. -0.009 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:166 Resp: 3538
Ion Ratio Lower Upper
166 100
165 102.9 80.6 120.8
167 13.7 11.4 17.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.186 min Scan# 11
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:188 Resp: 4379
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.3 7.3 10.9#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.381 ng
RT: 15.560 min Scan# 1544
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:198 Resp: 290
Ion Ratio Lower Upper
198 100
51 77.9 64.3 96.5
105 59.3 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.412 ng
RT: 16.380 min Scan# 1610
Delta R.T. -0.024 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:248 Resp: 1238
Ion Ratio Lower Upper
248 100
250 91.1 76.8 115.2
141 105.2 63.6 95.4#



#22
Hexachlorobenzene
Concen: 0.398 ng
RT: 16.504 min Scan# 1620
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:284 Resp: 1631
Ion Ratio Lower Upper
284 100
142 41.8 31.4 47.0
249 36.2 27.8 41.8









#23
Atrazine
Concen: 0.503 ng
RT: 16.653 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:200 Resp: 1013
Ion Ratio Lower Upper
200 100
173 27.6 35.4 53.0#
215 48.7 42.4 63.6



#24
Pentachlorophenol
Concen: 0.809 ng
RT: 16.839 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:266 Resp: 1173
Ion Ratio Lower Upper
266 100
264 64.5 49.9 74.9
268 65.8 50.2 75.2









#25
 Phenanthrene
 Concen: 0.441 ng
 RT: 17.224 min Scan# 1686
 Delta R.T. -0.024 min
 Lab File: BN035998.D
 Acq: 20 Jan 2025 14:38




Instrument :
 BNA_N
 ClientSampleId :
 PB166108BS


Tgt Ion:	178	Resp:	5634
Ion Ratio	Lower	Upper	
178	100		
176	19.9	15.9	23.9
179	15.3	12.9	19.3



#26
 Anthracene
 Concen: 0.448 ng
 RT: 17.323 min Scan# 1686
 Delta R.T. -0.011 min
 Lab File: BN035998.D
 Acq: 20 Jan 2025 14:38

Tgt Ion:	178	Resp:	5206
Ion Ratio	Lower	Upper	
178	100		
176	19.4	15.0	22.6
179	15.3	13.0	19.6









#27
Fluoranthene-d10
Concen: 0.443 ng
RT: 19.216 min Scan# 1940
Delta R.T. -0.022 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38




Instrument :
BNA_N
ClientSampleId :
PB166108BS


Tgt Ion:212 Resp: 4814
Ion Ratio Lower Upper
212 100
106 12.3 9.0 13.6
104 7.5 5.4 8.2



#28
Fluoranthene
Concen: 0.424 ng
RT: 19.248 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:202 Resp: 6317
Ion Ratio Lower Upper
202 100
101 10.2 7.2 10.8
203 17.0 13.9 20.9





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.367 min Scan# 21

Delta R.T. -0.017 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS





Tgt Ion:240 Resp: 4015

Ion Ratio Lower Upper

240 100

120 12.7 11.1 16.7

236 28.7 24.6 36.8



#30

Pyrene

Concen: 0.403 ng

RT: 19.611 min Scan# 2018

Delta R.T. -0.017 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38




Tgt Ion:202 Resp: 6564


Ion Ratio Lower Upper

202 100

200 21.8 17.0 25.6


203 17.8 14.6 22.0








#31
 Terphenyl-d14
 Concen: 0.448 ng
 RT: 19.815 min Scan# 20
 Delta R.T. -0.022 min
 Lab File: BN035998.D
 Acq: 20 Jan 2025 14:38

Instrument :
 BNA_N
 ClientSampleId :
 PB166108BS







Tgt Ion:244 Resp: 3584
 Ion Ratio Lower Upper
 244 100
 212 11.0 10.1 15.1
 122 14.4 12.2 18.4



#32
 Benzo(a)anthracene
 Concen: 0.427 ng
 RT: 21.349 min Scan# 2253
 Delta R.T. -0.026 min
 Lab File: BN035998.D
 Acq: 20 Jan 2025 14:38


Tgt Ion:228 Resp: 6038
 Ion Ratio Lower Upper
 228 100
 226 28.0 22.7 34.1
 229 19.9 17.1 25.7








#33
Chrysene
Concen: 0.414 ng
RT: 21.403 min Scan# 21
Delta R.T. -0.017 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38


Instrument :
BNA_N
ClientSampleId :
PB166108BS





Tgt Ion:228 Resp: 6127
Ion Ratio Lower Upper
228 100
226 31.4 25.2 37.8
229 19.7 16.0 24.0




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.564 ng
RT: 21.295 min Scan# 2247
Delta R.T. -0.026 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38



Tgt Ion:149 Resp: 3249
Ion Ratio Lower Upper
149 100
167 26.5 21.4 32.0
279 3.3 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.663 min Scan# 21

Delta R.T. -0.031 min

Lab File: BN035998.D


Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS






Tgt Ion:264 Resp: 4099

Ion Ratio Lower Upper

264 100

260 27.1 21.7 32.5

265 66.0 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.416 ng

RT: 26.002 min Scan# 3601

Delta R.T. -0.046 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38




Tgt Ion:276 Resp: 6731


Ion Ratio Lower Upper

276 100

138 25.0 18.9 28.3


277 25.4 19.5 29.3








#37
Benzo(b)fluoranthene
Concen: 0.438 ng
RT: 23.014 min Scan# 2579
Delta R.T. 0.016 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Instrument :
BNA_N
ClientSampleId :
PB166108BS







Tgt Ion:252 Resp: 6168
Ion Ratio Lower Upper
252 100
253 26.3 20.1 30.1
125 14.6 9.9 14.9



#38
Benzo(k)fluoranthene
Concen: 0.442 ng
RT: 23.014 min Scan# 2579
Delta R.T. -0.031 min
Lab File: BN035998.D
Acq: 20 Jan 2025 14:38

Tgt Ion:252 Resp: 6168
Ion Ratio Lower Upper
252 100
253 26.3 20.5 30.7
125 14.6 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.455 ng

RT: 23.561 min Scan# 21

Delta R.T. -0.028 min

Lab File: BN035998.D


Acq: 20 Jan 2025 14:38

Instrument :

BNA_N

ClientSampleId :

PB166108BS






Tgt Ion:252 Resp: 5550

Ion Ratio Lower Upper

252 100

253 27.3 21.5 32.3

125 17.1 12.6 19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.415 ng

RT: 26.017 min Scan# 3606

Delta R.T. -0.054 min

Lab File: BN035998.D

Acq: 20 Jan 2025 14:38




Tgt Ion:278 Resp: 5340


Ion Ratio Lower Upper

278 100

139 19.8 15.9 23.9


279 28.5 23.0 34.4







#41
 Benzo(g,h,i)perylene
 Concen: 0.368 ng
 RT: 26.712 min Scan# 3844
 Delta R.T. -0.051 min
 Lab File: BN035998.D
 Acq: 20 Jan 2025 14:38

Instrument :
 BNA_N
 ClientSampleId :
 PB166108BS



Tgt Ion:	276	Resp:	5293
Ion	Ratio	Lower	Upper
276	100		
277	27.1	22.8	34.2
138	24.3	20.1	30.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
 Data File : BN036000.D
 Acq On : 20 Jan 2025 16:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Jan 20 17:29:00 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2210	0.400	ng	-0.02
7) Naphthalene-d8	10.611	136	4309	0.400	ng	#-0.01
13) Acenaphthene-d10	14.452	164	2121	0.400	ng	0.00
19) Phenanthrene-d10	17.186	188	3983	0.400	ng	-0.02
29) Chrysene-d12	21.367	240	3167	0.400	ng	-0.02
35) Perylene-d12	23.666	264	3290	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.390	112	2542	0.469	ng	-0.02
5) Phenol-d6	6.972	99	2979	0.442	ng	-0.01
8) Nitrobenzene-d5	8.956	82	1855	0.544	ng	-0.03
11) 2-Methylnaphthalene-d10	12.198	152	2501	0.434	ng	-0.02
14) 2,4,6-Tribromophenol	15.933	330	497	0.488	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	4137	0.444	ng	-0.02
27) Fluoranthene-d10	19.221	212	4081	0.413	ng	-0.02
31) Terphenyl-d14	19.815	244	2866	0.454	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.318	88	1068	0.487	ng	99
3) n-Nitrosodimethylamine	3.621	42	1932	0.504	ng	# 98
6) bis(2-Chloroethyl)ether	7.232	93	2472	0.482	ng	97
9) Naphthalene	10.654	128	5411	0.447	ng	97
10) Hexachlorobutadiene	10.953	225	1718	0.437	ng	# 99
12) 2-Methylnaphthalene	12.269	142	3275	0.438	ng	99
16) Acenaphthylene	14.164	152	4225	0.423	ng	99
17) Acenaphthene	14.506	154	2843	0.434	ng	99
18) Fluorene	15.500	166	3587	0.498	ng	98
20) 4,6-Dinitro-2-methylph...	15.560	198	256	0.370	ng	94
21) 4-Bromophenyl-phenylether	16.379	248	1236	0.453	ng	# 88
22) Hexachlorobenzene	16.504	284	1595	0.428	ng	97
23) Atrazine	16.653	200	839	0.458	ng	# 87
24) Pentachlorophenol	16.839	266	597	0.453	ng	97
25) Phenanthrene	17.223	178	5041	0.434	ng	99
26) Anthracene	17.323	178	4488	0.424	ng	99
28) Fluoranthene	19.248	202	5508	0.407	ng	99
30) Pyrene	19.611	202	5550	0.432	ng	99
32) Benzo(a)anthracene	21.349	228	4679	0.419	ng	98
33) Chrysene	21.403	228	4915	0.421	ng	99
34) Bis(2-ethylhexyl)phtha...	21.295	149	2577	0.567	ng	98
36) Indeno(1,2,3-cd)pyrene	26.002	276	5439	0.419	ng	99
37) Benzo(b)fluoranthene	23.020	252	4806	0.425	ng	96
38) Benzo(k)fluoranthene	23.020	252	4806	0.430	ng	98
39) Benzo(a)pyrene	23.567	252	4125	0.422	ng	96
40) Dibenzo(a,h)anthracene	26.019	278	4327	0.419	ng	100
41) Benzo(g,h,i)perylene	26.718	276	4707	0.408	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012025\
Data File : BN036000.D
Acq On : 20 Jan 2025 16:11
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Jan 20 17:29:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 02 15:39:17 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.817 min Scan# 64
 Delta R.T. -0.019 min
 Lab File: BN036000.D
 Acq: 20 Jan 2025 16:11


Instrument :
 BNA_N
 ClientSampleId :
 SSTDC00.4EC





Tgt Ion:152 Resp: 2210
 Ion Ratio Lower Upper
 152 100
 150 150.6 117.8 176.6
 115 63.5 51.0 76.4




#2
 1,4-Dioxane
 Concen: 0.487 ng
 RT: 3.318 min Scan# 19
 Delta R.T. -0.004 min
 Lab File: BN036000.D
 Acq: 20 Jan 2025 16:11



Tgt Ion: 88 Resp: 1068
 Ion Ratio Lower Upper
 88 100
 43 43.3 32.7 49.1
 58 78.6 63.0 94.4









#3
n-Nitrosodimethylamine
Concen: 0.504 ng
RT: 3.621 min Scan# 61
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 1932
Ion Ratio Lower Upper
42 100
74 76.5 62.2 93.2
44 5.8 7.1 10.7#



#4
2-Fluorophenol
Concen: 0.469 ng
RT: 5.390 min Scan# 306
Delta R.T. -0.018 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion: 112 Resp: 2542
Ion Ratio Lower Upper
112 100
64 62.7 48.2 72.2
63 39.1 30.0 45.0









#5
Phenol-d6
Concen: 0.442 ng
RT: 6.972 min Scan# 51
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	2979
Ion	Ratio	Lower	Upper
99	100		
42	31.4	23.5	35.3
71	45.4	35.5	53.3



#6
bis(2-Chloroethyl)ether
Concen: 0.482 ng
RT: 7.232 min Scan# 561
Delta R.T. -0.026 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:	93	Resp:	2472
Ion	Ratio	Lower	Upper
93	100		
63	80.8	62.0	93.0
95	32.5	25.5	38.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.611 min Scan# 91
Delta R.T. -0.015 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	4309
Ion Ratio	Lower	Upper	
136	100		
137	12.6	10.6	15.8
54	9.5	9.8	14.6#
68	6.3	6.6	10.0#



#8
Nitrobenzene-d5
Concen: 0.544 ng
RT: 8.956 min Scan# 771
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:	82	Resp:	1855
Ion Ratio	Lower	Upper	
82	100		
128	36.0	36.9	55.3#
54	69.0	50.4	75.6









#9
Naphthalene
Concen: 0.447 ng
RT: 10.654 min Scan# 911
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 5411
Ion Ratio Lower Upper
128 100
129 12.5 10.6 16.0
127 14.5 12.8 19.2



#10
Hexachlorobutadiene
Concen: 0.437 ng
RT: 10.953 min Scan# 958
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:225 Resp: 1718
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.5 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.434 ng
RT: 12.198 min Scan# 1198
Delta R.T. -0.019 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 2501
Ion Ratio Lower Upper
152 100
151 21.9 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.438 ng
RT: 12.269 min Scan# 1160
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:142 Resp: 3275
Ion Ratio Lower Upper
142 100
141 90.2 71.9 107.9
115 51.2 39.6 59.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 14
Delta R.T. -0.010 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	164	Resp:	2121
Ion	Ratio	Lower	Upper
164	100		
162	101.5	83.1	124.7
160	51.7	46.0	69.0



#14
2,4,6-Tribromophenol
Concen: 0.488 ng
RT: 15.933 min Scan# 1574
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:	330	Resp:	497
Ion	Ratio	Lower	Upper
330	100		
332	101.8	77.2	115.8
141	45.3	31.6	47.4









#15
2-Fluorobiphenyl
Concen: 0.444 ng
RT: 13.073 min Scan# 1410
Delta R.T. -0.020 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	4137
Ion Ratio	Lower	Upper	
172	100		
171	36.7	30.5	45.7
170	25.1	20.8	31.2



#16
Acenaphthylene
Concen: 0.423 ng
RT: 14.164 min Scan# 1414
Delta R.T. -0.020 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:	152	Resp:	4225
Ion Ratio	Lower	Upper	
152	100		
151	20.8	16.3	24.5
153	12.8	10.6	15.8









#17
Acenaphthene
Concen: 0.434 ng
RT: 14.506 min Scan# 1446
Delta R.T. -0.020 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 2843
Ion Ratio Lower Upper
154 100
153 112.8 90.5 135.7
152 59.8 48.6 73.0



#18
Fluorene
Concen: 0.498 ng
RT: 15.500 min Scan# 1539
Delta R.T. -0.010 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:166 Resp: 3587
Ion Ratio Lower Upper
166 100
165 98.9 80.6 120.8
167 13.1 11.4 17.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.186 min Scan# 1

Delta R.T. -0.024 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:188 Resp: 3983

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.6 7.3 10.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.370 ng

RT: 15.560 min Scan# 1544

Delta R.T. -0.011 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11




Tgt Ion:198 Resp: 256


Ion Ratio Lower Upper

198 100

51 75.8 64.3 96.5

105 57.7 50.0 75.0









#21
4-Bromophenyl-phenylether
Concen: 0.453 ng
RT: 16.379 min Scan# 1610
Delta R.T. -0.024 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1236
Ion Ratio Lower Upper
248 100
250 93.2 76.8 115.2
141 100.0 63.6 95.4#



#22
Hexachlorobenzene
Concen: 0.428 ng
RT: 16.504 min Scan# 1620
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:284 Resp: 1595
Ion Ratio Lower Upper
284 100
142 42.0 31.4 47.0
249 35.5 27.8 41.8









#23
Atrazine
Concen: 0.458 ng
RT: 16.653 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 839
Ion Ratio Lower Upper
200 100
173 29.2 35.4 53.0#
215 49.4 42.4 63.6



#24
Pentachlorophenol
Concen: 0.453 ng
RT: 16.839 min Scan# 1647
Delta R.T. -0.011 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:266 Resp: 597
Ion Ratio Lower Upper
266 100
264 61.3 49.9 74.9
268 66.2 50.2 75.2









#25
 Phenanthrene
 Concen: 0.434 ng
 RT: 17.223 min Scan# 1686
 Delta R.T. -0.024 min
 Lab File: BN036000.D
 Acq: 20 Jan 2025 16:11




Instrument : BNA_N
 ClientSampleId : SSTDCCC0.4EC


Tgt Ion:178	Resp:	5041
Ion Ratio	Lower	Upper
178	100	
176	19.7	15.9 23.9
179	15.7	12.9 19.3



#26
 Anthracene
 Concen: 0.424 ng
 RT: 17.323 min Scan# 1686
 Delta R.T. -0.011 min
 Lab File: BN036000.D
 Acq: 20 Jan 2025 16:11

Tgt Ion:178	Resp:	4488
Ion Ratio	Lower	Upper
178	100	
176	19.3	15.0 22.6
179	15.7	13.0 19.6









#27
Fluoranthene-d10
Concen: 0.413 ng
RT: 19.221 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 4081
Ion Ratio Lower Upper
212 100
106 11.9 9.0 13.6
104 7.2 5.4 8.2



#28
Fluoranthene
Concen: 0.407 ng
RT: 19.248 min Scan# 1940
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:202 Resp: 5508
Ion Ratio Lower Upper
202 100
101 9.4 7.2 10.8
203 17.2 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.367 min Scan# 21
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 3167
Ion Ratio Lower Upper
240 100
120 15.5 11.1 16.7
236 30.4 24.6 36.8



#30
Pyrene
Concen: 0.432 ng
RT: 19.611 min Scan# 2018
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:202 Resp: 5550
Ion Ratio Lower Upper
202 100
200 21.5 17.0 25.6
203 17.9 14.6 22.0





#31

Terphenyl-d14

Concen: 0.454 ng

RT: 19.815 min Scan# 20

Delta R.T. -0.022 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:244 Resp: 2866

Ion Ratio Lower Upper

244 100

212 11.8 10.1 15.1

122 15.6 12.2 18.4



#32

Benzo(a)anthracene

Concen: 0.419 ng

RT: 21.349 min Scan# 2253

Delta R.T. -0.026 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11




Tgt Ion:228 Resp: 4679


Ion Ratio Lower Upper

228 100

226 29.3 22.7 34.1

229 20.3 17.1 25.7









#33
Chrysene
Concen: 0.421 ng
RT: 21.403 min Scan# 21
Delta R.T. -0.017 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:228 Resp: 4915
Ion Ratio Lower Upper
228 100
226 32.3 25.2 37.8
229 20.2 16.0 24.0



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.567 ng
RT: 21.295 min Scan# 2247
Delta R.T. -0.026 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:149 Resp: 2577
Ion Ratio Lower Upper
149 100
167 27.8 21.4 32.0
279 3.5 3.0 4.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.666 min Scan# 21

Delta R.T. -0.028 min

Lab File: BN036000.D


Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:264 Resp: 3290

Ion Ratio Lower Upper

264 100

260 25.9 21.7 32.5

265 67.1 33.9 50.9#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.419 ng

RT: 26.002 min Scan# 3601

Delta R.T. -0.046 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11




Tgt Ion:276 Resp: 5439


Ion Ratio Lower Upper

276 100

138 24.6 18.9 28.3

277 24.7 19.5 29.3









#37
Benzo(b)fluoranthene
Concen: 0.425 ng
RT: 23.020 min Scan# 2581
Delta R.T. 0.022 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 4806
Ion Ratio Lower Upper
252 100
253 26.7 20.1 30.1
125 14.3 9.9 14.9



#38
Benzo(k)fluoranthene
Concen: 0.430 ng
RT: 23.020 min Scan# 2581
Delta R.T. -0.025 min
Lab File: BN036000.D
Acq: 20 Jan 2025 16:11

Tgt Ion:252 Resp: 4806
Ion Ratio Lower Upper
252 100
253 26.7 20.5 30.7
125 14.3 10.6 16.0





#39

Benzo(a)pyrene

Concen: 0.422 ng

RT: 23.567 min Scan# 21

Delta R.T. -0.022 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 4125

Ion Ratio Lower Upper

252 100

253 29.4 21.5 32.3

125 17.0 12.6 19.0



#40

Dibenzo(a,h)anthracene

Concen: 0.419 ng

RT: 26.019 min Scan# 3607

Delta R.T. -0.052 min

Lab File: BN036000.D

Acq: 20 Jan 2025 16:11




Tgt Ion:278 Resp: 4327


Ion Ratio Lower Upper

278 100

139 20.3 15.9 23.9


279 28.7 23.0 34.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.408 ng
 RT: 26.718 min Scan# 3846
 Delta R.T. -0.046 min
 Lab File: BN036000.D
 Acq: 20 Jan 2025 16:11

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion: 276 Resp: 4707

Ion	Ratio	Lower	Upper
276	100		
277	27.1	22.8	34.2
138	24.7	20.1	30.1

