

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1121

SDG NO.: Q1121

Lab File ID: BN035990.D

DFTPP Injection Date: 01/20/2025

Instrument ID: BNA\_N

DFTPP Injection Time: 09:13

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.3
68	Less than 2.0% of mass 69	0.4 ( 0.8 ) 1
69	Mass 69 relative abundance	47.1
70	Less than 2.0% of mass 69	0.3 ( 0.7 ) 1
127	10.0 - 80.0% of mass 198	49
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.4
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	9.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.1 (18.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035991.D	01/20/2025	10:27
PB166108BL	PB166108BL	BN035992.D	01/20/2025	11:03
RW10A-20250116	Q1121-01	BN035997.D	01/20/2025	14:03
PB166108BS	PB166108BS	BN035998.D	01/20/2025	14:38
PB166108BSD	PB166108BSD	BN035999.D	01/20/2025	15:14
SSTDCCC0.4EC	SSTDCCC0.4	BN036000.D	01/20/2025	16:11