

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG No.: Q1121

Instrument ID: MSVOA_N Calibration Date/Time: 01/17/2025 09:06

Lab File ID: VN085480.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.733	0.683	0.1	-6.82	20
Vinyl Chloride	0.737	0.682		-7.46	20
Bromomethane	0.445	0.415		-6.74	20
Chloroethane	0.467	0.463		-0.86	20
Trichlorofluoromethane	1.069	1.061		-0.75	20
1,1,2-Trichlorotrifluoroethane	0.602	0.603		0.17	20
1,1-Dichloroethene	0.537	0.542		0.93	20
Acetone	0.261	0.244		-6.51	20
Carbon Disulfide	1.652	1.499		-9.26	20
Methyl tert-butyl Ether	1.742	1.821		4.53	20
Methylene Chloride	0.646	0.626		-3.1	20
trans-1,2-Dichloroethene	0.573	0.565		-1.4	20
1,1-Dichloroethane	1.178	1.171	0.1	-0.59	20
2-Butanone	0.384	0.350		-8.85	20
Carbon Tetrachloride	0.557	0.584		4.85	20
cis-1,2-Dichloroethene	0.675	0.685		1.48	20
Chloroform	1.218	1.213		-0.41	20
1,1,1-Trichloroethane	1.068	1.081		1.22	20
Methylcyclohexane	0.457	0.538		17.72	20
Benzene	1.463	1.537		5.06	20
1,2-Dichloroethane	0.551	0.574		4.17	20
Trichloroethene	0.341	0.349		2.35	20
1,2-Dichloropropane	0.374	0.387		3.48	20
Bromodichloromethane	0.549	0.594		8.2	20
4-Methyl-2-Pentanone	0.457	0.464		1.53	20
Toluene	0.848	0.951		12.15	20
t-1,3-Dichloropropene	0.519	0.585		12.72	20
cis-1,3-Dichloropropene	0.554	0.621		12.09	20
1,1,2-Trichloroethane	0.335	0.359		7.16	20
2-Hexanone	0.321	0.330		2.8	20
Dibromochloromethane	0.405	0.427		5.43	20
Tetrachloroethene	0.341	0.367		7.63	20
Chlorobenzene	1.095	1.136	0.3	3.74	20
Ethyl Benzene	1.784	2.030		13.79	20
m/p-Xylenes	0.659	0.770		16.84	20
o-Xylene	0.630	0.722		14.6	20
Styrene	1.043	1.239		18.79	20
Bromoform	0.288	0.307	0.1	6.6	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG No.: Q1121

Instrument ID: MSVOA_N Calibration Date/Time: 01/17/2025 09:06

Lab File ID: VN085480.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.375	3.839		13.75	20
1,1,2,2-Tetrachloroethane	1.192	1.098	0.3	-7.89	20
1,3-Dichlorobenzene	1.624	1.681		3.51	20
1,4-Dichlorobenzene	1.683	1.690		0.42	20
1,2-Dichlorobenzene	1.620	1.618		-0.12	20
1,2-Dichloroethane-d4	0.807	0.803		-0.5	20
Dibromofluoromethane	0.347	0.370		6.63	20
Toluene-d8	1.232	1.364		10.71	20
4-Bromofluorobenzene	0.422	0.475		12.56	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.