

Surrogate Summary

SDG No.: Q1121

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1121-01	RW10A-20250116	1,2-Dichloroethane-d4	50	52.3	105	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	49.6	99	89	112
		4-Bromofluorobenzene	50	46.4	93	85	114
VN0117WBL01	VN0117WBL01	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	51.7	103	80	119
		Toluene-d8	50	50.1	100	89	112
		4-Bromofluorobenzene	50	47.3	95	85	114
VN0117WBS01	VN0117WBS01	1,2-Dichloroethane-d4	50	50.6	101	81	118
		Dibromofluoromethane	50	52.1	104	80	119
		Toluene-d8	50	52.9	106	89	112
		4-Bromofluorobenzene	50	52.3	105	85	114
VN0117WBSD0	VN0117WBSD01	1,2-Dichloroethane-d4	50	52.6	105	81	118
		Dibromofluoromethane	50	52.1	104	80	119
		Toluene-d8	50	53.3	107	89	112
		4-Bromofluorobenzene	50	54.2	108	85	114