

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06  
 Lab Code: CHEM Case No.: Q1121 SAS No.: Q1121 SDG NO.: Q1121  
 Lab File ID: VN085480.D Date Analyzed: 01/17/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 09:06  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 155506        | 13.788 |  |  |  |  |
| UPPER LIMIT    | 311012        | 14.288 |  |  |  |  |
| LOWER LIMIT    | 77753         | 13.288 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| RW10A-20250116 | 115974        | 13.79  |  |  |  |  |
| VN0117WBL01    | 125050        | 13.79  |  |  |  |  |
| VN0117WBS01    | 143650        | 13.79  |  |  |  |  |
| VN0117WBSD01   | 132750        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.