

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Project: NWIRP Bethpage 112G08005-WE13

RunNo.: LB134337

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	01/20/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	01/20/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	01/20/2025

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Project: NWIRP Bethpage 112G08005-WE13

RunNo.: LB134369

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.482	0.50	96	90-110	01/22/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.490	0.50	98	90-110	01/22/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.490	0.50	98	90-110	01/22/2025