

Matrix Spike Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q1122
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q1122-01
Client ID:	RW10A-20250116MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.48		0.016	J	0.5	1	93		01/22/2025

Matrix Spike Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q1122
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q1122-01
Client ID:	RW10A-20250116MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Phosphorus, Total	mg/L	90-110	0.48		0.016	J	0.5	1	93		01/22/2025