

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB134326

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134326BS							
TDS	mg/L	100	94.0		94	1	90-110	01/17/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB134342

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134342BS							
TSS	mg/L	550	520		94	1	90-110	01/17/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB134367

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134367BS							
Alkalinity	mg/L	50	50.4		101	1	80-120	01/22/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB134369

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166181BS							
Phosphorus, Total	mg/L	0.50	0.49		98	1	90-110	01/22/2025