



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. SDG No.: Q1122
Contract: TETR06 Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	91.5	100	92	40 - 160	P	01/21/2025	11:59	LB134358
	Antimony	48.9	50.0	98	40 - 160	P	01/21/2025	11:59	LB134358
	Arsenic	20.8	20.0	104	40 - 160	P	01/21/2025	11:59	LB134358
	Barium	87.2	100	87	40 - 160	P	01/21/2025	11:59	LB134358
	Beryllium	5.73	6.0	96	40 - 160	P	01/21/2025	11:59	LB134358
	Cadmium	5.73	6.0	96	40 - 160	P	01/21/2025	11:59	LB134358
	Calcium	1890	2000	94	40 - 160	P	01/21/2025	11:59	LB134358
	Chromium	9.51	10.0	95	40 - 160	P	01/21/2025	11:59	LB134358
	Cobalt	29.0	30.0	97	40 - 160	P	01/21/2025	11:59	LB134358
	Copper	21.6	20.0	108	40 - 160	P	01/21/2025	11:59	LB134358
	Iron	101	100	101	40 - 160	P	01/21/2025	11:59	LB134358
	Lead	11.6	12.0	97	40 - 160	P	01/21/2025	11:59	LB134358
	Magnesium	1990	2000	100	40 - 160	P	01/21/2025	11:59	LB134358
	Manganese	19.3	20.0	96	40 - 160	P	01/21/2025	11:59	LB134358
	Nickel	38.7	40.0	97	40 - 160	P	01/21/2025	11:59	LB134358
	Potassium	1910	2000	96	40 - 160	P	01/21/2025	11:59	LB134358
	Selenium	19.8	20.0	99	40 - 160	P	01/21/2025	11:59	LB134358
	Silver	11.1	10.0	111	40 - 160	P	01/21/2025	11:59	LB134358
	Sodium	1960	2000	98	40 - 160	P	01/21/2025	11:59	LB134358
	Thallium	40.3	40.0	101	40 - 160	P	01/21/2025	11:59	LB134358
	Vanadium	38.1	40.0	95	40 - 160	P	01/21/2025	11:59	LB134358
	Zinc	42.1	40.0	105	40 - 160	P	01/21/2025	11:59	LB134358
CRA	Mercury	0.17	0.2	85	40 - 160	CV	01/23/2025	10:28	LB134375