



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Tetra Tech NUS, Inc.				SDG No.:	Q1122				
Contract:	TETR06	Lab Code:	CHEM	Case No.:	Q1122	SAS No.:	Q1122			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB66	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	01/23/2025	10:21	LB134375

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Contract: TETR06

Lab Code: CHEM

Case No.: Q1122

SAS No.: Q1122

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	01/23/2025	10:26	LB134375
CCB63	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	01/23/2025	10:58	LB134375
CCB64	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	01/23/2025	11:12	LB134375

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech NUS, Inc.</u>		SDG No.: <u>Q1122</u>								
Contract: <u>TETR06</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1122</u>	SAS No.: <u>Q1122</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	80.0	100	P	01/21/2025	11:51	LB134358
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	01/21/2025	11:51	LB134358
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	11:51	LB134358
	Barium	100	+/-100	U	25.0	100	P	01/21/2025	11:51	LB134358
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	11:51	LB134358
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	11:51	LB134358
	Calcium	2000	+/-2000	U	500	2000	P	01/21/2025	11:51	LB134358
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	11:51	LB134358
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	01/21/2025	11:51	LB134358
	Copper	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	11:51	LB134358
	Iron	100	+/-100	U	80.0	100	P	01/21/2025	11:51	LB134358
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/21/2025	11:51	LB134358
	Magnesium	2000	+/-2000	U	500	2000	P	01/21/2025	11:51	LB134358
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	01/21/2025	11:51	LB134358
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	11:51	LB134358
	Potassium	2000	+/-2000	U	1600	2000	P	01/21/2025	11:51	LB134358
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	11:51	LB134358
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	11:51	LB134358
	Sodium	2000	+/-2000	U	1000	2000	P	01/21/2025	11:51	LB134358
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	11:51	LB134358
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	11:51	LB134358
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	11:51	LB134358

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Contract: TETR06

Lab Code: CHEM

Case No.: Q1122

SAS No.: Q1122

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	80.0	100	P	01/21/2025	12:16	LB134358
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	01/21/2025	12:16	LB134358
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	12:16	LB134358
	Barium	100	+/-100	U	25.0	100	P	01/21/2025	12:16	LB134358
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	12:16	LB134358
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	12:16	LB134358
	Calcium	2000	+/-2000	U	500	2000	P	01/21/2025	12:16	LB134358
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	12:16	LB134358
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	01/21/2025	12:16	LB134358
	Copper	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	12:16	LB134358
	Iron	100	+/-100	U	80.0	100	P	01/21/2025	12:16	LB134358
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/21/2025	12:16	LB134358
	Magnesium	2000	+/-2000	U	500	2000	P	01/21/2025	12:16	LB134358
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	01/21/2025	12:16	LB134358
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	12:16	LB134358
	Potassium	2000	+/-2000	U	1600	2000	P	01/21/2025	12:16	LB134358
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	12:16	LB134358
	Silver	1.24	+/-10.0	J	5.00	10.0	P	01/21/2025	12:16	LB134358
	Sodium	2000	+/-2000	U	1000	2000	P	01/21/2025	12:16	LB134358
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	12:16	LB134358
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	12:16	LB134358
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	12:16	LB134358
CCB02	Aluminum	100	+/-100	U	80.0	100	P	01/21/2025	13:06	LB134358
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	01/21/2025	13:06	LB134358
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	13:06	LB134358
	Barium	100	+/-100	U	25.0	100	P	01/21/2025	13:06	LB134358
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	13:06	LB134358
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	13:06	LB134358
	Calcium	2000	+/-2000	U	500	2000	P	01/21/2025	13:06	LB134358
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	13:06	LB134358
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	01/21/2025	13:06	LB134358
	Copper	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	13:06	LB134358
	Iron	100	+/-100	U	80.0	100	P	01/21/2025	13:06	LB134358
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/21/2025	13:06	LB134358
	Magnesium	2000	+/-2000	U	500	2000	P	01/21/2025	13:06	LB134358
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	01/21/2025	13:06	LB134358
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	13:06	LB134358
	Potassium	2000	+/-2000	U	1600	2000	P	01/21/2025	13:06	LB134358
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	13:06	LB134358

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Contract: TETR06

Lab Code: CHEM

Case No.: Q1122

SAS No.: Q1122

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	13:06	LB134358
	Sodium	2000	+/-2000	U	1000	2000	P	01/21/2025	13:06	LB134358
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	13:06	LB134358
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	13:06	LB134358
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	13:06	LB134358
CCB03	Aluminum	100	+/-100	U	80.0	100	P	01/21/2025	13:22	LB134358
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	01/21/2025	13:22	LB134358
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	13:22	LB134358
	Barium	100	+/-100	U	25.0	100	P	01/21/2025	13:22	LB134358
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	13:22	LB134358
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	13:22	LB134358
	Calcium	2000	+/-2000	U	500	2000	P	01/21/2025	13:22	LB134358
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	13:22	LB134358
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	01/21/2025	13:22	LB134358
	Copper	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	13:22	LB134358
	Iron	100	+/-100	U	80.0	100	P	01/21/2025	13:22	LB134358
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/21/2025	13:22	LB134358
	Magnesium	2000	+/-2000	U	500	2000	P	01/21/2025	13:22	LB134358
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	01/21/2025	13:22	LB134358
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	13:22	LB134358
	Potassium	2000	+/-2000	U	1600	2000	P	01/21/2025	13:22	LB134358
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	13:22	LB134358
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	13:22	LB134358
	Sodium	2000	+/-2000	U	1000	2000	P	01/21/2025	13:22	LB134358
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	13:22	LB134358
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	13:22	LB134358
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	13:22	LB134358
CCB04	Aluminum	100	+/-100	U	80.0	100	P	01/21/2025	14:00	LB134358
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	01/21/2025	14:00	LB134358
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	14:00	LB134358
	Barium	100	+/-100	U	25.0	100	P	01/21/2025	14:00	LB134358
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	14:00	LB134358
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	14:00	LB134358
	Calcium	2000	+/-2000	U	500	2000	P	01/21/2025	14:00	LB134358
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	14:00	LB134358
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	01/21/2025	14:00	LB134358
	Copper	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	14:00	LB134358
	Iron	100	+/-100	U	80.0	100	P	01/21/2025	14:00	LB134358
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/21/2025	14:00	LB134358

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tetra Tech NUS, Inc.</u>		SDG No.: <u>Q1122</u>								
Contract: <u>TETR06</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1122</u>	SAS No.: <u>Q1122</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	500	2000	P	01/21/2025	14:00	LB134358
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	01/21/2025	14:00	LB134358
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	14:00	LB134358
	Potassium	2000	+/-2000	U	1600	2000	P	01/21/2025	14:00	LB134358
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	14:00	LB134358
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	14:00	LB134358
	Sodium	2000	+/-2000	U	1000	2000	P	01/21/2025	14:00	LB134358
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	14:00	LB134358
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	14:00	LB134358
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	14:00	LB134358
CCB05	Aluminum	100	+/-100	U	80.0	100	P	01/21/2025	15:07	LB134358
	Antimony	50.0	+/-50.0	U	12.5	50.0	P	01/21/2025	15:07	LB134358
	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	15:07	LB134358
	Barium	100	+/-100	U	25.0	100	P	01/21/2025	15:07	LB134358
	Beryllium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	15:07	LB134358
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	01/21/2025	15:07	LB134358
	Calcium	2000	+/-2000	U	500	2000	P	01/21/2025	15:07	LB134358
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	15:07	LB134358
	Cobalt	30.0	+/-30.0	U	7.50	30.0	P	01/21/2025	15:07	LB134358
	Copper	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	15:07	LB134358
	Iron	100	+/-100	U	80.0	100	P	01/21/2025	15:07	LB134358
	Lead	12.0	+/-12.0	U	9.60	12.0	P	01/21/2025	15:07	LB134358
	Magnesium	2000	+/-2000	U	500	2000	P	01/21/2025	15:07	LB134358
	Manganese	20.0	+/-20.0	U	5.00	20.0	P	01/21/2025	15:07	LB134358
	Nickel	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	15:07	LB134358
	Potassium	2000	+/-2000	U	1600	2000	P	01/21/2025	15:07	LB134358
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	01/21/2025	15:07	LB134358
	Silver	10.0	+/-10.0	U	5.00	10.0	P	01/21/2025	15:07	LB134358
	Sodium	2000	+/-2000	U	1000	2000	P	01/21/2025	15:07	LB134358
	Thallium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	15:07	LB134358
	Vanadium	40.0	+/-40.0	U	20.0	40.0	P	01/21/2025	15:07	LB134358
	Zinc	40.0	+/-40.0	U	10.0	40.0	P	01/21/2025	15:07	LB134358