

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166192BL		WATER		Batch Number:		PB166192		Prep Date:	01/22/2025	
	Mercury	0.20	<0.20	U	0.16	0.20	CV	01/23/2025	10:35	LB134375

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166153BL	WATER			Batch Number:		PB166153		Prep Date:	01/21/2025	
	Aluminum	50.0	<50.0	U	40.0	50.0	P	01/21/2025	14:55	LB134358
	Antimony	25.0	<25.0	U	6.25	25.0	P	01/21/2025	14:55	LB134358
	Arsenic	10.0	<10.0	U	8.00	10.0	P	01/21/2025	14:55	LB134358
	Barium	50.0	<50.0	U	12.5	50.0	P	01/21/2025	14:55	LB134358
	Beryllium	3.00	<3.00	U	0.75	3.00	P	01/21/2025	14:55	LB134358
	Cadmium	3.00	<3.00	U	0.75	3.00	P	01/21/2025	14:55	LB134358
	Calcium	1000	<1000	U	250	1000	P	01/21/2025	14:55	LB134358
	Chromium	5.00	<5.00	U	2.50	5.00	P	01/21/2025	14:55	LB134358
	Cobalt	15.0	<15.0	U	3.75	15.0	P	01/21/2025	14:55	LB134358
	Copper	10.0	<10.0	U	8.00	10.0	P	01/21/2025	14:55	LB134358
	Iron	50.0	<50.0	U	40.0	50.0	P	01/21/2025	14:55	LB134358
	Lead	6.00	<6.00	U	4.80	6.00	P	01/21/2025	14:55	LB134358
	Magnesium	1000	<1000	U	250	1000	P	01/21/2025	14:55	LB134358
	Manganese	10.0	<10.0	U	2.50	10.0	P	01/21/2025	14:55	LB134358
	Nickel	20.0	<20.0	U	5.00	20.0	P	01/21/2025	14:55	LB134358
	Potassium	1000	<1000	U	800	1000	P	01/21/2025	14:55	LB134358
	Selenium	10.0	<10.0	U	8.00	10.0	P	01/21/2025	14:55	LB134358
	Silver	5.00	<5.00	U	2.50	5.00	P	01/21/2025	14:55	LB134358
	Sodium	1000	<1000	U	500	1000	P	01/21/2025	14:55	LB134358
	Thallium	20.0	<20.0	U	10.0	20.0	P	01/21/2025	14:55	LB134358
	Vanadium	20.0	<20.0	U	10.0	20.0	P	01/21/2025	14:55	LB134358
	Zinc	20.0	<20.0	U	5.00	20.0	P	01/21/2025	14:55	LB134358