

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1122
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1122 **SAS No.:** Q1122

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166153BS							
Aluminum	ug/L	1000	903		90	86 - 115	P
Antimony	ug/L	400	381		95	88 - 113	P
Arsenic	ug/L	400	385		96	87 - 113	P
Barium	ug/L	100	88.3		88	88 - 113	P
Beryllium	ug/L	100	90.1		90	89 - 112	P
Cadmium	ug/L	100	92.3		92	88 - 113	P
Calcium	ug/L	500	479	J	96	87 - 113	P
Chromium	ug/L	200	191		96	90 - 113	P
Cobalt	ug/L	100	93.3		93	89 - 114	P
Copper	ug/L	150	149		99	86 - 114	P
Iron	ug/L	1500	1470		98	87 - 115	P
Lead	ug/L	500	464		93	86 - 113	P
Magnesium	ug/L	1000	897	J	90	85 - 113	P
Manganese	ug/L	100	92.9		93	90 - 114	P
Nickel	ug/L	250	234		94	88 - 113	P
Potassium	ug/L	5000	4830		97	86 - 114	P
Selenium	ug/L	1000	955		96	83 - 114	P
Silver	ug/L	37.5	36.3		97	84 - 115	P
Sodium	ug/L	1500	1510		101	87 - 115	P
Thallium	ug/L	1000	1020		102	85 - 114	P
Vanadium	ug/L	150	136		91	90 - 111	P
Zinc	ug/L	100	97.6		98	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q1122

Contract: TETR06

Lab Code: CHEM

Case No.: Q1122

SAS No.: Q1122

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166192BS Mercury	ug/L	4.0	4.02		100	82 - 119	CV