

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: TETR06

Lab Code: CHEM **Case No.:** Q1122 **SAS No.:** Q1122 **SDG NO.:** Q1122

Instrument ID: ECD_P **Date(s) Analyzed:** 01/06/2025 01/07/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.87	4.77	4.97	17785200
		2	4.96	4.86	5.06	12871900
		3	5.04	4.94	5.14	38767800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.04	4.94	5.14	29682400
		2	5.57	5.47	5.67	16809200
		3	5.86	5.76	5.96	31337200
		4	6.02	5.92	6.12	15643200
		5	6.11	6.01	6.21	12044900
Aroclor-1262	500	1	8.28	8.18	8.38	79133000
		2	8.62	8.52	8.72	144417000
		3	8.95	8.85	9.05	102560000
		4	9.04	8.94	9.14	83101200
		5	9.72	9.62	9.82	50821400

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: TETR06

Lab Code: CHEM **Case No.:** Q1122 **SAS No.:** Q1122 **SDG NO.:** Q1122

Instrument ID: ECD_P **Date(s) Analyzed:** 01/06/2025 01/07/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.20	4.10	4.30	13491500
		2	4.29	4.19	4.39	10378200
		3	4.36	4.26	4.46	31049200
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.36	4.26	4.46	24630800
		2	5.10	5.00	5.20	23918600
		3	5.28	5.18	5.38	12560200
		4	5.37	5.27	5.47	11992000
		5	5.54	5.44	5.64	12836200
Aroclor-1262	500	1	7.14	7.04	7.24	79852400
		2	7.40	7.30	7.50	69928200
		3	7.93	7.83	8.03	57593000
		4	7.99	7.89	8.09	100831000
		5	8.51	8.41	8.61	48819800