

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122 SDG NO.: Q1122

Continuing Calib Date: 01/20/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 15:32 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.83       | 5.83      | 5.73      | 5.93  | 0.01       |
| Aroclor-1016-2       | (2) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1016-3       | (3) | 5.91       | 5.92      | 5.82      | 6.02  | 0.01       |
| Aroclor-1016-4       | (4) | 6.01       | 6.02      | 5.92      | 6.12  | 0.01       |
| Aroclor-1016-5       | (5) | 6.30       | 6.31      | 6.21      | 6.41  | 0.01       |
| Aroclor-1260-1       | (1) | 7.43       | 7.43      | 7.33      | 7.53  | 0.00       |
| Aroclor-1260-2       | (2) | 7.68       | 7.69      | 7.59      | 7.79  | 0.01       |
| Aroclor-1260-3       | (3) | 8.04       | 8.05      | 7.95      | 8.15  | 0.01       |
| Aroclor-1260-4       | (4) | 8.27       | 8.28      | 8.18      | 8.38  | 0.01       |
| Aroclor-1260-5       | (5) | 8.60       | 8.61      | 8.51      | 8.71  | 0.01       |
| Tetrachloro-m-xylene |     | 4.67       | 4.67      | 4.57      | 4.77  | 0.00       |
| Decachlorobiphenyl   |     | 10.50      | 10.52     | 10.42     | 10.62 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122 SDG NO.: Q1122

Continuing Calib Date: 01/20/2025 Initial Calibration Date(s): 01/06/2025 01/07/2025

Continuing Calib Time: 15:32 Initial Calibration Time(s): 19:57 03:16

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 5.08       | 5.09      | 4.99      | 5.19 | 0.01       |
| Aroclor-1016-2       | (2) | 5.10       | 5.10      | 5.00      | 5.20 | 0.00       |
| Aroclor-1016-3       | (3) | 5.28       | 5.28      | 5.18      | 5.38 | 0.00       |
| Aroclor-1016-4       | (4) | 5.32       | 5.32      | 5.22      | 5.42 | 0.00       |
| Aroclor-1016-5       | (5) | 5.53       | 5.54      | 5.44      | 5.64 | 0.01       |
| Aroclor-1260-1       | (1) | 6.58       | 6.58      | 6.48      | 6.68 | 0.01       |
| Aroclor-1260-2       | (2) | 6.76       | 6.77      | 6.67      | 6.87 | 0.01       |
| Aroclor-1260-3       | (3) | 6.92       | 6.93      | 6.83      | 7.03 | 0.01       |
| Aroclor-1260-4       | (4) | 7.39       | 7.40      | 7.30      | 7.50 | 0.01       |
| Aroclor-1260-5       | (5) | 7.63       | 7.64      | 7.54      | 7.74 | 0.01       |
| Tetrachloro-m-xylene |     | 3.98       | 3.98      | 3.88      | 4.08 | 0.00       |
| Decachlorobiphenyl   |     | 9.10       | 9.11      | 9.01      | 9.21 | 0.01       |