

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122 SDG NO.: Q1122

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL01 Date Analyzed: 01/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069214.D Time Analyzed: 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.831	5.732	5.932	497.280	500.000	-0.5
Aroclor-1016-2	5.853	5.754	5.954	493.530	500.000	-1.3
Aroclor-1016-3	5.916	5.818	6.018	490.120	500.000	-2.0
Aroclor-1016-4	6.014	5.916	6.116	494.340	500.000	-1.1
Aroclor-1016-5	6.308	6.210	6.410	492.200	500.000	-1.6
Aroclor-1260-1	7.430	7.333	7.533	453.900	500.000	-9.2
Aroclor-1260-2	7.684	7.587	7.787	459.800	500.000	-8.0
Aroclor-1260-3	8.043	7.947	8.147	456.560	500.000	-8.7
Aroclor-1260-4	8.276	8.180	8.380	450.130	500.000	-10.0
Aroclor-1260-5	8.607	8.513	8.713	474.460	500.000	-5.1
Decachlorobiphenyl	10.515	10.424	10.624	45.130	50.000	-9.7
Tetrachloro-m-xylene	4.671	4.572	4.772	51.090	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122 SDG NO.: Q1122

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL01 Date Analyzed: 01/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069214.D Time Analyzed: 09:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.081	4.985	5.185	499.750	500.000	-0.1
Aroclor-1016-2	5.100	5.004	5.204	486.420	500.000	-2.7
Aroclor-1016-3	5.280	5.184	5.384	503.910	500.000	0.8
Aroclor-1016-4	5.320	5.224	5.424	483.450	500.000	-3.3
Aroclor-1016-5	5.537	5.442	5.642	484.280	500.000	-3.1
Aroclor-1260-1	6.578	6.484	6.684	447.610	500.000	-10.5
Aroclor-1260-2	6.765	6.672	6.872	461.170	500.000	-7.8
Aroclor-1260-3	6.922	6.828	7.028	440.440	500.000	-11.9
Aroclor-1260-4	7.395	7.302	7.502	434.350	500.000	-13.1
Aroclor-1260-5	7.635	7.541	7.741	451.490	500.000	-9.7
Decachlorobiphenyl	9.101	9.011	9.211	43.710	50.000	-12.6
Tetrachloro-m-xylene	3.981	3.883	4.083	50.460	50.000	0.9