

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122 SDG NO.: Q1122

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL02 Date Analyzed: 01/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069229.D Time Analyzed: 15:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.825	5.732	5.932	482.550	500.000	-3.5
Aroclor-1016-2	5.847	5.754	5.954	500.800	500.000	0.2
Aroclor-1016-3	5.910	5.818	6.018	487.650	500.000	-2.5
Aroclor-1016-4	6.008	5.916	6.116	490.880	500.000	-1.8
Aroclor-1016-5	6.302	6.210	6.410	495.370	500.000	-0.9
Aroclor-1260-1	7.425	7.333	7.533	462.560	500.000	-7.5
Aroclor-1260-2	7.678	7.587	7.787	461.280	500.000	-7.7
Aroclor-1260-3	8.037	7.947	8.147	443.610	500.000	-11.3
Aroclor-1260-4	8.270	8.180	8.380	457.840	500.000	-8.4
Aroclor-1260-5	8.600	8.513	8.713	472.770	500.000	-5.4
Decachlorobiphenyl	10.503	10.424	10.624	45.750	50.000	-8.5
Tetrachloro-m-xylene	4.666	4.572	4.772	50.640	50.000	1.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR06

Lab Code: CHEM Case No.: Q1122 SAS No.: Q1122 SDG NO.: Q1122

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL02 Date Analyzed: 01/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP069229.D Time Analyzed: 15:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.077	4.985	5.185	484.470	500.000	-3.1
Aroclor-1016-2	5.097	5.004	5.204	465.170	500.000	-7.0
Aroclor-1016-3	5.276	5.184	5.384	497.370	500.000	-0.5
Aroclor-1016-4	5.316	5.224	5.424	473.090	500.000	-5.4
Aroclor-1016-5	5.534	5.442	5.642	450.430	500.000	-9.9
Aroclor-1260-1	6.575	6.484	6.684	420.560	500.000	-15.9
Aroclor-1260-2	6.762	6.672	6.872	434.700	500.000	-13.1
Aroclor-1260-3	6.918	6.828	7.028	404.100	500.000	-19.2
Aroclor-1260-4	7.391	7.302	7.502	393.810	500.000	-21.2
Aroclor-1260-5	7.631	7.541	7.741	404.540	500.000	-19.1
Decachlorobiphenyl	9.096	9.011	9.211	40.770	50.000	-18.5
Tetrachloro-m-xylene	3.978	3.883	4.083	48.120	50.000	-3.8