

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010625\
 Data File : PP068935.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jan 2025 01:22
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/07/2025
 Supervised By :Ankita Jodhani 01/07/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 07 04:55:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 07 04:31:10 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.672	3.983	37802873	26914889	25.828	25.480
2) SA Decachlor...	10.524	9.112	30405057	32949823	26.113	27.242
Target Compounds						
26) L6 AR-1254-1	6.686	5.898	14519383	16528283	279.371m	274.651
27) L6 AR-1254-2	6.904	6.045	21595154	14763310	263.305	276.914
28) L6 AR-1254-3	7.268	6.454	21385186	22387356	263.197	269.680
29) L6 AR-1254-4	7.551	6.682	16198411	12382863	263.510	273.516
30) L6 AR-1254-5	7.968	7.102	16873594	19999727	252.947m	263.558

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP010625\
Data File : PP068935.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jan 2025 01:22
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/07/2025
Supervised By :Ankita Jodhani 01/07/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 07 04:55:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP010625.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 07 04:31:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.672 min
Delta R.T.: 0.000 min
Response: 37802873
Conc: 25.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/07/2025
Supervised By :Ankita Jodhani 01/07/2025




#1 Tetrachloro-m-xylene

R.T.: 3.983 min
Delta R.T.: 0.000 min
Response: 26914889
Conc: 25.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.524 min
Delta R.T.: 0.000 min
Response: 30405057
Conc: 26.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.112 min
Delta R.T.: 0.000 min
Response: 32949823
Conc: 27.24 ng/ml




#26 AR-1254-1

R.T.: 6.686 min
Delta R.T.: 0.000 min
Response: 14519383
Conc: 279.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/07/2025
Supervised By :Ankita Jodhani 01/07/2025




#26 AR-1254-1

R.T.: 5.898 min
Delta R.T.: 0.000 min
Response: 16528283
Conc: 274.65 ng/ml




#27 AR-1254-2

R.T.: 6.904 min
Delta R.T.: 0.000 min
Response: 21595154
Conc: 263.30 ng/ml



#27 AR-1254-2

R.T.: 6.045 min
Delta R.T.: 0.000 min
Response: 14763310
Conc: 276.91 ng/ml




#28 AR-1254-3

R.T.: 7.268 min
Delta R.T.: -0.001 min
Response: 21385186
Conc: 263.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/07/2025
Supervised By :Ankita Jodhani 01/07/2025




#28 AR-1254-3

R.T.: 6.454 min
Delta R.T.: 0.000 min
Response: 22387356
Conc: 269.68 ng/ml




#29 AR-1254-4

R.T.: 7.551 min
Delta R.T.: 0.000 min
Response: 16198411
Conc: 263.51 ng/ml



#29 AR-1254-4

R.T.: 6.682 min
Delta R.T.: 0.000 min
Response: 12382863
Conc: 273.52 ng/ml




#30 AR-1254-5

R.T.: 7.968 min
Delta R.T.: -0.001 min
Response: 16873594
Conc: 252.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/07/2025
Supervised By :Ankita Jodhani 01/07/2025



#30 AR-1254-5

R.T.: 7.102 min
Delta R.T.: 0.000 min
Response: 19999727
Conc: 263.56 ng/ml