

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP010625

Review By	yogesh	Review On	1/7/2025 7:49:59 AM
Supervise By	Ankita	Supervise On	1/7/2025 10:29:04 AM
SubDirectory	PP010625	HP Acquire Method	HP Processing Method PP010625

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP068913.D	06 Jan 2025 19:24		YP\AJ	Ok
2	I.BLK	I.BLK	PP068914.D	06 Jan 2025 19:41		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP068915.D	06 Jan 2025 19:57		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP068916.D	06 Jan 2025 20:13		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP068917.D	06 Jan 2025 20:30		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP068918.D	06 Jan 2025 20:46		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP068919.D	06 Jan 2025 21:02		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP068920.D	06 Jan 2025 21:19		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP068921.D	06 Jan 2025 21:35		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP068922.D	06 Jan 2025 21:51		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP068923.D	06 Jan 2025 22:07		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP068924.D	06 Jan 2025 22:24		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP068925.D	06 Jan 2025 22:40		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP068926.D	06 Jan 2025 22:56		YP\AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP068927.D	06 Jan 2025 23:12		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP068928.D	06 Jan 2025 23:29		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP068929.D	06 Jan 2025 23:45		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP068930.D	07 Jan 2025 00:01		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP010625

Review By	yogesh	Review On	1/7/2025 7:49:59 AM
Supervise By	Ankita	Supervise On	1/7/2025 10:29:04 AM
SubDirectory	PP010625	HP Acquire Method	HP Processing Method PP010625
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP068931.D	07 Jan 2025 00:17		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP068932.D	07 Jan 2025 00:34		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP068933.D	07 Jan 2025 00:50		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP068934.D	07 Jan 2025 01:06		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP068935.D	07 Jan 2025 01:22		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP068936.D	07 Jan 2025 01:39		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP068937.D	07 Jan 2025 01:55		YPIAJ	Ok,M
26	AR1268ICC1000	AR1268ICC1000	PP068938.D	07 Jan 2025 02:11		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP068939.D	07 Jan 2025 02:27		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP068940.D	07 Jan 2025 02:44		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP068941.D	07 Jan 2025 03:00		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP068942.D	07 Jan 2025 03:16		YPIAJ	Ok,M
31	PP010625ICV500	ICVPP010625	PP068943.D	07 Jan 2025 03:32		YPIAJ	Ok
32	AR1242ICV500	ICVPP010625AR1242	PP068944.D	07 Jan 2025 03:48		YPIAJ	Ok
33	AR1248ICV500	ICVPP010625AR1248	PP068945.D	07 Jan 2025 04:05		YPIAJ	Ok
34	AR1254ICV500	ICVPP010625AR1254	PP068946.D	07 Jan 2025 04:21		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP010625AR1268	PP068947.D	07 Jan 2025 04:37		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP012025

Review By	yogesh	Review On	1/21/2025 8:42:23 AM
Supervise By	Ankita	Supervise On	1/21/2025 8:42:42 AM
SubDirectory	PP012025	HP Acquire Method	HP Processing Method PP010625

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069213.D	20 Jan 2025 08:44		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP069214.D	20 Jan 2025 09:00		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP069215.D	20 Jan 2025 09:16		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP069216.D	20 Jan 2025 09:32		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP069217.D	20 Jan 2025 09:49		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP069218.D	20 Jan 2025 10:05		YPIAJ	Ok
7	PB166124BL	PB166124BL	PP069219.D	20 Jan 2025 10:21		YPIAJ	Ok
8	PB166124BS	PB166124BS	PP069220.D	20 Jan 2025 10:37		YPIAJ	Ok
9	PB166124BSD	PB166124BSD	PP069221.D	20 Jan 2025 10:54		YPIAJ	Ok
10	Q1122-01	RW10A-20250116	PP069222.D	20 Jan 2025 11:10		YPIAJ	Ok
11	PB166127BL	PB166127BL	PP069223.D	20 Jan 2025 11:59		YPIAJ	Ok
12	PB166127BS	PB166127BS	PP069224.D	20 Jan 2025 12:15		YPIAJ	Ok
13	Q1123-01	BU-03-1172025	PP069225.D	20 Jan 2025 12:31		YPIAJ	Ok
14	Q1132-01	HD-1-011725	PP069226.D	20 Jan 2025 12:47		YPIAJ	Ok
15	Q1132-04	HD-2-011725	PP069227.D	20 Jan 2025 13:04		YPIAJ	Ok
16	Q1133-01	TR-04-1172025	PP069228.D	20 Jan 2025 13:20	Reanalysed in PP069234.D	YPIAJ	Not Ok
17	AR1660CCC500	AR1660CCC500	PP069229.D	20 Jan 2025 15:32		YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP069230.D	20 Jan 2025 15:48		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP012025

Review By	yogesh	Review On	1/21/2025 8:42:23 AM
Supervise By	Ankita	Supervise On	1/21/2025 8:42:42 AM
SubDirectory	PP012025	HP Acquire Method	HP Processing Method PP010625
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP069231.D	20 Jan 2025 16:04		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP069232.D	20 Jan 2025 16:21		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP069233.D	20 Jan 2025 16:37		YPIAJ	Ok
22	Q1133-01	TR-04-1172025	PP069234.D	20 Jan 2025 16:53	AR1254 Hit, Closing CCC fail	YPIAJ	Not Ok
23	Q1134-01	EO-1-011725	PP069235.D	20 Jan 2025 17:09	Closing CCC fail	YPIAJ	Not Ok
24	Q1134-01MS	EO-1-011725MS	PP069236.D	20 Jan 2025 17:26	Closing CCC fail	YPIAJ	Not Ok
25	Q1134-01MSD	EO-1-011725MSD	PP069237.D	20 Jan 2025 17:42	Closing CCC fail	YPIAJ	Not Ok
26	AR1660CCC500	AR1660CCC500	PP069238.D	20 Jan 2025 18:51	CCC Fail	YPIAJ	Not Ok
27	AR1242CCC500	AR1242CCC500	PP069239.D	20 Jan 2025 19:08		YPIAJ	Ok
28	AR1248CCC500	AR1248CCC500	PP069240.D	20 Jan 2025 19:24		YPIAJ	Ok
29	AR1254CCC500	AR1254CCC500	PP069241.D	20 Jan 2025 19:40		YPIAJ	Ok,M
30	I.BLK	I.BLK	PP069242.D	20 Jan 2025 19:56		YPIAJ	Ok

M : Manual Integration