

Surrogate Summary

SDG No.: Q1122

Client: Tetra Tech NUS, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL093481.D	PIBLK-PL093481.D	Decachlorobiphenyl	1	20	22.0	110		30	135
		Tetrachloro-m-xylene	1	20	20.5	102		44	124
		Decachlorobiphenyl	2	20	21.4	107		30	135
		Tetrachloro-m-xylene	2	20	19.9	100		44	124
I.BLK-PL093694.D	PIBLK-PL093694.D	Decachlorobiphenyl	1	20	21.9	110		30	135
		Tetrachloro-m-xylene	1	20	21.7	109		44	124
		Decachlorobiphenyl	2	20	19.4	97		30	135
		Tetrachloro-m-xylene	2	20	21.5	107		44	124
PB166101BL	PB166101BL	Decachlorobiphenyl	1	20	22.2	111		30	135
		Tetrachloro-m-xylene	1	20	20.0	100		44	124
		Decachlorobiphenyl	2	20	21.1	105		30	135
		Tetrachloro-m-xylene	2	20	19.2	96		44	124
PB166101BS	PB166101BS	Decachlorobiphenyl	1	20	20.4	102		30	135
		Tetrachloro-m-xylene	1	20	18.4	92		44	124
		Decachlorobiphenyl	2	20	19.7	99		30	135
		Tetrachloro-m-xylene	2	20	17.9	90		44	124
PB166101BSD	PB166101BSD	Decachlorobiphenyl	1	20	20.4	102		30	135
		Tetrachloro-m-xylene	1	20	18.1	90		44	124
		Decachlorobiphenyl	2	20	19.7	99		30	135
		Tetrachloro-m-xylene	2	20	17.3	87		44	124
Q1122-01	RW10A-20250116	Decachlorobiphenyl	1	20	23.0	115		30	135
		Tetrachloro-m-xylene	1	20	21.0	105		44	124
		Decachlorobiphenyl	2	20	22.8	114		30	135
		Tetrachloro-m-xylene	2	20	21.3	106		44	124
I.BLK-PL093706.D	PIBLK-PL093706.D	Decachlorobiphenyl	1	20	23.3	117		30	135
		Tetrachloro-m-xylene	1	20	21.6	108		44	124
		Decachlorobiphenyl	2	20	23.0	115		30	135
		Tetrachloro-m-xylene	2	20	21.1	105		44	124