

Surrogate Summary

SW-846

SDG No.: Q1122

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166117BL	PB166117BL	2-Fluorophenol	150	144	96		19	119
		Phenol-d6	150	141	94		10	130
		Nitrobenzene-d5	100	97.8	98		44	120
		2-Fluorobiphenyl	100	98.6	99		44	119
		2,4,6-Tribromophenol	150	162	108		43	140
		Terphenyl-d14	100	92.3	92		50	134
PB166117BS	PB166117BS	2-Fluorophenol	150	136	91		19	119
		Phenol-d6	150	135	90		10	130
		Nitrobenzene-d5	100	94.7	95		44	120
		2-Fluorobiphenyl	100	95.4	95		44	119
		2,4,6-Tribromophenol	150	157	105		43	140
		Terphenyl-d14	100	96.5	97		50	134
PB166117BSD	PB166117BSD	2-Fluorophenol	150	134	90		19	119
		Phenol-d6	150	135	90		10	130
		Nitrobenzene-d5	100	96.2	96		44	120
		2-Fluorobiphenyl	100	96.2	96		44	119
		2,4,6-Tribromophenol	150	160	106		43	140
		Terphenyl-d14	100	99.3	99		50	134
Q1122-01	RW10A-20250116	2-Fluorophenol	150	51.2	34		19	119
		Phenol-d6	150	30.6	20		10	130
		Nitrobenzene-d5	100	86.1	86		44	120
		2-Fluorobiphenyl	100	88.3	88		44	119
		2,4,6-Tribromophenol	150	161	107		43	140
		Terphenyl-d14	100	80.4	80		50	134