

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1124 SAS No.: Q1124 SDG NO.: Q1124

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/16/2025 01/16/2025

Client Sample No.: CCAL01 Date Analyzed: 01/21/2025

Lab Sample No.: AR1660CCC400 Data File : PR070041.D Time Analyzed: 08:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.848	4.748	4.948	28147810.000	26952600.000	-4.2
Aroclor-1016-2	4.871	4.770	4.970	44488680.000	43163000.000	-3.0
Aroclor-1016-3	4.934	4.834	5.034	28020020.000	27000200.000	-3.6
Aroclor-1016-4	5.036	4.936	5.136	21300440.000	20393400.000	-4.3
Aroclor-1016-5	5.351	5.250	5.450	23842170.000	22261100.000	-6.6
Aroclor-1260-1	6.562	6.463	6.663	41448590.000	38650900.000	-6.8
Aroclor-1260-2	6.846	6.745	6.945	43262170.000	41378800.000	-4.4
Aroclor-1260-3	7.236	7.135	7.335	36130430.000	34617800.000	-4.2
Aroclor-1260-4	7.478	7.379	7.579	39102050.000	37045900.000	-5.3
Aroclor-1260-5	7.820	7.720	7.920	62626270.000	59721300.000	-4.6
DCB	9.517	9.417	9.617	517457800.000	496549000.000	-4.1
TCX	3.629	3.528	3.728	1116257000.000	1046960000.000	-6.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1124 SAS No.: Q1124 SDG NO.: Q1124

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/16/2025 01/16/2025

Client Sample No.: CCAL01 Date Analyzed: 01/21/2025

Lab Sample No.: AR1660CCC400 Data File : PR070041.D Time Analyzed: 08:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.077	3.978	4.178	180235200.000	165658000.000	-8.1
Aroclor-1016-2	4.095	3.996	4.196	264477200.000	248764000.000	-5.9
Aroclor-1016-3	4.275	4.176	4.376	140422600.000	129783000.000	-7.6
Aroclor-1016-4	4.326	4.227	4.427	124420800.000	113069000.000	-9.1
Aroclor-1016-5	4.545	4.445	4.645	150359800.000	135431000.000	-9.9
Aroclor-1260-1	5.639	5.540	5.740	249242700.000	226107000.000	-9.3
Aroclor-1260-2	5.846	5.747	5.947	272617200.000	252146000.000	-7.5
Aroclor-1260-3	6.004	5.905	6.105	272281900.000	249646000.000	-8.3
Aroclor-1260-4	6.511	6.413	6.613	228965500.000	208207000.000	-9.1
Aroclor-1260-5	6.777	6.679	6.879	488567400.000	450283000.000	-7.8
DCB	8.228	8.133	8.333	3514619000.000	3247140000.000	-7.6
TCX	2.959	2.861	3.061	6357247000.000	5892850000.000	-7.3