

**CALIBRATION VERIFICATION SUMMARY**

Contract: CHEM02

Lab Code: CHEM Case No.: Q1124 SAS No.: Q1124 SDG NO.: Q1124

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/16/2025 01/16/2025

Client Sample No.: CCAL02 Date Analyzed: 01/21/2025

Lab Sample No.: AR1660CCC400 Data File : PR070045.D Time Analyzed: 11:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.848	4.748	4.948	28147810.000	27774100.000	-1.3
Aroclor-1016-2	4.870	4.770	4.970	44488680.000	43594600.000	-2.0
Aroclor-1016-3	4.934	4.834	5.034	28020020.000	27123500.000	-3.2
Aroclor-1016-4	5.036	4.936	5.136	21300440.000	20578000.000	-3.4
Aroclor-1016-5	5.350	5.250	5.450	23842170.000	22673700.000	-4.9
Aroclor-1260-1	6.562	6.463	6.663	41448590.000	40000200.000	-3.5
Aroclor-1260-2	6.845	6.745	6.945	43262170.000	41752200.000	-3.5
Aroclor-1260-3	7.235	7.135	7.335	36130430.000	34802000.000	-3.7
Aroclor-1260-4	7.478	7.379	7.579	39102050.000	37576900.000	-3.9
Aroclor-1260-5	7.819	7.720	7.920	62626270.000	60036200.000	-4.1
DCB	9.514	9.417	9.617	517457800.000	528141000.000	2.1
TCX	3.628	3.528	3.728	1116257000.000	1064020000.000	-4.7

**CALIBRATION VERIFICATION SUMMARY**

Contract: CHEM02

Lab Code: CHEM Case No.: Q1124 SAS No.: Q1124 SDG NO.: Q1124

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/16/2025 01/16/2025

Client Sample No.: CCAL02 Date Analyzed: 01/21/2025

Lab Sample No.: AR1660CCC400 Data File : PR070045.D Time Analyzed: 11:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.078	3.978	4.178	180235200.000	170128000.000	-5.6
Aroclor-1016-2	4.096	3.996	4.196	264477200.000	252505000.000	-4.5
Aroclor-1016-3	4.275	4.176	4.376	140422600.000	132526000.000	-5.6
Aroclor-1016-4	4.326	4.227	4.427	124420800.000	115608000.000	-7.1
Aroclor-1016-5	4.544	4.445	4.645	150359800.000	136614000.000	-9.1
Aroclor-1260-1	5.640	5.540	5.740	249242700.000	231669000.000	-7.1
Aroclor-1260-2	5.846	5.747	5.947	272617200.000	256016000.000	-6.1
Aroclor-1260-3	6.004	5.905	6.105	272281900.000	254129000.000	-6.7
Aroclor-1260-4	6.511	6.413	6.613	228965500.000	211203000.000	-7.8
Aroclor-1260-5	6.778	6.679	6.879	488567400.000	459995000.000	-5.8
DCB	8.229	8.133	8.333	3514619000.000	3505120000.000	-0.3
TCX	2.959	2.861	3.061	6357247000.000	5989180000.000	-5.8