

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1124 SAS No.: Q1124 SDG No.: Q1124

Instrument ID: MSVOA\_N Calibration Date/Time: 01/17/2025 09:06

Lab File ID: VN085480.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND           | Amount | Conc. | %Dev    | MAX%D |
|--------------------|--------|-------|---------|-------|
| p-Isopropyltoluene | 50.000 | 55.07 | -10.142 | 20    |