

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: Q1124 SAS No.: Q1124 SDG NO.: Q1124
Lab File ID: VN085437.D BFB Injection Date: 01/14/2025
Instrument ID: MSVOA_N BFB Injection Time: 14:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.3
75	30.0 - 60.0% of mass 95	58
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.8
173	Less than 2.0% of mass 174	1.4 (1.8) 1
174	50.0 - 100.0% of mass 95	76
175	5.0 - 9.0% of mass 174	5.4 (7.1) 1
176	95.0 - 101.0% of mass 174	74.1 (97.4) 1
177	5.0 - 9.0% of mass 176	4.9 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085438.D	01/14/2025	14:56
VSTDICCC050	VSTDICCC050	VN085439.D	01/14/2025	15:19
VSTDICC020	VSTDICC020	VN085440.D	01/14/2025	15:43
VSTDICC010	VSTDICC010	VN085441.D	01/14/2025	16:07
VSTDICC005	VSTDICC005	VN085442.D	01/14/2025	16:31
VSTDICC001	VSTDICC001	VN085443.D	01/14/2025	17:19