


LB134496

SFAM01.1

INSTRUMENT ITS: CV2



Linear

A= 0.0000e+000
B= 2.0278e-004 slope
C= 6.6237e-003 y-intercept
Rho= 0.9992387
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	400
0.00	0.000	0.008	0.008	9	0.000	9					
0.05	0.050										
0.20	0.200	0.246	0.046	1181	0.0 %	1181					
2.50	2.500	2.450	-0.050	12048	0.0 %	12048					
5.00	5.000	4.853	-0.147	23897	0.0 %	23897					
7.50	7.500	7.781	0.281	38338	0.0 %	38338					
10.0	10.000	9.862	-0.138	48602	0.0 %	48602					

400
= 23
- 2
- 3
4
- 1

LB134496

INSTRUMENT ID:CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	9	-		0 SFAM01.1	PPB	1/31/2025 10:44	Std
	0.2 S01	1181	-		0.2 SFAM01.1	PPB	1/31/2025 10:47	Std
	2.5 S02	12048	-		2.5 SFAM01.1	PPB	1/31/2025 10:52	Std
	5 S03	23897	-		5 SFAM01.1	PPB	1/31/2025 10:54	Std
	7.5 S04	38338	-		7.5 SFAM01.1	PPB	1/31/2025 10:58	Std
	10 S05	48602	-		10 SFAM01.1	PPB	1/31/2025 11:00	Std
ICV076	ICV076	18392	3.7362 -		SFAM01.1	PPB	1/31/2025 11:04	SMPL
ICB076	ICB076	-217	-0.0374 -		SFAM01.1	PPB	1/31/2025 11:07	SMPL
CCV090	CCV090	24185	4.9109 -		SFAM01.1	PPB	1/31/2025 11:09	SMPL
CCB090	CCB090	-234	-0.0408 -		SFAM01.1	PPB	1/31/2025 11:11	SMPL
PB166430BL	PBW430	-70	-0.0076 -		SFAM01.1	PPB	1/31/2025 11:13	SMPL
Q1089-01	MC0B00	297	0.0669 -		SFAM01.1	PPB	1/31/2025 11:16	SMPL
Q1089-02	MC0B01	77	0.0222 -		SFAM01.1	PPB	1/31/2025 11:18	SMPL
Q1089-03	MC0B01D	309	0.0693 -		SFAM01.1	PPB	1/31/2025 11:20	SMPL
Q1089-04	MC0B01S	4997	1.0199 -		SFAM01.1	PPB	1/31/2025 11:23	SMPL
Q1089-05	MC0B06	14	0.0095 -		SFAM01.1	PPB	1/31/2025 11:25	SMPL
Q1089-06	MC0B07	-6	0.0054 -		SFAM01.1	PPB	1/31/2025 11:27	SMPL
Q1089-07	MC0AL9	22	0.0111 -		SFAM01.1	PPB	1/31/2025 11:29	SMPL
Q1089-08	MC0AM2	14	0.0095 -		SFAM01.1	PPB	1/31/2025 11:32	SMPL
Q1089-09	MC0AZ9	78	0.0224 -		SFAM01.1	PPB	1/31/2025 11:34	SMPL
Q1089-10	MC0B02	292	0.0658 -		SFAM01.1	PPB	1/31/2025 11:36	SMPL
Q1089-11	MC0B04	133	0.0336 -		SFAM01.1	PPB	1/31/2025 11:38	SMPL
Q1089-12	MC0B09	106	0.0281 -		SFAM01.1	PPB	1/31/2025 11:41	SMPL
Q1089-13	MC0B10	21	0.0109 -		SFAM01.1	PPB	1/31/2025 11:43	SMPL
Q1089-14	MC0B11	60	0.0188 -		SFAM01.1	PPB	1/31/2025 11:45	SMPL
Q1127-09	MJNLE5	135	0.034 -		SFAM01.1	PPB	1/31/2025 11:47	SMPL
Q1127-10	MJNLE5D	107	0.0283 -		SFAM01.1	PPB	1/31/2025 11:50	SMPL
Q1127-11	MJNLE5S	5982	1.2197 -		SFAM01.1	PPB	1/31/2025 11:52	SMPL
CCV091	CCV091	23141	4.6992 -		SFAM01.1	PPB	1/31/2025 11:54	SMPL
CCB091	CCB091	-170	-0.0278 -		SFAM01.1	PPB	1/31/2025 11:57	SMPL