

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51956 MA No.: _____ SDG No.: MJNLB3
Instrument ID: P8 Start Date: 01/21/2025
Analytical Method: ICP-MS Run Batch: LB134585
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0	1	1.3	30	50	52	4
Copper	0	0.0	0	2	2.3	15	500	560	12
Lead	0	0.0	0	1	0.99	-1	250	240	-4
Zinc	0	0.0	0	5	5.8	16	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51956 MA No.: SDG No.: MJNLB3
Instrument ID: P8 Start Date: 01/21/2025
Analytical Method: ICP-MS Run Batch: LB134585
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	125	130	4	250	270	8	500	520	4
Copper	1250	1400	12	2500	2700	8	5000	5400	8
Lead	625	640	2	1250	1300	4	2500	2500	0
Zinc	1250	1300	4	2500	2600	4	5000	5400	8

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51956 MA No.: _____ SDG No.: MJNLB3
Instrument ID: P8 Start Date: 01/21/2025
Analytical Method: ICP-MS Run Batch: LB134585
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	1000	990	-1						
Copper	10000	9700	-3						
Lead	5000	5000	0						
Zinc	10000	9800	-2						