

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51956 MA No. : SDG No.: MJNLB3
 Initial Calibration Verification Source : MP84051
 Continuing Calibration Verification Source : MP84052
 Run Batch: LB134585 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	200	200	101	1	500	490	99	2	490	97	1
Copper	100	110	106	1	5000	4700	94	1	4700	94	1
Lead	200	210	104	13	2500	2500	101	1	2500	100	2
Zinc	200	200	101	1	5000	4700	95	1	4800	95	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51956 MA No. : SDG No.: MJNLB3
 Initial Calibration Verification Source : MP84051
 Continuing Calibration Verification Source : MP84052
 Run Batch: LB134585 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID: CCV004		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					500	470	95	3	490	97	1
Copper					5000	4700	93	1	4700	94	0
Lead					2500	2600	103	2	2500	101	1
Zinc					5000	4700	94	1	4800	95	1