EPA SAMPLE NO.

MJNLB5L

FORM 8-IN

ICP-AES AND ICP-MS SERIAL DILUTIONS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51956 MA No.: SDG No.: MJNLB3

Matrix : Water Analytical Method: ICP-MS

% Solids:

Concentration Units (µg/L, mg/L or mg/kg dry weight): ug/L

Analyte	Initial Sample Result (I)	Q	Serial Dilution Result(S)	Q	% Difference	Q
Arsenic	15		15		5	
Copper	100		110		8	
Lead	3.4		3.3	J	4	
Zinc	9.3		9.6	J	3	