

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51956 MA No. :                      SDG No.: MJNLB4  
 Initial Calibration Verification Source : MP84272  
 Continuing Calibration Verification Source : MP84052  
 Run Batch: LB134585 Analytical Method: ICP-MS  
 Concentration Units: µg/L

|         | Initial Calibration Verification |       |     |      | Continuing Calibration Verification |       |     |      |            |     |      |
|---------|----------------------------------|-------|-----|------|-------------------------------------|-------|-----|------|------------|-----|------|
|         | ID: ICV001                       |       |     |      | ID: CCV001                          |       |     |      | ID: CCV002 |     |      |
| Analyte | True                             | Found | %R  | %RSD | True                                | Found | %R  | %RSD | Found      | %R  | %RSD |
| Arsenic | 200                              | 200   | 101 | 1    | 500                                 | 490   | 99  | 2    | 490        | 97  | 1    |
| Copper  | 100                              | 110   | 106 | 1    | 5000                                | 4700  | 94  | 1    | 4700       | 94  | 1    |
| Lead    | 200                              | 210   | 104 | 13   | 2500                                | 2500  | 101 | 1    | 2500       | 100 | 2    |
| Zinc    | 200                              | 200   | 101 | 1    | 5000                                | 4700  | 95  | 1    | 4800       | 95  | 0    |

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51956 MA No. :                      SDG No.: MJNLB4  
 Initial Calibration Verification Source : MP84272  
 Continuing Calibration Verification Source : MP84052  
 Run Batch: LB134585 Analytical Method: ICP-MS  
 Concentration Units: µg/L

|         | Initial Calibration Verification |       |    |      | Continuing Calibration Verification |       |     |      |            |     |      |
|---------|----------------------------------|-------|----|------|-------------------------------------|-------|-----|------|------------|-----|------|
|         | ID:                              |       |    |      | ID: CCV003                          |       |     |      | ID: CCV004 |     |      |
| Analyte | True                             | Found | %R | %RSD | True                                | Found | %R  | %RSD | Found      | %R  | %RSD |
| Arsenic |                                  |       |    |      | 500                                 | 470   | 95  | 3    | 490        | 97  | 1    |
| Copper  |                                  |       |    |      | 5000                                | 4700  | 93  | 1    | 4700       | 94  | 0    |
| Lead    |                                  |       |    |      | 2500                                | 2600  | 103 | 2    | 2500       | 101 | 1    |
| Zinc    |                                  |       |    |      | 5000                                | 4700  | 94  | 1    | 4800       | 95  | 1    |