

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51956 MA No. : SDG No.: MJNLA3
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV013				ID: CCV014		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1000	100	0	5000	4800	96	0	4800	96	0
Barium	520	550	105	0	10000	10000	103	1	9800	98	1
Cadmium	510	490	97	0	2500	2500	98	0	2500	98	0
Chromium	520	530	103	0	15500	16000	101	0	15000	100	0
Lead	1000	980	98	0	25000	24000	97	0	24000	97	0
Selenium	1000	1100	107	0	5000	4800	97	1	4800	97	0
Silver	250	250	100	0	1250	1300	103	0	1300	101	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51956 MA No. : SDG No.: MJNLA3
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV015				ID: CCV016		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	4700	94	0	4900	98	2
Barium					10000	9400	94	1	10000	102	2
Cadmium					2500	2400	97	0	2400	95	2
Chromium					15500	15000	99	0	15000	95	0
Lead					25000	24000	96	0	24000	96	2
Selenium					5000	4700	94	0	5000	100	2
Silver					1250	1300	101	0	1200	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51956 MA No. : SDG No.: MJNLA3
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV017				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	4700	95	0			
Barium					10000	11000	107	1			
Cadmium					2500	2300	93	0			
Chromium					15500	15000	95	0			
Lead					25000	23000	93	0			
Selenium					5000	4800	97	0			
Silver					1250	1300	101	0			