

LABORATORY CONTROL SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51956 MA No.: _____ SDG No.: MJNLA3
Matrix : Water Preparation Method: 200.7
Analytical Method: ICP-AES Preparation Batch: PB166162
Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg): $\mu\text{g/L}$

Analyte	True	Found	%R
Arsenic	20	19	93
Barium	400	380	94
Cadmium	10	9.6	96
Chromium	20	21	105
Lead	20	18	89
Selenium	70	62	88
Silver	20	19	94