


LB134366

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
B= 2.4269e-004 slope
C= -3.7195e-002 + intercept
Rho= 0.9996646
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%))
0.00	0.000	-0.028	-0.028	39	0.000	39	0				=
0.05	0.050					442	0				-6
0.20	0.200	0.189	-0.011	932	0.0 %	932					-1
2.50	2.500	2.486	-0.014	10398	0.0 %	10398					0
5.00	5.000	5.019	0.019	20833	0.0 %	20833					2
7.50	7.500	7.679	0.179	31795	0.0 %	31795					-1
10.0	10.000	9.854	-0.146	40757	0.0 %	40757					

LB134366

INSTRUMENT ID: CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	39	-		0 SFAM01.1	PPB	01/22/2025 13:59	Std
	0.2 S01	932	-		0.2 SFAM01.1	PPB	01/22/2025 14:02	Std
	2.5 S02	10398	-		2.5 SFAM01.1	PPB	01/22/2025 14:04	Std
	5 S03	20833	-		5 SFAM01.1	PPB	01/22/2025 14:11	Std
	7.5 S04	31795	-		7.5 SFAM01.1	PPB	01/22/2025 14:13	Std
	10 S05	40757	-		10 SFAM01.1	PPB	01/22/2025 14:15	Std
ICV065	ICV065	15165	3.6433	-	SFAM01.1	PPB	01/22/2025 14:18	SMPL
ICB065	ICB065	-174	-0.0794	-	SFAM01.1	PPB	01/22/2025 14:21	SMPL
CCV059	CCV059	20679	4.9815	-	SFAM01.1	PPB	01/22/2025 14:23	SMPL
CCB059	CCB059	-200	-0.0857	-	SFAM01.1	PPB	01/22/2025 14:25	SMPL
PB166178BL	PBW178	-52	-0.0498	-	SFAM01.1	PPB	01/22/2025 14:28	SMPL
Q1106-01	YE8G3	4	-0.0362	-	SFAM01.1	PPB	01/22/2025 14:30	SMPL
Q1106-02	YE8G5	2633	0.6018	-	SFAM01.1	PPB	01/22/2025 14:32	SMPL
Q1106-03	YE8G6	-182	-0.0814	-	SFAM01.1	PPB	01/22/2025 14:34	SMPL
Q1106-04	YE8G6D	-43	-0.0476	-	SFAM01.1	PPB	01/22/2025 14:37	SMPL
Q1106-05	YE8G6S	3982	0.9292	-	SFAM01.1	PPB	01/22/2025 14:39	SMPL
Q1106-06	YE8H0	539	0.0936	-	SFAM01.1	PPB	01/22/2025 14:41	SMPL
Q1106-07	YE8H1	1466	0.3186	-	SFAM01.1	PPB	01/22/2025 14:43	SMPL
Q1106-08	YE8H2	547	0.0956	-	SFAM01.1	PPB	01/22/2025 14:46	SMPL
Q1106-09	YE8H3	274	0.0293	-	SFAM01.1	PPB	01/22/2025 14:48	SMPL
Q1106-10	YE8H4	-46	-0.0484	-	SFAM01.1	PPB	01/22/2025 14:50	SMPL
Q1106-11	YE8H5	366	0.0516	-	SFAM01.1	PPB	01/22/2025 14:53	SMPL
Q1106-12	YE8H6	21	-0.0321	-	SFAM01.1	PPB	01/22/2025 14:55	SMPL
Q1106-13	YE8F6	16	-0.0333	-	SFAM01.1	PPB	01/22/2025 14:57	SMPL
Q1106-14	YE8F9	44	-0.0265	-	SFAM01.1	PPB	01/22/2025 14:59	SMPL
Q1106-15	YE8G0	134	-0.0047	-	SFAM01.1	PPB	01/22/2025 15:02	SMPL
Q1106-16	YE8G1	54	-0.0241	-	SFAM01.1	PPB	01/22/2025 15:04	SMPL
Q1106-17	YE8G2	-3	-0.0379	-	SFAM01.1	PPB	01/22/2025 15:06	SMPL
Q1106-18	YE8H9	20	-0.0323	-	SFAM01.1	PPB	01/22/2025 15:08	SMPL
Q1106-19	YE8J0	76	-0.0187	-	SFAM01.1	PPB	01/22/2025 15:11	SMPL
CCV060	CCV060	20453	4.9266	-	SFAM01.1	PPB	01/22/2025 15:17	SMPL
CCB060	CCB160	-120	-0.0663	-	SFAM01.1	PPB	01/22/2025 15:20	SMPL
Q1106-20	YE8J1	499	0.0839	-	SFAM01.1	PPB	01/22/2025 15:22	SMPL
PB166179BL	PBW179	-40	-0.0469	-	SFAM01.1	PPB	01/22/2025 15:24	SMPL
Q1129-01	MJNLA3	168	0.0036	-	SFAM01.1	PPB	01/22/2025 15:27	SMPL
Q1129-02	MJNLA3D	89	-0.0156	-	SFAM01.1	PPB	01/22/2025 15:29	SMPL
Q1129-03	MJNLA3S	3677	0.8552	-	SFAM01.1	PPB	01/22/2025 15:31	SMPL
PB166178TB	LEB178	-7	-0.0389	-	SFAM01.1	PPB	01/22/2025 15:33	SMPL
PB166132TB	LEB132	92	-0.0149	-	SFAM01.1	PPB	01/22/2025 15:36	SMPL
CCV061	CCV061	22324	5.3807	-	SFAM01.1	PPB	01/22/2025 15:41	SMPL
CCB061	CCB061	-152	-0.0741	-	SFAM01.1	PPB	01/22/2025 15:43	SMPL