

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV013				ID: CCV014		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	101	0	400000	410000	103	1	400000	99	1
Calcium	10000	10000	104	0	400000	410000	102	0	390000	98	1
Iron	10000	9800	98	0	400000	420000	104	0	400000	100	0
Magnesium	6000	5800	97	0	400000	410000	102	0	390000	98	0
Potassium	9900	9300	94	1	150000	160000	105	1	150000	101	0
Sodium	10000	10000	100	2	400000	380000	95	0	370000	92	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV015				ID: CCV016		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	380000	95	1	410000	102	1
Calcium					400000	380000	94	0	400000	99	1
Iron					400000	380000	96	0	410000	101	1
Magnesium					400000	380000	94	0	390000	98	1
Potassium					150000	150000	97	0	160000	104	1
Sodium					400000	360000	91	0	370000	92	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV017				ID: CCV018		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	420000	105	0	390000	98	1
Calcium					400000	410000	102	1	380000	96	1
Iron					400000	410000	103	1	390000	98	0
Magnesium					400000	400000	100	0	380000	95	0
Potassium					150000	160000	105	1	150000	99	1
Sodium					400000	380000	95	0	360000	91	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84069
 Run Batch: LB134411 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV019				ID: CCV020		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	100	0	390000	98	1
Calcium					400000	390000	98	1	390000	97	1
Iron					400000	400000	101	0	400000	100	0
Magnesium					400000	390000	98	1	390000	97	0
Potassium					150000	150000	99	0	150000	98	0
Sodium					400000	380000	94	0	370000	93	0