

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No.: SDG No.: YE8F8
Instrument ID: CV1 Start Date: 02/07/2025
Analytical Method: CVAA Run Batch: LB134618
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.19	-7	2.5	2.5	1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No.: SDG No.: YE8F8
Instrument ID: CV1 Start Date: 02/07/2025
Analytical Method: CVAA Run Batch: LB134618
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.0	0	7.5	7.6	2	10	9.9	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No.: SDG No.: YE8F8
Instrument ID: CV1 Start Date: 02/07/2025
Analytical Method: CVAA Run Batch: LB134635
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.039	0	0.2	0.24	20	2.5	2.5	1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No.: SDG No.: YE8F8
Instrument ID: CV1 Start Date: 02/07/2025
Analytical Method: CVAA Run Batch: LB134635
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.7	-7	7.5	7.9	6	10	9.8	-2