

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
 Initial Calibration Verification Source : MP84371
 Continuing Calibration Verification Source : MP84373
 Run Batch: LB134618 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV090				ID: CCV037				ID: CCV038		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.0	100		5.0	5.1	101		4.9	98	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
 Initial Calibration Verification Source : MP84401
 Continuing Calibration Verification Source : MP84403
 Run Batch: LB134635 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV092				ID: CCV044				ID: CCV045		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	94		5.0	4.9	97		5.3	105	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8F8
Initial Calibration Verification Source : MP84401
Continuing Calibration Verification Source : MP84403
Run Batch: LB134635 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV046				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.2	104				