

LB134618


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

40373



A= 0.0000e+000

B= 2.4707e-004 slope

C= -6.3044e-002 y-intercept

Rho= 0.9998511

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.016	-0.016	191	0.000	191					
0.05	0.050										=
0.20	0.200	0.187	-0.013	1013	0.0 %	1013					-7
2.50	2.500	2.514	0.014	10432	0.0 %	10432					1
5.00	5.000	4.978	-0.022	20402	0.0 %	20402					0
7.50	7.500	7.624	0.124	31114	0.0 %	31114					2
10.0	10.000	9.912	-0.088	40373	0.0 %	40373					1

LB134618
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	191	-		0 SFAM01.1	PPB	2/7/2025 11:14	Std
	0.2 S01	1013	-		0.2 SFAM01.1	PPB	2/7/2025 11:16	Std
	2.5 S02	10432	-		2.5 SFAM01.1	PPB	2/7/2025 11:19	Std
	5 S03	20402	-		5 SFAM01.1	PPB	2/7/2025 11:21	Std
	7.5 S04	31114	-		7.5 SFAM01.1	PPB	2/7/2025 11:23	Std
	10 S05	40373	-		10 SFAM01.1	PPB	2/7/2025 11:25	Std
ICV090	ICV090	16401	3.9892 -		SFAM01.1	PPB	2/7/2025 11:31	SMPL
ICB090	ICB090	-84	-0.0838 -		SFAM01.1	PPB	2/7/2025 11:34	SMPL
CCV037	CCV037	20765	5.0674 -		SFAM01.1	PPB	2/7/2025 11:36	SMPL
CCB037	CCB037	-55	-0.0766 -		SFAM01.1	PPB	2/7/2025 11:41	SMPL
PB166616BL	PBW616	109	-0.0361 -		SFAM01.1	PPB	2/7/2025 11:43	SMPL
Q1135-09	YE8H7	113	-0.0351 -		SFAM01.1	PPB	2/7/2025 11:45	SMPL
Q1176-01	ME2931	218	-0.0092 -		SFAM01.1	PPB	2/7/2025 11:48	SMPL
Q1176-02	ME2933	134	-0.0299 -		SFAM01.1	PPB	2/7/2025 11:50	SMPL
Q1176-03	ME2937	292	0.0091 -		SFAM01.1	PPB	2/7/2025 11:52	SMPL
Q1176-04	ME2945	191	-0.0159 -		SFAM01.1	PPB	2/7/2025 11:54	SMPL
Q1176-05	ME2942	165	-0.0223 -		SFAM01.1	PPB	2/7/2025 11:57	SMPL
Q1176-06	ME2943	114	-0.0349 -		SFAM01.1	PPB	2/7/2025 11:59	SMPL
Q1176-07	ME2940	389	0.0331 -		SFAM01.1	PPB	2/7/2025 12:01	SMPL
Q1176-08	ME2941	79	-0.0435 -		SFAM01.1	PPB	2/7/2025 12:03	SMPL
Q1176-09	ME2944	139	-0.0287 -		SFAM01.1	PPB	2/7/2025 12:06	SMPL
Q1176-10	ME2944D	217	-0.0094 -		SFAM01.1	PPB	2/7/2025 12:08	SMPL
Q1176-11	ME2944S	4335	1.008 -		SFAM01.1	PPB	2/7/2025 12:10	SMPL
Q1176-12	ME2938	230	-0.0062 -		SFAM01.1	PPB	2/7/2025 12:12	SMPL
Q1176-13	ME2939	264	0.0022 -		SFAM01.1	PPB	2/7/2025 12:15	SMPL
Q1176-14	ME2932	340	0.021 -		SFAM01.1	PPB	2/7/2025 12:17	SMPL
Q1176-15	ME2936	101	-0.0381 -		SFAM01.1	PPB	2/7/2025 12:19	SMPL
Q1176-16	ME2934	319	0.0158 -		SFAM01.1	PPB	2/7/2025 12:22	SMPL
Q1176-17	ME2935	93	-0.0401 -		SFAM01.1	PPB	2/7/2025 12:24	SMPL
Q1176-18	ME2950	150	-0.026 -		SFAM01.1	PPB	2/7/2025 12:26	SMPL
CCV038	CCV038	20175	4.9217 -		SFAM01.1	PPB	2/7/2025 12:28	SMPL
CCB038	CCB038	-37	-0.0722 -		SFAM01.1	PPB	2/7/2025 12:31	SMPL
Q1176-19	ME2951	192	-0.0156 -		SFAM01.1	PPB	2/7/2025 12:33	SMPL
Q1176-20	ME2953	259	0.0009 -		SFAM01.1	PPB	2/7/2025 12:35	SMPL
Q1176-21	ME2954	180	-0.0186 -		SFAM01.1	PPB	2/7/2025 12:37	SMPL
PB166617BL	PBW617	86	-0.0418 -		SFAM01.1	PPB	2/7/2025 12:40	SMPL
Q1151-13	MCOB02	106	-0.0369 -		SFAM01.1	PPB	2/7/2025 12:42	SMPL
Q1186-01	ME2948	170	-0.021 -		SFAM01.1	PPB	2/7/2025 12:44	SMPL
Q1186-02	ME2949	135	-0.0297 -		SFAM01.1	PPB	2/7/2025 12:47	SMPL
Q1186-03	ME2955	143	-0.0277 -		SFAM01.1	PPB	2/7/2025 12:49	SMPL
Q1186-04	ME2956	159	-0.0238 -		SFAM01.1	PPB	2/7/2025 12:51	SMPL
Q1186-05	ME2957	186	-0.0171 -		SFAM01.1	PPB	2/7/2025 12:53	SMPL
Q1186-06	ME2960	184	-0.0176 -		SFAM01.1	PPB	2/7/2025 12:56	SMPL
Q1186-07	ME2961	124	-0.0324 -		SFAM01.1	PPB	2/7/2025 12:58	SMPL
Q1186-08	ME2962	181	-0.0183 -		SFAM01.1	PPB	2/7/2025 13:00	SMPL
Q1186-09	ME2959	127	-0.0317 -		SFAM01.1	PPB	2/7/2025 13:02	SMPL

LB134618
INSTRUMENT ID : CV1

Q1186-10	ME2959D	119	-0.0336 -	SFAM01.1	PPB	2/7/2025 13:05 SMPL
Q1186-11	ME2959S	4265	0.9907 -	SFAM01.1	PPB	2/7/2025 13:07 SMPL
Q1186-12	ME2963	82	-0.0428 -	SFAM01.1	PPB	2/7/2025 13:09 SMPL
Q1186-13	ME2967	171	-0.0208 -	SFAM01.1	PPB	2/7/2025 13:11 SMPL
Q1186-14	ME2965	263	0.0019 -	SFAM01.1	PPB	2/7/2025 13:14 SMPL
Q1186-15	ME2966	166	-0.022 -	SFAM01.1	PPB	2/7/2025 13:16 SMPL
CCV039	CCV039	20280	4.9476 -	SFAM01.1	PPB	2/7/2025 13:18 SMPL
CCB039	CCB039	-94	-0.0863 -	SFAM01.1	PPB	2/7/2025 13:21 SMPL
Q1186-16	ME2958	76	-0.0443 -	SFAM01.1	PPB	2/7/2025 13:23 SMPL
Q1186-17	ME2968	95	-0.0396 -	SFAM01.1	PPB	2/7/2025 13:25 SMPL
Q1186-18	ME2974	121	-0.0331 -	SFAM01.1	PPB	2/7/2025 13:27 SMPL
Q1186-19	ME2977	149	-0.0262 -	SFAM01.1	PPB	2/7/2025 13:30 SMPL
Q1186-20	ME2980	177	-0.0193 -	SFAM01.1	PPB	2/7/2025 13:32 SMPL
CCV040	CCV040	20372	4.9703 -	SFAM01.1	PPB	2/7/2025 13:34 SMPL
CCB040	CCB040	-33	-0.0712 -	SFAM01.1	PPB	2/7/2025 13:36 SMPL