

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134674

Review By	Janvi Patel	Review On	2/12/2025 11:01:11 AM
Supervise By	Mohan Bera	Supervise On	2/12/2025 11:10:40 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84414,MP84423,MP84422,MP84420,MP84419,MP84418,MP84417,MP84416,MP84415		
ICV Standard	MP84473		
CCV Standard	MP84438		
ICSA Standard	MP84440,MP84441		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84444,MP84443		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE005	TUNE	02/11/25 11:22		Jaswal	OK
2	S00	S0	CAL1	02/11/25 12:05		Jaswal	OK
3	S02	S02	CAL3	02/11/25 12:08		Jaswal	OK
4	S03	S03	CAL4	02/11/25 12:11		Jaswal	OK
5	S04	S04	CAL5	02/11/25 12:15		Jaswal	OK
6	S05	S05	CAL6	02/11/25 12:18		Jaswal	OK
7	S06	S06	CAL7	02/11/25 12:20		Jaswal	OK
8	S07	S07	CAL8	02/11/25 12:23		Jaswal	OK
9	S08	S08	CAL9	02/11/25 12:26		Jaswal	OK
10	ICV005	ICV005	ICV	02/11/25 13:29	Inst. paused for Calibration review	Jaswal	OK
11	ICB005	ICB005	ICB	02/11/25 13:33		Jaswal	OK
12	ICSA005	ICSA005	ICSA	02/11/25 13:40	Inst. paused for review	Jaswal	OK
13	ICSAB005	ICSAB005	ICSAB	02/11/25 13:44		Jaswal	OK
14	CCV021	CCV021	CCV	02/11/25 13:47		Jaswal	OK
15	CCB021	CCB021	CCB	02/11/25 13:50		Jaswal	OK
16	Q1159-01DL	YE8C9	SAM	02/11/25 13:53	Sr high in MS	Jaswal	Not Ok
17	Q1159-02DL	YE8C9D	DUP	02/11/25 13:57	Sr high in MS	Jaswal	Not Ok
18	Q1159-01LDL	YE8C9L	SD	02/11/25 14:00	Sr high in MS	Jaswal	Not Ok

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134674

Review By	Janvi Patel	Review On	2/12/2025 11:01:11 AM
Supervise By	Mohan Bera	Supervise On	2/12/2025 11:10:40 AM

STD. NAME	STD REF.#
ICAL Standard	MP84414,MP84423,MP84422,MP84420,MP84419,MP84418,MP84417,MP84416,MP84415
ICV Standard	MP84473
CCV Standard	MP84438
ICSA Standard	MP84440,MP84441
CRI Standard	
LCS Standard	MP84099
Chk Standard	MP84444,MP84443

19	Q1159-03DL	YE8C9S	MS	02/11/25 14:03	2x for Sr still Sr high	Jaswal	Not Ok
20	PB166570BL	PBW570	MB	02/11/25 14:15		Jaswal	OK
21	PB166570BS	LCS570	LCS	02/11/25 14:18		Jaswal	OK
22	Q1135-09	YE8H7	SAM	02/11/25 14:22		Jaswal	OK
23	PB166317BL	PBW317	MB	02/11/25 14:25		Jaswal	OK
24	PB166317BS	LCS317	LCS	02/11/25 14:29		Jaswal	OK
25	Q1186-01	ME2948	SAM	02/11/25 14:32		Jaswal	OK
26	Q1159-01DL2	YE8C9	SAM	02/11/25 14:35	4x For Sr	Jaswal	Confirms
27	Q1159-02DL2	YE8C9D	DUP	02/11/25 14:38	4x For Sr	Jaswal	Confirms
28	Q1159-01LDL2	YE8C9L	SD	02/11/25 14:42	20x For Sr	Jaswal	Confirms
29	Q1159-03DL2	YE8C9S	MS	02/11/25 14:45	4x For Sr	Jaswal	Confirms
30	CCV022	CCV022	CCV	02/11/25 14:50		Jaswal	OK
31	CCB022	CCB022	CCB	02/11/25 14:53		Jaswal	OK
32	Q1204-01A	ME2975A	PS	02/11/25 15:21	Se 0.1ml of mp84474	Jaswal	OK
33	Q1245-09A	A6335A	PS	02/11/25 15:25	Se 0.1ml of mp84474	Jaswal	OK
34	PB166636BL	PBW636	MB	02/11/25 15:28		Jaswal	OK
35	PB166636BS	LCS636	LCS	02/11/25 15:31		Jaswal	OK
36	Q1255-01	MDCZM3	SAM	02/11/25 15:41		Jaswal	OK
37	Q1255-02	MDCZP0	SAM	02/11/25 15:45		Jaswal	OK
38	Q1255-03	MDCZP1	SAM	02/11/25 15:48		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134674

Review By	Janvi Patel	Review On	2/12/2025 11:01:11 AM
Supervise By	Mohan Bera	Supervise On	2/12/2025 11:10:40 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84414,MP84423,MP84422,MP84420,MP84419,MP84418,MP84417,MP84416,MP84415		
ICV Standard	MP84473		
CCV Standard	MP84438		
ICSA Standard	MP84440,MP84441		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84444,MP84443		

39	Q1255-04	MDCZP5	SAM	02/11/25 15:51		Jaswal	OK
40	Q1255-05	MDCZM4	SAM	02/11/25 15:54		Jaswal	OK
41	Q1255-06	MDCZN3	SAM	02/11/25 15:58		Jaswal	OK
42	Q1255-07	MDCZP2	SAM	02/11/25 16:01		Jaswal	OK
43	Q1255-08	MDCZN8	SAM	02/11/25 16:04		Jaswal	OK
44	Q1255-09	MDCZN8D	DUP	02/11/25 16:08		Jaswal	OK
45	Q1255-08L	MDCZN8L	SD	02/11/25 16:11		Jaswal	OK
46	Q1255-10	MDCZN8S	MS	02/11/25 16:14	MS fail for Se(Below RL)	Jaswal	OK
47	Q1255-11	MDCZL6	SAM	02/11/25 16:17		Jaswal	OK
48	CCV023	CCV023	CCV	02/11/25 16:21		Jaswal	OK
49	CCB023	CCB023	CCB	02/11/25 16:24		Jaswal	OK
50	Q1255-12	MDCZL9	SAM	02/11/25 16:32		Jaswal	OK
51	Q1255-13	MDCZN9	SAM	02/11/25 16:36		Jaswal	OK
52	Q1255-14	MDCZN2	SAM	02/11/25 16:39		Jaswal	OK
53	Q1255-15	MDCZN7	SAM	02/11/25 16:42		Jaswal	OK
54	Q1255-16	MDCZP4	SAM	02/11/25 16:46		Jaswal	OK
55	CCV024	CCV024	CCV	02/11/25 16:49		Jaswal	OK
56	CCB024	CCB024	CCB	02/11/25 16:52		Jaswal	OK