

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1145 SAS No.: Q1145 SDG No.: Q1145

Instrument ID: MSVOA_X Calibration Date/Time: 01/23/2025 08:21

Lab File ID: VX044699.D Init. Calib. Date(s): 01/16/2025 01/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:39 10:12

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.665	1.678	0.5	0.78	
Toluene	1.975	2.000	0.4	1.27	
Ethyl Benzene	2.156	2.096	0.1	-2.78	
m/p-Xylenes	0.792	0.759	0.3	-4.17	
o-Xylene	0.795	0.794	0.3	-0.13	
1,2-Dichloroethane-d4	2.896	2.644	0.01	-8.7	
Toluene-d8	1.632	1.723	0.01	5.58	
4-Bromofluorobenzene	0.583	0.546	0.2	-6.35	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.