

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1145

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VX044700.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0123WBS01	Benzene	20	19.9	ug/L	100			65	135	
	Toluene	20	19.2	ug/L	96			70	130	
	Ethyl Benzene	20	19.2	ug/L	96			60	140	
	m/p-Xylenes	40	39.1	ug/L	98			87	111	
	o-Xylene	20	19.5	ug/L	98			87	111	