

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q1145 SAS No.: Q1145 SDG NO.: Q1145
 Lab File ID: VX044699.D Date Analyzed: 01/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:21
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	17047	4.89	92191	6.75	78268	10.05
UPPER LIMIT	34094	5.385	184382	7.251	156536	10.549
LOWER LIMIT	8523.5	4.385	46095.5	6.251	39134	9.549
EPA SAMPLE NO.						
001 WILLETS PT BLVD (DEC)	29592	4.89	163912	6.76	147341	10.05
002 35TH AVE (DEC)	28355	4.89	158559	6.76	144308	10.05
VX0123WBL01	30850	4.90	168575	6.76	149390	10.05
VX0123WBS01	31428	4.89	167207	6.75	150020	10.05
VX0123WBSD01	30388	4.89	171874	6.75	151608	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.